

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052716\
 Data File : BF087462.D
 Acq On : 28 May 2016 2:04
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: May 28 04:20:11 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 16:26:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	185995	20.00	ng	-0.03
21) Naphthalene-d8	9.52	136	774785	20.00	ng	-0.03
38) Acenaphthene-d10	12.35	164	365984	20.00	ng	-0.03
63) Phenanthrene-d10	14.75	188	676453	20.00	ng	-0.02
75) Chrysene-d12	18.46	240	497555	20.00	ng	-0.02
86) Perylene-d12	20.11	264	393609	20.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.38	112	936224	88.45	ng	-0.03
7) Phenol-d6	7.05	99	1187979	83.06	ng	-0.02
23) Nitrobenzene-d5	8.42	82	1250719	81.36	ng	-0.02
41) 2,4,6-Tribromophenol	13.65	330	284928	81.01	ng	-0.03
44) 2-Fluorobiphenyl	11.30	172	1959010	75.46	ng	-0.03
78) Terphenyl-d14	17.11	244	1922574	82.15	ng	-0.02

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	1.78	88	216774	38.81	ng	# 74
3) Pyridine	2.36	79	463862	37.99	ng	# 64
4) n-Nitrosodimethylamine	2.34	42	408637	43.43	ng	# 45
6) Aniline	6.99	93	747179	40.38	ng	94
8) 2-Chlorophenol	7.18	128	534331	40.48	ng	94
9) Benzaldehyde	6.81	77	303168	38.39	ng	98
10) Phenol	7.07	94	662215	42.40	ng	86
11) bis(2-Chloroethyl)ether	7.14	93	514940	40.74	ng	87
12) 1,3-Dichlorobenzene	7.38	146	565645	39.29	ng	# 90
13) 1,4-Dichlorobenzene	7.52	146	591382	40.01	ng	96
14) 1,2-Dichlorobenzene	7.75	146	544170	39.80	ng	97
15) Benzyl Alcohol	7.79	79	448814	43.04	ng	# 88
16) 2,2'-oxybis(1-Chloropropan	7.98	45	1094983	38.50	ng	87
17) 2-Methylphenol	8.01	107	444496	42.02	ng	# 89
18) Hexachloroethane	8.26	117	223165	40.47	ng	96
19) n-Nitroso-di-n-propylamine	8.23	70	439946	41.25	ng	# 73
20) 3+4-Methylphenols	8.27	107	567864	44.40	ng	# 61
22) Acetophenone	8.19	105	768564	41.52	ng	# 98
24) Nitrobenzene	8.44	77	638742	39.36	ng	96
25) Isophorone	8.84	82	1141724	41.41	ng	98
26) 2-Nitrophenol	8.95	139	281723	42.47	ng	# 72
27) 2,4-Dimethylphenol	9.10	122	486930	42.29	ng	87
28) bis(2-Chloroethoxy)methane	9.22	93	630354	40.59	ng	99
29) 2,4-Dichlorophenol	9.36	162	436822	42.09	ng	96
30) 1,2,4-Trichlorobenzene	9.45	180	469267	39.50	ng	97
31) Naphthalene	9.55	128	1528608	39.37	ng	99
32) Benzoic acid	9.47	122	241010	41.54	ng	88
33) 4-Chloroaniline	9.70	127	555786	38.87	ng	# 91
34) Hexachlorobutadiene	9.77	225	291513	40.37	ng	98
35) Caprolactam	10.38	113	106351	35.11	ng	# 56
36) 4-Chloro-3-methylphenol	10.57	107	496628	42.33	ng	88
37) 2-Methylnaphthalene	10.68	142	976510	40.26	ng	97
39) 1,2,4,5-Tetrachlorobenzene	10.96	216	453050	38.67	ng	98
40) Hexachlorocyclopentadiene	10.92	237	131933	33.59	ng	96


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052716\
 Data File : BF087462.D
 Acq On : 28 May 2016 2:04
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: May 28 04:20:11 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 16:26:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	11.19	196	270764	40.04	ng	94
43) 2,4,5-Trichlorophenol	11.27	196	266967	38.77	ng #	92
45) 1,1'-Biphenyl	11.45	154	1174665	38.12	ng	96
46) 2-Chloronaphthalene	11.47	162	920562	39.05	ng	97
47) 2-Nitroaniline	11.69	65	360832	42.48	ng #	77
48) Acenaphthylene	12.12	152	1413069	37.86	ng	99
49) Dimethylphthalate	12.01	163	1146224	39.10	ng	100
50) 2,6-Dinitrotoluene	12.11	165	241884	40.95	ng #	80
51) Acenaphthene	12.41	154	898139	39.15	ng	98
52) 3-Nitroaniline	12.38	138	249524	41.16	ng	86
53) 2,4-Dinitrophenol	12.58	184	114838	41.42	ng #	84
54) Dibenzofuran	12.70	168	1198132	38.58	ng	97
55) 4-Nitrophenol	12.76	139	114929	39.74	ng #	21
56) 2,4-Dinitrotoluene	12.76	165	312084	39.53	ng #	89
57) Fluorene	13.25	166	1007109	37.39	ng	98
58) 2,3,4,6-Tetrachlorophenol	12.92	232	211975	40.02	ng	96
59) Diethylphthalate	13.15	149	1114947	39.12	ng	99
60) 4-Chlorophenyl-phenylether	13.27	204	498533	37.96	ng	99
61) 4-Nitroaniline	13.38	138	224791	39.72	ng #	63
62) Azobenzene	13.53	77	1201786	40.64	ng	86
64) 4,6-Dinitro-2-methylphenol	13.44	198	176099	40.94	ng #	83
65) n-Nitrosodiphenylamine	13.50	169	871013	39.81	ng	99
66) 4-Bromophenyl-phenylether	14.06	248	285310	38.55	ng	99
67) Hexachlorobenzene	14.13	284	300541	38.83	ng #	67
68) Atrazine	14.41	200	219528	31.48	ng	93
69) Pentachlorophenol	14.49	266	103819	46.00	ng	96
70) Phenanthrene	14.79	178	1440445	38.44	ng	100
71) Anthracene	14.88	178	1502354	39.69	ng	100
72) Carbazole	15.17	167	1266107	39.22	ng	98
73) Di-n-butylphthalate	15.74	149	1679557	40.23	ng #	98
74) Fluoranthene	16.55	202	1414458	37.65	ng	98
76) Benzidine	16.78	184	344939	26.94	ng	98
77) Pyrene	16.86	202	1507351	42.09	ng	100
79) Butylbenzylphthalate	17.77	149	700242	44.09	ng #	85
80) Benzo(a)anthracene	18.43	228	1111800	39.28	ng	99
81) 3,3'-Dichlorobenzidine	18.42	252	633950	40.55	ng	99
82) Chrysene	18.49	228	1132099	40.30	ng	100
83) Bis(2-ethylhexyl)phthalate	18.51	149	967681	41.75	ng	98
84) Di-n-octyl phthalate	19.28	149	1368085	38.71	ng #	84
85) Indeno(1,2,3-cd)pyrene	21.34	276	935259	41.54	ng	94
87) Benzo(b)fluoranthene	19.71	252	1837883	73.30	ng #	95
88) Benzo(k)fluoranthene	19.71	252	1838997	85.21	ng #	97
89) Benzo(a)pyrene	20.06	252	859095	39.62	ng #	95
90) Dibenzo(a,h)anthracene	21.34	278	783975	42.39	ng	97
91) Benzo(g,h,i)perylene	21.68	276	777397	42.60	ng #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.48 min Scan# 516

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:152 Resp: 185995

Ion Ratio Lower Upper

152 100

150 156.5 125.8 188.6




115 70.6 51.5 77.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 38.81 ng

RT: 1.78 min Scan# 17

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 88 Resp: 216774

Ion Ratio Lower Upper

88 100

58 94.2 59.6 89.4#



43 44.2 21.8 32.6#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 37.99 ng

RT: 2.36 min Scan# 68

Delta R.T. -0.05 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 79 Resp: 463862

Ion Ratio Lower Upper

79 100

52 100.8 55.9 83.9#




51 53.9 28.1 42.1#



Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0



#4

n-Nitrosodimethylamine

Concen: 43.43 ng

RT: 2.34 min Scan# 66

Delta R.T. -0.05 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 42 Resp: 408637

Ion Ratio Lower Upper

42 100

74 80.5 123.4 185.0#



44 5.9 5.8 8.6



Abundance Ion 42.00 (41.70 to 42.70): BF0

Ion 74.00 (73.70 to 74.70): BF0

Ion 44.00 (43.70 to 44.70): BF0





#5

2-Fluorophenol

Concen: 88.45 ng

RT: 5.38 min Scan# 332

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 112 Resp: 936224

Ion Ratio Lower Upper

112 100

64 68.4 52.1 78.1




63 37.9 25.7 38.5



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#6

Aniline

Concen: 40.38 ng

RT: 6.99 min Scan# 473

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 93 Resp: 747179

Ion Ratio Lower Upper

93 100

66 44.5 39.0 58.6



65 23.1 20.6 31.0




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 66.00 (65.70 to 66.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#7

Phenol-d6

Concen: 83.06 ng

RT: 7.05 min Scan# 478

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 99 Resp: 1187979

Ion Ratio Lower Upper

99 100

42 30.2 13.6 20.4#




71 41.5 24.6 36.8#



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 40.48 ng

RT: 7.18 min Scan# 489

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 128 Resp: 534331

Ion Ratio Lower Upper

128 100

130 33.3 12.5 52.5



64 53.8 27.3 67.3




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 38.39 ng

RT: 6.81 min Scan# 457

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 77 Resp: 303168

Ion Ratio Lower Upper

77 100

105 95.0 72.7 112.7




106 88.7 70.8 110.8



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 42.40 ng

RT: 7.07 min Scan# 480

Delta R.T. -0.01 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 94 Resp: 662215

Ion Ratio Lower Upper

94 100

65 32.7 7.2 47.2



66 44.1 14.9 54.9




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0





#11

bis(2-Chloroethyl)ether

Concen: 40.74 ng

RT: 7.14 min Scan# 486

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 93 Resp: 514940

Ion Ratio Lower Upper

93 100

63 92.3 58.0 98.0




95 33.9 11.9 51.9



Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 63.00 (62.70 to 63.70): BF0

Ion 95.00 (94.70 to 95.70): BF0



#12

1,3-Dichlorobenzene

Concen: 39.29 ng

RT: 7.38 min Scan# 507

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 146 Resp: 565645

Ion Ratio Lower Upper

146 100

148 62.3 52.4 78.6



75 41.1 22.8 34.2#




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 75.00 (74.70 to 75.70): BF0









#13
 1,4-Dichlorobenzene
 Concen: 40.01 ng
 RT: 7.52 min Scan# 519
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.2	52.2	78.4
111	44.3	30.7	46.1



#14
 1,2-Dichlorobenzene
 Concen: 39.80 ng
 RT: 7.75 min Scan# 539
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion	Ratio	Lower	Upper
146	100		
148	64.1	51.0	76.6
111	50.0	36.2	54.4









#15
Benzyl Alcohol
Concen: 43.04 ng
RT: 7.79 min Scan# 543
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 448814
Ion Ratio Lower Upper
79 100
108 68.2 68.4 102.6#
77 64.9 50.6 76.0



#16
2,2'-oxybis(1-Chloropropane)
Concen: 38.50 ng
RT: 7.98 min Scan# 559
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion: 45 Resp: 1094983
Ion Ratio Lower Upper
45 100
77 10.8 0.0 36.4
79 7.9 0.0 33.4





#17

2-Methylphenol

Concen: 42.02 ng

RT: 8.01 min Scan# 562

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:107 Resp: 444496





Ion Ratio Lower Upper

107 100

108 112.2 93.7 140.5

77 55.9 33.4 50.0#

79 54.0 34.3 51.5#



#18

Hexachloroethane

Concen: 40.47 ng

RT: 8.26 min Scan# 584

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04




Tgt Ion:117 Resp: 223165




Ion Ratio Lower Upper

117 100

119 96.0 76.5 114.7

201 85.4 63.0 94.6









#19
n-Nitroso-di-n-propylamine
Concen: 41.25 ng
RT: 8.23 min Scan# 581
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 70 Resp: 439946
Ion Ratio Lower Upper
70 100
42 81.4 43.1 64.7#
101 8.9 8.1 12.1
130 18.5 18.6 28.0#



#20
3+4-Methylphenols
Concen: 44.40 ng
RT: 8.27 min Scan# 585
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion: 107 Resp: 567864
Ion Ratio Lower Upper
107 100
108 88.3 68.5 108.5
77 38.6 101.6 141.6#
79 33.0 7.4 47.4









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 9.52 min Scan# 694
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:	136	Resp:	774785
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.8	13.2
54	12.1	5.0	7.4#
68	5.9	4.4	6.6



#22
Acetophenone
Concen: 41.52 ng
RT: 8.19 min Scan# 578
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:	105	Resp:	768564
Ion	Ratio	Lower	Upper
105	100		
71	0.9	4.8	7.2#
51	40.2	32.6	49.0
120	20.6	16.5	24.7









#23
 Nitrobenzene-d5
 Concen: 81.36 ng
 RT: 8.42 min Scan# 598
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion: 82 Resp: 1250719
 Ion Ratio Lower Upper
 82 100
 128 39.3 40.6 61.0#
 54 69.8 38.1 57.1#



#24
 Nitrobenzene
 Concen: 39.36 ng
 RT: 8.44 min Scan# 600
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion: 77 Resp: 638742
 Ion Ratio Lower Upper
 77 100
 123 38.7 33.0 49.4
 65 15.5 10.8 16.2





#25

Isophorone

Concen: 41.41 ng

RT: 8.84 min Scan# 635

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 82 Resp: 1141724

Ion Ratio Lower Upper

82 100

95 7.7 5.1 7.7

138 15.8 12.4 18.6



#26

2-Nitrophenol

Concen: 42.47 ng

RT: 8.95 min Scan# 644

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04




Tgt Ion: 139 Resp: 281723

Ion Ratio Lower Upper

139 100

109 35.3 19.5 29.3#

65 68.0 37.5 56.3#









#27
 2,4-Dimethylphenol
 Concen: 42.29 ng
 RT: 9.10 min Scan# 657
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:122 Resp: 486930
 Ion Ratio Lower Upper
 122 100
 107 121.6 82.0 123.0
 121 59.6 46.1 69.1



#28
 bis(2-Chloroethoxy)methane
 Concen: 40.59 ng
 RT: 9.22 min Scan# 668
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion: 93 Resp: 630354
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.0 39.0
 123 10.0 9.5 14.3









#29
 2,4-Dichlorophenol
 Concen: 42.09 ng
 RT: 9.36 min Scan# 680
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:162 Resp: 436822
 Ion Ratio Lower Upper
 162 100
 164 63.5 42.2 82.2
 98 43.0 19.0 59.0



#30
 1,2,4-Trichlorobenzene
 Concen: 39.50 ng
 RT: 9.45 min Scan# 688
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:180 Resp: 469267
 Ion Ratio Lower Upper
 180 100
 182 94.8 78.0 117.0
 145 32.9 24.2 36.2





#31

Naphthalene

Concen: 39.37 ng

RT: 9.55 min Scan# 697

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:128 Resp: 1528608

Ion Ratio Lower Upper

128 100

129 11.6 8.6 13.0




127 13.7 10.8 16.2



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 41.54 ng

RT: 9.47 min Scan# 690

Delta R.T. 0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:122 Resp: 241010

Ion Ratio Lower Upper

122 100

105 125.4 92.6 132.6



77 90.8 60.0 100.0




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0





#33

4-Chloroaniline

Concen: 38.87 ng

RT: 9.70 min Scan# 710

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:127 Resp: 555786


Ion Ratio Lower Upper

127 100

129 32.8 26.2 39.4

65 39.9 23.0 34.4#

92 21.7 15.1 22.7




Abundance Ion 127.00 (126.70 to 127.70): E



Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



Time-->



#34

Hexachlorobutadiene

Concen: 40.37 ng

RT: 9.77 min Scan# 716

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:225 Resp: 291513

Ion Ratio Lower Upper

225 100

223 62.2 51.5 77.3


227 64.1 52.2 78.2




Abundance Ion 225.00 (224.70 to 225.70): E


Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



Time-->









#35
 Caprolactam
 Concen: 35.11 ng
 RT: 10.38 min Scan# 769
 Delta R.T. -0.01 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:113 Resp: 106351
 Ion Ratio Lower Upper
 113 100
 55 247.5 162.0 202.0#
 56 185.2 115.3 155.3#




Abundance Ion 113.00 (112.70 to 113.70): E
 Ion 55.00 (54.70 to 55.70): BF0
 Ion 56.00 (55.70 to 56.70): BF0






#36
 4-Chloro-3-methylphenol
 Concen: 42.33 ng
 RT: 10.57 min Scan# 786
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:107 Resp: 496628
 Ion Ratio Lower Upper
 107 100
 144 21.8 20.7 31.1
 142 67.3 63.2 94.8



Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 40.26 ng

RT: 10.68 min Scan# 796

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:142 Resp: 976510

Ion Ratio Lower Upper

142 100

141 88.7 71.0 106.6




115 39.5 26.6 39.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 12.35 min Scan# 942

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:164 Resp: 365984

Ion Ratio Lower Upper

164 100

162 104.8 82.8 124.2



160 47.9 36.9 55.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 38.67 ng

RT: 10.96 min Scan# 820

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:216 Resp: 453050


Ion Ratio Lower Upper

216 100

214 80.1 62.6 93.8

179 24.5 18.2 27.4

108 23.2 17.5 26.3






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 33.59 ng

RT: 10.92 min Scan# 817

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:237 Resp: 131933

Ion Ratio Lower Upper

237 100

235 63.5 47.2 87.2



272 11.8 0.0 31.1



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 81.01 ng
 RT: 13.65 min Scan# 1055
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.2	79.0	118.4
141	41.0	31.2	46.8



#42
 2,4,6-Trichlorophenol
 Concen: 40.04 ng
 RT: 11.19 min Scan# 840
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion	Ratio	Lower	Upper
196	100		
198	100.1	74.7	112.1
200	31.9	23.8	35.6









#43
 2,4,5-Trichlorophenol
 Concen: 38.77 ng
 RT: 11.27 min Scan# 847
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:	196	Resp:	266967
Ion	Ratio	Lower	Upper
196	100		
198	98.6	77.5	116.3
97	58.2	34.8	52.2#
132	30.9	23.0	34.4




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 75.46 ng
 RT: 11.30 min Scan# 850
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:	172	Resp:	1959010
Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.0	43.6
170	24.7	19.8	29.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 38.12 ng
 RT: 11.45 min Scan# 863
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:154 Resp: 1174665
 Ion Ratio Lower Upper
 154 100
 153 41.1 20.3 60.3
 76 15.5 0.0 30.2



#46
 2-Chloronaphthalene
 Concen: 39.05 ng
 RT: 11.47 min Scan# 865
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:162 Resp: 920562
 Ion Ratio Lower Upper
 162 100
 127 42.9 31.8 47.6
 164 33.0 27.0 40.6





#47

2-Nitroaniline

Concen: 42.48 ng

RT: 11.69 min Scan# 884

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 65 Resp: 360832

Ion Ratio Lower Upper




65 100

92 59.3 57.4 86.2

138 82.0 90.1 135.1#



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): E



#48

Acenaphthylene

Concen: 37.86 ng

RT: 12.12 min Scan# 922

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion: 152 Resp: 1413069

Ion Ratio Lower Upper



152 100


151 20.9 16.8 25.2

153 14.0 10.9 16.3



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#49
Dimethylphthalate
Concen: 39.10 ng
RT: 12.01 min Scan# 912
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:163 Resp: 1146224
Ion Ratio Lower Upper
163 100
194 3.2 2.6 4.0
164 10.3 8.2 12.4




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 40.95 ng
RT: 12.11 min Scan# 921
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:165 Resp: 241884
Ion Ratio Lower Upper
165 100
63 89.5 51.8 77.8#
89 57.3 50.7 76.1



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 39.15 ng

RT: 12.41 min Scan# 947

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:154 Resp: 898139

Ion Ratio Lower Upper

154 100

153 114.3 89.8 134.6




152 52.4 43.4 65.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 41.16 ng

RT: 12.38 min Scan# 944

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:138 Resp: 249524

Ion Ratio Lower Upper

138 100

108 10.8 7.8 11.6



92 125.6 87.8 131.8




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
2,4-Dinitrophenol
Concen: 41.42 ng
RT: 12.58 min Scan# 962
Delta R.T. -0.01 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC




Tgt Ion:184 Resp: 114838
Ion Ratio Lower Upper
184 100
63 88.5 55.8 83.8#
154 70.0 62.4 93.6



#54
Dibenzofuran
Concen: 38.58 ng
RT: 12.70 min Scan# 972
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:168 Resp: 1198132
Ion Ratio Lower Upper
168 100
139 41.1 31.4 47.0
169 13.1 10.7 16.1





#55

4-Nitrophenol

Concen: 39.74 ng

RT: 12.76 min Scan# 978

Delta R.T. -0.01 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:139 Resp: 114929

Ion Ratio Lower Upper

139 100

109 90.4 36.9 76.9#

65 175.7 64.0 104.0#



#56

2,4-Dinitrotoluene

Concen: 39.53 ng

RT: 12.76 min Scan# 978

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Tgt Ion:165 Resp: 312084


Ion Ratio Lower Upper

165 100

63 71.9 44.3 66.5#

89 90.5 70.0 105.0

182 2.2 1.8 2.6





#57

Fluorene

Concen: 37.39 ng

RT: 13.25 min Scan# 1020

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:166 Resp: 1007109

Ion Ratio Lower Upper

166 100

165 100.9 79.0 118.6

167 13.4 10.6 15.8



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E

1000000

800000

600000



400000

200000

0

13.10 13.20 13.30 13.40

Time-->



#58

2,3,4,6-Tetrachlorophenol

Concen: 40.02 ng

RT: 12.92 min Scan# 992

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Tgt Ion:232 Resp: 211975


Ion Ratio Lower Upper

232 100

131 55.8 41.9 62.9

130 3.0 2.2 3.4

166 34.9 26.8 40.2



Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E

1200000

1000000

800000

600000


400000

200000

0

12.80 13.00 13.20

Time-->









#59
Diethylphthalate
Concen: 39.12 ng
RT: 13.15 min Scan# 1012
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:149 Resp: 1114947
Ion Ratio Lower Upper
149 100
177 20.6 16.6 24.8
150 13.0 10.0 15.0



#60
4-Chlorophenyl-phenylether
Concen: 37.96 ng
RT: 13.27 min Scan# 1022
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:204 Resp: 498533
Ion Ratio Lower Upper
204 100
206 32.3 25.4 38.0
141 76.5 61.6 92.4









#61
4-Nitroaniline
Concen: 39.72 ng
RT: 13.38 min Scan# 1032
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:138 Resp: 224791
Ion Ratio Lower Upper
138 100
92 58.7 24.7 64.7
108 93.9 37.4 77.4#



#62
Azobenzene
Concen: 40.64 ng
RT: 13.53 min Scan# 1045
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion: 77 Resp: 1201786
Ion Ratio Lower Upper
77 100
182 19.8 0.0 38.8
105 19.9 3.2 43.2
51 42.9 8.8 48.8









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 14.75 min Scan# 1152
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:188 Resp: 676453
Ion Ratio Lower Upper
188 100
94 8.5 6.8 10.2
80 9.3 7.1 10.7



#64
4,6-Dinitro-2-methylphenol
Concen: 40.94 ng
RT: 13.44 min Scan# 1037
Delta R.T. -0.01 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:198 Resp: 176099
Ion Ratio Lower Upper
198 100
51 62.4 20.5 60.5#
105 44.3 24.5 64.5





#65

n-Nitrosodiphenylamine

Concen: 39.81 ng

RT: 13.50 min Scan# 1042

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:169 Resp: 871013

Ion Ratio Lower Upper

169 100

168 68.0 53.8 80.6




167 36.9 29.5 44.3



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 38.55 ng

RT: 14.06 min Scan# 1091

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:248 Resp: 285310

Ion Ratio Lower Upper

248 100

250 97.1 78.6 117.8



141 94.2 76.0 114.0




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 38.83 ng
RT: 14.13 min Scan# 1097
Delta R.T. -0.03 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:284 Resp: 300541
Ion Ratio Lower Upper
284 100
142 45.6 16.7 25.1#
249 35.4 21.3 31.9#




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 31.48 ng
RT: 14.41 min Scan# 1122
Delta R.T. -0.02 min
Lab File: BF087462.D
Acq: 28 May 2016 2:04

Tgt Ion:200 Resp: 219528
Ion Ratio Lower Upper
200 100
173 25.3 8.5 48.5
215 42.1 27.2 67.2



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 46.00 ng
 RT: 14.49 min Scan# 1129
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion: 266 Resp: 103819
 Ion Ratio Lower Upper
 266 100
 268 62.6 51.5 77.3
 264 61.5 52.9 79.3




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 38.44 ng
 RT: 14.79 min Scan# 1155
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion: 178 Resp: 1440445
 Ion Ratio Lower Upper
 178 100
 176 20.0 16.0 24.0
 179 15.9 12.6 18.8



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E









#71
 Anthracene
 Concen: 39.69 ng
 RT: 14.88 min Scan# 1163
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:178 Resp: 1502354
 Ion Ratio Lower Upper
 178 100
 176 19.5 15.6 23.4
 179 15.5 12.7 19.1



#72
 Carbazole
 Concen: 39.22 ng
 RT: 15.17 min Scan# 1188
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:167 Resp: 1266107
 Ion Ratio Lower Upper
 167 100
 166 22.5 17.7 26.5
 139 14.9 10.8 16.2





#73

Di-n-butylphthalate

Concen: 40.23 ng

RT: 15.74 min Scan# 1238

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion:149 Resp: 1679557

Ion Ratio Lower Upper

149 100

150 9.7 7.6 11.4

104 6.5 3.8 5.6#



#74

Fluoranthene

Concen: 37.65 ng

RT: 16.55 min Scan# 1309

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04




Tgt Ion:202 Resp: 1414458


Ion Ratio Lower Upper

202 100

101 13.3 0.0 35.4

203 18.1 0.0 38.1









#75
 Chrysene-d12
 Concen: 20.00 ng
 RT: 18.46 min Scan# 1476
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:240 Resp: 497555
 Ion Ratio Lower Upper
 240 100
 120 7.8 11.2 16.8#
 236 26.5 21.2 31.8



#76
 Benzidine
 Concen: 26.94 ng
 RT: 16.78 min Scan# 1329
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:184 Resp: 344939
 Ion Ratio Lower Upper
 184 100
 185 15.9 13.6 20.4
 183 11.8 9.8 14.6









#77
 Pyrene
 Concen: 42.09 ng
 RT: 16.86 min Scan# 1336
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:202 Resp: 1507351
 Ion Ratio Lower Upper
 202 100
 200 22.3 17.7 26.5
 203 17.9 14.2 21.4



#78
 Terphenyl-d14
 Concen: 82.15 ng
 RT: 17.11 min Scan# 1358
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion:244 Resp: 1922574
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.2 9.2
 122 11.3 12.0 18.0#





#79

Butylbenzylphthalate

Concen: 44.09 ng

RT: 17.77 min Scan# 1416

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 700242

Ion Ratio Lower Upper

149 100

91 74.7 48.0 72.0#

206 19.3 15.8 23.6



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF0



Ion 206.00 (205.70 to 206.70): E

17.77

400000

200000

Time-->



#80

Benzo(a)anthracene

Concen: 39.28 ng

RT: 18.43 min Scan# 1474

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:228 Resp: 1111800

Ion Ratio Lower Upper

228 100

226 28.6 22.5 33.7

229 20.1 16.6 25.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E


Ion 229.00 (228.70 to 229.70): E

18.43

800000

400000

Time-->









#81
 3,3'-Dichlorobenzidine
 Concen: 40.55 ng
 RT: 18.42 min Scan# 1473
 Delta R.T. -0.03 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.7	50.7	76.1
126	15.9	12.7	19.1




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#82
 Chrysene
 Concen: 40.30 ng
 RT: 18.49 min Scan# 1479
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.4	36.6
229	19.8	15.8	23.8



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 226.00 (225.70 to 226.70): E
 Ion 229.00 (228.70 to 229.70): E





#83

Bis(2-ethylhexyl)phthalate

Concen: 41.75 ng

RT: 18.51 min Scan# 1481

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 967681

Ion Ratio Lower Upper

149 100

167 26.1 21.8 32.6




279 3.6 2.6 4.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 38.71 ng

RT: 19.28 min Scan# 1548

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:149 Resp: 1368085

Ion Ratio Lower Upper

149 100

167 1.5 1.2 1.8



43 17.0 8.2 12.4#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 41.54 ng
 RT: 21.34 min Scan# 1728
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.0	25.6	38.4
277	25.9	20.5	30.7



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 20.11 min Scan# 1621
 Delta R.T. -0.02 min
 Lab File: BF087462.D
 Acq: 28 May 2016 2:04

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.2	19.5	29.3
265	22.5	17.3	25.9





#87

Benzo(b)fluoranthene

Concen: 73.30 ng

RT: 19.71 min Scan# 1586

Delta R.T. -0.01 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:252 Resp: 1837883

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6




125 9.0 11.4 17.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 85.21 ng

RT: 19.71 min Scan# 1586

Delta R.T. -0.05 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:252 Resp: 1838997

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6



125 9.0 9.1 13.7#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 39.62 ng

RT: 20.06 min Scan# 1616

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:252 Resp: 859095

Ion Ratio Lower Upper

252 100

253 22.8 17.2 25.8




125 15.0 9.0 13.6#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 42.39 ng

RT: 21.34 min Scan# 1728

Delta R.T. -0.02 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04


Tgt Ion:278 Resp: 783975

Ion Ratio Lower Upper

278 100

139 18.6 16.6 24.8



279 24.0 19.6 29.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 42.60 ng

RT: 21.68 min Scan# 1758

Delta R.T. -0.03 min

Lab File: BF087462.D

Acq: 28 May 2016 2:04

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 276 Resp: 777397

Ion Ratio Lower Upper

276 100

277 24.3 19.6 29.4

138 23.4 23.6 35.4#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

