

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052716\
 Data File : BF087474.D
 Acq On : 28 May 2016 8:54
 Operator : UM/SJ
 Sample : H3258-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-1

Manual Integrations
 APPROVED

umangi
 5/31/2016 7:02:14 PM

Quant Time: May 30 01:17:16 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 30 00:59:34 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	128014	20.00	ng	0.00
21) Naphthalene-d8	9.51	136	512898	20.00	ng	-0.01
38) Acenaphthene-d10	12.34	164	218129	20.00	ng	-0.01
63) Phenanthrene-d10	14.74	188	338153	20.00	ng	-0.01
75) Chrysene-d12	18.45	240	246089	20.00	ng	-0.01
86) Perylene-d12	20.11	264	187682	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	864735	118.70	ng	0.01
7) Phenol-d6	7.04	99	1172259	119.08	ng	-0.01
23) Nitrobenzene-d5	8.40	82	753176	74.01	ng	-0.02
41) 2,4,6-Tribromophenol	13.65	330	211977	101.13	ng	0.00
44) 2-Fluorobiphenyl	11.29	172	1190858	76.97	ng	-0.01
78) Terphenyl-d14	17.10	244	729256	63.00	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	11.99	163	221427	12.67	ng	99
59) Diethylphthalate	13.13	149	35623	2.10	ng	98
70) Phenanthrene	14.78	178	57050	3.05	ng	99
74) Fluoranthene	16.55	202	70907	3.78	ng	93
77) Pyrene	16.85	202	62990	3.56	ng	96
80) Benzo(a)anthracene	18.42	228	34485m	2.46	ng	
82) Chrysene	18.47	228	31715	2.28	ng	98
87) Benzo(b)fluoranthene	19.71	252	31459m	2.63	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052716\
Data File : BF087474.D
Acq On : 28 May 2016 8:54
Operator : UM/SJ
Sample : H3258-01
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-1

Manual Integrations
APPROVED

umangi
5/31/2016 7:02:14 PM

Quant Time: May 30 01:17:16 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 30 00:59:34 2016
Response via : Initial Calibration

