



Data Path : Z:\HPCHEM1\BNA F\Data\BF052717\
Data File : BF095525.D
Acq On : 28 May 2017 7:43
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 30 00:50:18 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration



TIC: BF095525.D

(69) Pentachlorophenol (C)



14.801min (+0.047) 0.00ng m

response 17389

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	65.07
264.00	64.60	71.66
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF052717\
Data File : BF095525.D
Acq On : 28 May 2017 7:43
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 30 00:50:18 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration



TIC: BF095525.D

(76) Benzidine

17.018min (+0.012) 0.00ng

response 163760

Ion	Exp%	Act%
184.00	100	100
185.00	19.40	14.08#
183.00	12.20	12.19
0.00	0.00	0.00