


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052717\
 Data File : BF095525.D
 Acq On : 28 May 2017 7:43
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri May 26 16:50:16 2017



AutoFind: Scans 2310, 2311, 2312; Background Corrected with Scan 2304

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.6	12300	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.7	12348	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.9	17037	PASS
197	198	0.00	2	0.4	155	PASS
198	198	100	100	100.0	35536	PASS
199	198	5	9	7.0	2470	PASS
275	198	10	60	24.4	8668	PASS
365	198	1	100	2.0	723	PASS
441	442	0.01	24	16.0	5022	PASS
442	198	50	100	88.3	31392	PASS
443	442	15	24	18.6	5854	PASS