




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052722\
Data File : BF128513.D
Acq On : 27 May 2022 16:49
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 28 05:59:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 27 01:51:59 2022
Response via : Initial Calibration



TIC: BF128513.D\data.ms

(70) Pentachlorophenol (C)




11.163min (-0.006) 263935.14 ng

response 429409

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.70	62.22
264.00	61.80	61.91
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052722\
 Data File : BF128513.D
 Acq On : 27 May 2022 16:49
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 28 05:59:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration



TIC: BF128513.D\data.ms

(77) Benzidine

12.710min (+ 0.006) 61553.08 ng

response 2048711

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.90	14.55
183.00	11.60	12.18
0.00	0.00	0.00