

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052722\
 Data File : BF128499.D
 Acq On : 27 May 2022 09:38
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 27 23:17:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	39.887	0.3	103	0.00
3	Pyridine	40.000	40.307	-0.8	105	0.00
4	n-Nitrosodimethylamine	40.000	40.229	-0.6	102	0.00
5 S	2-Fluorophenol	80.000	76.796	4.0	102	0.00
6	Aniline	40.000	40.244	-0.6	101	0.00
7 S	Phenol-d6	80.000	76.823	4.0	102	0.00
8	2-Chlorophenol	40.000	38.825	2.9	103	0.00
9	Benzaldehyde	40.000	49.612	-24.0	116	0.00
10 C	Phenol	40.000	38.197	4.5	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.374	4.1	103	0.00
12	1,3-Dichlorobenzene	40.000	39.402	1.5	104	0.00
13 C	1,4-Dichlorobenzene	40.000	38.919	2.7	103	0.00
14	1,2-Dichlorobenzene	40.000	38.268	4.3	102	0.00
15	Benzyl Alcohol	40.000	37.000	7.5	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.393	4.0	101	0.00
17	2-Methylphenol	40.000	38.612	3.5	101	0.00
18	Hexachloroethane	40.000	39.735	0.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.176	7.1	101	0.00
20	3+4-Methylphenols	40.000	37.730	5.7	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	38.080	4.8	102	0.00
23 S	Nitrobenzene-d5	80.000	84.553	-5.7	107	0.00
24	Nitrobenzene	40.000	41.071	-2.7	105	0.00
25	Isophorone	40.000	38.550	3.6	101	0.00
26 C	2-Nitrophenol	40.000	46.375	-15.9	112	0.00
27	2,4-Dimethylphenol	40.000	40.382	-1.0	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.416	4.0	102	0.00
29 C	2,4-Dichlorophenol	40.000	39.368	1.6	103	0.00
30	1,2,4-Trichlorobenzene	40.000	39.360	1.6	104	0.00
31	Naphthalene	40.000	38.448	3.9	103	0.00
32	Benzoic acid	40.000	47.117	-17.8	112	0.00
33	4-Chloroaniline	40.000	39.103	2.2	102	0.00
34 C	Hexachlorobutadiene	40.000	39.524	1.2	103	0.00
35	Caprolactam	40.000	39.454	1.4	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.277	1.8	103	0.00
37	2-Methylnaphthalene	40.000	39.009	2.5	104	0.00
38	1-Methylnaphthalene	40.000	38.695	3.3	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.291	1.8	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.822	-2.1	102	0.00
42 S	2,4,6-Tribromophenol	80.000	78.043	2.4	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.763	8.1	96	0.00
44	2,4,5-Trichlorophenol	40.000	40.338	-0.8	105	0.00
45 S	2-Fluorobiphenyl	80.000	74.469	6.9	103	0.00
46	1,1'-Biphenyl	40.000	37.968	5.1	103	0.00
47	2-Chloronaphthalene	40.000	38.520	3.7	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052722\
 Data File : BF128499.D
 Acq On : 27 May 2022 09:38
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 27 23:17:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.759	-14.4	112	0.00
49	Acenaphthylene	40.000	38.677	3.3	104	0.00
50	Dimethylphthalate	40.000	38.440	3.9	103	0.00
51	2,6-Dinitrotoluene	40.000	44.904	-12.3	110	0.00
52 C	Acenaphthene	40.000	39.079	2.3	104	0.00
53	3-Nitroaniline	40.000	44.007	-10.0	109	0.00
54 P	2,4-Dinitrophenol	40.000	41.359	-3.4	114	0.00
55	Dibenzofuran	40.000	38.983	2.5	103	0.00
56 P	4-Nitrophenol	40.000	40.928	-2.3	99	0.00
57	2,4-Dinitrotoluene	40.000	48.484	-21.2	113	0.00
58	Fluorene	40.000	37.444	6.4	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.652	3.4	101	0.00
60	Diethylphthalate	40.000	39.188	2.0	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.052	4.9	104	0.00
62	4-Nitroaniline	40.000	44.524	-11.3	108	0.00
63	Azobenzene	40.000	38.653	3.4	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.794	-7.0	114	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.037	2.4	102	0.00
67	4-Bromophenyl-phenylether	40.000	39.537	1.2	103	0.00
68	Hexachlorobenzene	40.000	39.464	1.3	103	0.00
69	Atrazine	40.000	39.306	1.7	102	0.00
70 C	Pentachlorophenol	40.000	39.169	2.1	95	0.00
71	Phenanthrene	40.000	38.641	3.4	102	0.00
72	Anthracene	40.000	38.625	3.4	103	0.00
73	Carbazole	40.000	38.461	3.8	101	0.00
74	Di-n-butylphthalate	40.000	38.779	3.1	101	0.00
75 C	Fluoranthene	40.000	38.393	4.0	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	43.796	-9.5	107	0.00
78	Pyrene	40.000	39.353	1.6	104	0.00
79 S	Terphenyl-d14	80.000	76.801	4.0	103	0.00
80	Butylbenzylphthalate	40.000	40.124	-0.3	102	0.00
81	Benzo(a)anthracene	40.000	38.057	4.9	104	0.00
82	3,3'-Dichlorobenzidine	40.000	40.132	-0.3	105	0.00
83	Chrysene	40.000	38.641	3.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.766	0.6	104	0.00
85 c	Di-n-octyl phthalate	40.000	39.917	0.2	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.525	-1.3	105	0.00
88	Benzo(b)fluoranthene	40.000	41.596	-4.0	103	0.00
89	Benzo(k)fluoranthene	40.000	38.777	3.1	103	0.00
90 C	Benzo(a)pyrene	40.000	39.346	1.6	102	0.00
91	Dibenzo(a,h)anthracene	40.000	39.906	0.2	104	0.00
92	Benzo(g,h,i)perylene	40.000	39.080	2.3	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052722\
Data File : BF128499.D
Acq On : 27 May 2022 09:38
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 27 23:17:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 27 01:51:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0