

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052725\
 Data File : BF142552.D
 Acq On : 27 May 2025 10:15
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 27 11:31:47 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 20 16:26:47 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.475	0.460	3.2	110	0.00
3	Pyridine	1.210	1.204	0.5	113	0.00
4	n-Nitrosodimethylamine	0.628	0.600	4.5	106	-0.02
5 S	2-Fluorophenol	1.187	1.141	3.9	111	0.00
6	Aniline	1.921	1.879	2.2	111	0.00
7 S	Phenol-d6	1.429	1.377	3.6	110	-0.01
8	2-Chlorophenol	1.293	1.246	3.6	109	0.00
9	Benzaldehyde	0.859	0.837	2.6	112	0.00
10 C	Phenol	1.608	1.566	2.6	112	-0.01
11	bis(2-Chloroethyl)ether	1.157	1.109	4.1	110	0.00
12	1,3-Dichlorobenzene	1.443	1.369	5.1	108	0.00
13 C	1,4-Dichlorobenzene	1.453	1.383	4.8	108	0.00
14	1,2-Dichlorobenzene	1.391	1.321	5.0	109	0.00
15	Benzyl Alcohol	1.056	1.025	2.9	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.944	1.867	4.0	109	-0.01
17	2-Methylphenol	1.010	0.977	3.3	110	0.00
18	Hexachloroethane	0.507	0.472	6.9	106	0.00
19 P	n-Nitroso-di-n-propylamine	0.886	0.841	5.1	109	-0.02
20	3+4-Methylphenols	1.293	1.236	4.4	109	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.443	0.418	5.6	108	0.00
23 S	Nitrobenzene-d5	0.367	0.350	4.6	109	-0.01
24	Nitrobenzene	0.331	0.320	3.3	110	-0.01
25	Isophorone	0.613	0.588	4.1	110	-0.01
26 C	2-Nitrophenol	0.177	0.173	2.3	110	0.00
27	2,4-Dimethylphenol	0.312	0.298	4.5	109	0.00
28	bis(2-Chloroethoxy)methane	0.383	0.363	5.2	111	-0.01
29 C	2,4-Dichlorophenol	0.283	0.277	2.1	111	0.00
30	1,2,4-Trichlorobenzene	0.308	0.292	5.2	110	0.00
31	Naphthalene	0.985	0.929	5.7	109	0.00
32	Benzoic acid	0.190	0.190	0.0	112	-0.04
33	4-Chloroaniline	0.399	0.389	2.5	111	0.00
34 C	Hexachlorobutadiene	0.192	0.181	5.7	107	0.00
35	Caprolactam	0.080	0.086	-7.5	122	-0.02
36 C	4-Chloro-3-methylphenol	0.291	0.283	2.7	111	-0.01
37	2-Methylnaphthalene	0.620	0.582	6.1	109	0.00
38	1-Methylnaphthalene	0.641	0.609	5.0	110	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	0.574	0.532	7.3	108	0.00
41 P	Hexachlorocyclopentadiene	0.387	0.355	8.3	101	0.00
42 S	2,4,6-Tribromophenol	0.222	0.219	1.4	117	-0.01
43 C	2,4,6-Trichlorophenol	0.387	0.368	4.9	111	0.00
44	2,4,5-Trichlorophenol	0.408	0.389	4.7	111	0.00
45 S	2-Fluorobiphenyl	1.490	1.346	9.7	109	-0.01
46	1,1'-Biphenyl	1.552	1.446	6.8	110	0.00
47	2-Chloronaphthalene	1.146	1.073	6.4	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052725\
 Data File : BF142552.D
 Acq On : 27 May 2025 10:15
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 27 11:31:47 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 20 16:26:47 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.331	0.332	-0.3	115	0.00
49	Acenaphthylene	1.936	1.837	5.1	112	0.00
50	Dimethylphthalate	1.320	1.258	4.7	113	-0.01
51	2,6-Dinitrotoluene	0.285	0.277	2.8	111	0.00
52 C	Acenaphthene	1.182	1.108	6.3	111	0.00
53	3-Nitroaniline	0.311	0.314	-1.0	118	0.00
54 P	2,4-Dinitrophenol	0.147	0.154	-4.8	116	0.00
55	Dibenzofuran	1.701	1.614	5.1	113	0.00
56 P	4-Nitrophenol	0.230	0.236	-2.6	117	0.00
57	2,4-Dinitrotoluene	0.372	0.384	-3.2	119	0.00
58	Fluorene	1.314	1.253	4.6	114	-0.01
59	2,3,4,6-Tetrachlorophenol	0.341	0.333	2.3	113	0.00
60	Diethylphthalate	1.289	1.240	3.8	113	-0.01
61	4-Chlorophenyl-phenylether	0.646	0.611	5.4	113	0.00
62	4-Nitroaniline	0.282	0.312	-10.6	127	-0.01
63	Azobenzene	1.158	1.120	3.3	114	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.116	-3.6	120	0.00
66 c	n-Nitrosodiphenylamine	0.684	0.625	8.6	115	0.00
67	4-Bromophenyl-phenylether	0.236	0.216	8.5	117	0.00
68	Hexachlorobenzene	0.262	0.242	7.6	116	0.00
69	Atrazine	0.182	0.185	-1.6	125	0.00
70 C	Pentachlorophenol	0.148	0.143	3.4	116	0.00
71	Phenanthrene	1.069	1.009	5.6	120	0.00
72	Anthracene	1.091	1.027	5.9	119	0.00
73	Carbazole	0.933	0.922	1.2	124	0.00
74	Di-n-butylphthalate	1.021	1.036	-1.5	127	0.00
75 C	Fluoranthene	0.999	1.034	-3.5	132	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	142	0.00
77	Benزيدine	0.744	0.818	-9.9	145	0.00
78	Pyrene	1.866	1.741	6.7	133	0.00
79 S	Terphenyl-d14	1.464	1.339	8.5	133	0.00
80	Butylbenzylphthalate	0.516	0.552	-7.0	149	0.00
81	Benzo(a)anthracene	1.336	1.296	3.0	144	0.00
82	3,3'-Dichlorobenzidine	0.405	0.399	1.5	139	0.00
83	Chrysene	1.200	1.105	7.9	137	0.00
84	Bis(2-ethylhexyl)phthalate	0.660	0.646	2.1	136	0.00
85 c	Di-n-octyl phthalate	1.290	1.042	19.2	117	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	1.501	1.390	7.4	109	-0.02
88	Benzo(b)fluoranthene	1.184	1.236	-4.4	125	0.00
89	Benzo(k)fluoranthene	1.109	0.974	12.2	106	-0.01
90 C	Benzo(a)pyrene	1.116	1.070	4.1	112	-0.01
91	Dibenzo(a,h)anthracene	1.218	1.127	7.5	108	-0.02
92	Benzo(g,h,i)perylene	1.218	1.132	7.1	109	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052725\
Data File : BF142552.D
Acq On : 27 May 2025 10:15
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 27 11:31:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0