

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052816\
 Data File : BF087495.D
 Acq On : 29 May 2016 00:03
 Operator : UM/SJ
 Sample : H3276-08
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-06

Quant Time: May 30 04:04:21 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 30 03:48:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.46	152	110900	20.00	ng	-0.01
21) Naphthalene-d8	9.50	136	452508	20.00	ng	-0.01
38) Acenaphthene-d10	12.32	164	204842	20.00	ng	-0.01
63) Phenanthrene-d10	14.73	188	381087	20.00	ng	0.00
75) Chrysene-d12	18.43	240	251981	20.00	ng	0.00
86) Perylene-d12	20.10	264	218805	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	393493	62.35	ng	0.01
7) Phenol-d6	7.03	99	352476	41.33	ng	-0.01
23) Nitrobenzene-d5	8.39	82	781193	87.01	ng	-0.01
41) 2,4,6-Tribromophenol	13.63	330	285161	144.86	ng	0.00
44) 2-Fluorobiphenyl	11.28	172	1268303	87.29	ng	-0.01
78) Terphenyl-d14	17.09	244	922770	77.86	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052816\
Data File : BF087495.D
Acq On : 29 May 2016 00:03
Operator : UM/SJ
Sample : H3276-08
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-06

Quant Time: May 30 04:04:21 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 30 03:48:52 2016
Response via : Initial Calibration

