

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052816\
 Data File : BF087517.D
 Acq On : 29 May 2016 12:50
 Operator : UM/SJ
 Sample : H3222-06
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-36

Quant Time: May 30 05:00:14 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 30 04:10:04 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	104189	20.00	ng	0.00
21) Naphthalene-d8	9.50	136	412699	20.00	ng	-0.01
38) Acenaphthene-d10	12.33	164	238771	20.00	ng	0.00
63) Phenanthrene-d10	14.74	188	351612	20.00	ng	0.01
75) Chrysene-d12	18.43	240	260323	20.00	ng	0.00
86) Perylene-d12	20.11	264	205566	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	400697	67.58	ng	0.02
7) Phenol-d6	7.05	99	442870	55.28	ng	0.01
23) Nitrobenzene-d5	8.39	82	592125	72.31	ng	-0.01
41) 2,4,6-Tribromophenol	13.66	330	210928	91.93	ng	0.02
44) 2-Fluorobiphenyl	11.28	172	951120	56.16	ng	-0.01
78) Terphenyl-d14	17.09	244	747203	61.02	ng	0.00
Target Compounds						
31) Naphthalene	9.54	128	51992	2.51	ng	# 85
49) Dimethylphthalate	12.00	163	62595	3.27	ng	# 68
59) Diethylphthalate	13.15	149	68724	3.70	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF052816\
Data File : BF087517.D
Acq On : 29 May 2016 12:50
Operator : UM/SJ
Sample : H3222-06
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-36

Quant Time: May 30 05:00:14 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 30 04:10:04 2016
Response via : Initial Calibration

