

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052916\
 Data File : BF087540.D
 Acq On : 30 May 2016 7:24
 Operator : UM/SJ
 Sample : H3303-03
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B3-12

Manual Integrations
 APPROVED

umangi
 5/31/2016 7:40:29 PM

Quant Time: May 31 13:07:13 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 31 00:45:12 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	70917	20.00	ng	0.02
21) Naphthalene-d8	9.54	136	274815	20.00	ng	0.03
38) Acenaphthene-d10	12.50	164	155951m	20.00	ng	0.17
63) Phenanthrene-d10	14.86	188	155623	20.00	ng	0.11
75) Chrysene-d12	18.46	240	243435	20.00	ng	0.01
86) Perylene-d12	20.14	264	202398	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	562510	139.38	ng	0.00
7) Phenol-d6	7.03	99	763491	140.00	ng	-0.01
23) Nitrobenzene-d5	8.41	82	720743	132.18	ng	0.00
41) 2,4,6-Tribromophenol	13.78	330	98433	65.68	ng	0.15
44) 2-Fluorobiphenyl	11.43	172	496420	44.88	ng	0.14
78) Terphenyl-d14	17.12	244	638600	55.77	ng	0.02
Target Compounds						
31) Naphthalene	9.61	128	3619670	262.85	ng	# 79
37) 2-Methylnaphthalene	10.87	142	8313487	966.36	ng	# 94
45) 1,1'-Biphenyl	11.60	154	892487	67.97	ng	88
57) Fluorene	13.38	166	730015m	63.61	ng	
70) Phenanthrene	14.89	178	2212544	256.67	ng	# 95
74) Fluoranthene	16.61	202	92269	10.67	ng	74
77) Pyrene	16.89	202	164331m	9.38	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF052916\
Data File : BF087540.D
Acq On : 30 May 2016 7:24
Operator : UM/SJ
Sample : H3303-03
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B3-12

Manual Integrations
APPROVED

umangi
5/31/2016 7:40:29 PM

Quant Time: May 31 13:07:13 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF052316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 31 00:45:12 2016
Response via : Initial Calibration

