

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052920\
 Data File : BF120456.D
 Acq On : 29 May 2020 23:02
 Operator : MA/SJ
 Sample : L2792-10
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NM96-COMP-2020-0-6

Manual Integrations
 APPROVED

mohammad
 6/1/2020 12:21:49 PM

Quant Time: May 30 05:06:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 12:39:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	276483	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	975732	20.00	ng	0.00
39) Acenaphthene-d10	9.87	164	444747	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	683577	20.00	ng	0.00
76) Chrysene-d12	13.99	240	545642	20.00	ng	0.00
86) Perylene-d12	15.44	264	548891	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	1529799	105.45	ng	0.00
7) Phenol-d6	6.47	99	1921638	107.46	ng	0.00
23) Nitrobenzene-d5	7.40	82	1240561	83.17	ng	0.00
42) 2,4,6-Tribromophenol	10.66	330	438929	106.52	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	2127236	81.15	ng	0.00
79) Terphenyl-d14	12.94	244	1820788	74.26	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.59	163	277467	9.979	ng	99
71) Phenanthrene	11.38	178	297499	9.877	ng	98
72) Anthracene	11.43	178	86482	2.862	ng	96
75) Fluoranthene	12.57	202	468264	14.448	ng	97
78) Pyrene	12.80	202	429725	11.325	ng	98
81) Benzo(a)anthracene	13.98	228	233453	7.983	ng	99
83) Chrysene	14.02	228	217903	7.108	ng	96
87) Indeno(1,2,3-cd)pyrene	16.88	276	111059	3.699	ng	94
88) Benzo(b)fluoranthene	15.03	252	297933m	9.772	ng	
89) Benzo(k)fluoranthene	15.05	252	69711m	2.449	ng	
90) Benzo(a)pyrene	15.39	252	192051	7.202	ng	99
92) Benzo(a,h,i)perylene	17.32	276	104576	4.332	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF052920\
Data File : BF120456.D
Acq On : 29 May 2020 23:02
Operator : MA/SJ
Sample : L2792-10
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NM96-COMP-2020-0-6

Manual Integrations
APPROVED

mohammad
6/1/2020 12:21:49 PM

Quant Time: May 30 05:06:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF052820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 29 12:39:43 2020
Response via : Initial Calibration

