

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052922\
 Data File : BF128563.D
 Acq On : 29 May 2022 11:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 30 00:49:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.538	0.514	4.5	66	-0.03
3	Pyridine	1.394	1.332	4.4	67	-0.02
4	n-Nitrosodimethylamine	0.730	0.746	-2.2	70	-0.02
5 S	2-Fluorophenol	1.207	1.173	2.8	69	0.00
6	Aniline	1.796	1.713	4.6	64	0.00
7 S	Phenol-d6	1.540	1.482	3.8	69	0.00
8	2-Chlorophenol	1.252	1.234	1.4	70	0.00
9	Benzaldehyde	0.745	0.793	-6.4	79	0.00
10 C	Phenol	1.526	1.504	1.4	71	0.00
11	bis(2-Chloroethyl)ether	1.249	1.177	5.8	68	0.00
12	1,3-Dichlorobenzene	1.422	1.417	0.4	71	0.00
13 C	1,4-Dichlorobenzene	1.439	1.413	1.8	70	0.00
14	1,2-Dichlorobenzene	1.324	1.309	1.1	71	0.00
15	Benzyl Alcohol	1.174	1.171	0.3	70	0.00
16	2,2'-oxybis(1-Chloropropane	2.210	2.050	7.2	65	0.00
17	2-Methylphenol	1.066	1.017	4.6	67	0.00
18	Hexachloroethane	0.508	0.520	-2.4	72	0.00
19 P	n-Nitroso-di-n-propylamine	0.944	0.884	6.4	69	0.00
20	3+4-Methylphenols	1.287	1.245	3.3	70	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
22	Acetophenone	0.461	0.460	0.2	72	0.00
23 S	Nitrobenzene-d5	0.347	0.387	-11.5	75	0.00
24	Nitrobenzene	0.343	0.367	-7.0	73	0.00
25	Isophorone	0.640	0.620	3.1	68	0.00
26 C	2-Nitrophenol	0.143	0.170	-18.9	77	0.00
27	2,4-Dimethylphenol	0.285	0.281	1.4	70	0.00
28	bis(2-Chloroethoxy)methane	0.415	0.394	5.1	67	0.00
29 C	2,4-Dichlorophenol	0.295	0.297	-0.7	70	0.00
30	1,2,4-Trichlorobenzene	0.319	0.321	-0.6	71	0.00
31	Naphthalene	0.979	0.964	1.5	70	0.00
32	Benzoic acid	0.201	0.226	-12.4	71	0.00
33	4-Chloroaniline	0.390	0.377	3.3	67	0.00
34 C	Hexachlorobutadiene	0.202	0.217	-7.4	75	0.00
35	Caprolactam	0.091	0.086	5.5	64	0.00
36 C	4-Chloro-3-methylphenol	0.303	0.314	-3.6	72	0.00
37	2-Methylnaphthalene	0.618	0.606	1.9	70	0.00
38	1-Methylnaphthalene	0.601	0.590	1.8	71	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	68	0.00
40	1,2,4,5-Tetrachlorobenzene	0.549	0.572	-4.2	74	0.00
41 P	Hexachlorocyclopentadiene	0.309	0.333	-7.8	71	0.00
42 S	2,4,6-Tribromophenol	0.190	0.211	-11.1	77	0.00
43 C	2,4,6-Trichlorophenol	0.405	0.410	-1.2	70	0.00
44	2,4,5-Trichlorophenol	0.417	0.431	-3.4	71	0.00
45 S	2-Fluorobiphenyl	1.231	1.264	-2.7	75	0.00
46	1,1'-Biphenyl	1.404	1.402	0.1	71	0.00
47	2-Chloronaphthalene	1.102	1.107	-0.5	70	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052922\
 Data File : BF128563.D
 Acq On : 29 May 2022 11:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 30 00:49:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.307	0.374	-21.8	79	0.00
49	Acenaphthylene	1.694	1.692	0.1	71	0.00
50	Dimethylphthalate	1.297	1.304	-0.5	71	0.00
51	2,6-Dinitrotoluene	0.238	0.274	-15.1	75	0.00
52 C	Acenaphthene	1.045	1.085	-3.8	73	0.00
53	3-Nitroaniline	0.275	0.301	-9.5	72	0.00
54 P	2,4-Dinitrophenol	0.100	0.134	-34.0#	89	0.00
55	Dibenzofuran	1.550	1.559	-0.6	71	0.00
56 P	4-Nitrophenol	0.223	0.238	-6.7	68	0.00
57	2,4-Dinitrotoluene	0.284	0.363	-27.8#	79	0.00
58	Fluorene	1.116	1.138	-2.0	74	0.00
59	2,3,4,6-Tetrachlorophenol	0.338	0.352	-4.1	72	0.00
60	Diethylphthalate	1.260	1.309	-3.9	72	0.00
61	4-Chlorophenyl-phenylether	0.566	0.593	-4.8	76	0.00
62	4-Nitroaniline	0.263	0.299	-13.7	73	0.00
63	Azobenzene	1.283	1.294	-0.9	71	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	0.083	0.107	-28.9#	84	0.00
66 c	n-Nitrosodiphenylamine	0.599	0.604	-0.8	71	0.00
67	4-Bromophenyl-phenylether	0.207	0.217	-4.8	73	0.00
68	Hexachlorobenzene	0.228	0.240	-5.3	73	0.00
69	Atrazine	0.186	0.182	2.2	68	0.00
70 C	Pentachlorophenol	0.139	0.144	-3.6	68	0.00
71	Phenanthrene	0.947	0.944	0.3	71	0.00
72	Anthracene	0.946	0.941	0.5	71	0.00
73	Carbazole	0.907	0.878	3.2	68	0.00
74	Di-n-butylphthalate	1.067	1.087	-1.9	71	0.00
75 C	Fluoranthene	1.013	0.967	4.5	67	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
77	Benzydine	0.461	0.445	3.5	60	0.00
78	Pyrene	1.507	1.523	-1.1	68	0.00
79 S	Terphenyl-d14	1.035	1.089	-5.2	72	0.00
80	Butylbenzylphthalate	0.636	0.647	-1.7	66	0.00
81	Benzo(a)anthracene	1.189	1.160	2.4	68	0.00
82	3,3'-Dichlorobenzidine	0.304	0.306	-0.7	67	0.00
83	Chrysene	1.217	1.180	3.0	65	0.00
84	Bis(2-ethylhexyl)phthalate	0.769	0.828	-7.7	72	0.00
85 c	Di-n-octyl phthalate	1.353	1.340	1.0	65	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	66	0.00
87	Indeno(1,2,3-cd)pyrene	1.180	1.309	-10.9	73	0.00
88	Benzo(b)fluoranthene	1.244	1.183	4.9	60	0.00
89	Benzo(k)fluoranthene	1.153	1.176	-2.0	70	0.00
90 C	Benzo(a)pyrene	0.998	0.993	0.5	66	0.00
91	Dibenzo(a,h)anthracene	0.978	1.090	-11.5	74	0.00
92	Benzo(g,h,i)perylene	0.970	1.093	-12.7	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF052922\
Data File : BF128563.D
Acq On : 29 May 2022 11:07
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 30 00:49:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 27 01:51:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0