

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF053015\
 Data File : BF079617.D
 Acq On : 30 May 2015 22:06
 Operator : TP/IZ
 Sample : G2420-28 5X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PENRD8.14(5)

Quant Time: Jun 01 03:11:45 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 01 02:59:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	65278	20.00	ng	0.00
21) Naphthalene-d8	8.51	136	266189	20.00	ng	0.00
38) Acenaphthene-d10	10.67	164	130653	20.00	ng	0.00
63) Phenanthrene-d10	12.51	188	241333	20.00	ng	0.00
75) Chrysene-d12	15.78	240	255684	20.00	ng	0.00
86) Perylene-d12	17.46	264	256729	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.21	112	44351	11.78	ng	0.01
7) Phenol-d6	6.55	99	80243	16.73	ng	0.01
23) Nitrobenzene-d5	7.63	82	40055	8.70	ng	0.00
41) 2,4,6-Tribromophenol	11.66	330	29274	20.40	ng	0.00
44) 2-Fluorobiphenyl	9.86	172	117686	12.85	ng	0.00
78) Terphenyl-d14	14.50	244	149965	13.76	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF053015\
Data File : BF079617.D
Acq On : 30 May 2015 22:06
Operator : TP/IZ
Sample : G2420-28 5X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PENRD8.14(5)

Quant Time: Jun 01 03:11:45 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF052615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 01 02:59:43 2015
Response via : Initial Calibration

