

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF053017\
 Data File : BF095549.D
 Acq On : 30 May 2017 22:33
 Operator : SJ/MA
 Sample : I3302-07
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NM12-COMP-0-6

Manual Integrations
 APPROVED

mohammad
 5/31/2017 5:02:59 PM

Quant Time: May 31 04:45:30 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 26 16:50:16 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	87762	20.00	ng	0.00
21) Naphthalene-d8	9.77	136	337800	20.00	ng	0.00
38) Acenaphthene-d10	12.60	164	135225	20.00	ng	0.00
63) Phenanthrene-d10	15.01	188	204517	20.00	ng	0.00
75) Chrysene-d12	18.68	240	201458	20.00	ng	0.00
86) Perylene-d12	20.35	264	131610	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.71	112	568256	107.95	ng	0.00
7) Phenol-d6	7.30	99	679721	107.62	ng	0.00
23) Nitrobenzene-d5	8.65	82	362682	67.70	ng	0.00
41) 2,4,6-Tribromophenol	13.91	330	95576	86.30	ng	0.00
44) 2-Fluorobiphenyl	11.54	172	600433	63.91	ng	0.00
78) Terphenyl-d14	17.31	244	393781	43.05	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	12.24	163	93216	8.40	ng	96
74) Fluoranthene	16.79	202	59317	4.92	ng	99
77) Pyrene	17.09	202	61782	4.10	ng	98
80) Benzo(a)anthracene	18.67	228	28152m	2.40	ng	
82) Chrysene	18.71	228	33080	2.92	ng	98
87) Benzo(b)fluoranthene	19.95	252	30634m	3.77	ng	
89) Benzo(a)pyrene	20.30	252	20774	2.77	ng	# 89
91) Benzo(g,h,i)perylene	22.10	276	12240	2.01	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF053017\
Data File : BF095549.D
Acq On : 30 May 2017 22:33
Operator : SJ/MA
Sample : I3302-07
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NM12-COMP-0-6

Manual Integrations
APPROVED

mohammad
5/31/2017 5:02:59 PM

Quant Time: May 31 04:45:30 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 26 16:50:16 2017
Response via : Initial Calibration

