

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053025\
 Data File : BF142577.D
 Acq On : 30 May 2025 18:46
 Operator : RC/JU
 Sample : Q2146-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP04-MHN-WC

Quant Time: May 30 22:28:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 20 16:26:47 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.898	152	121806	20.000	ng	0.00
21) Naphthalene-d8	8.181	136	458880	20.000	ng	0.00
39) Acenaphthene-d10	9.934	164	239948	20.000	ng	-0.01
64) Phenanthrene-d10	11.422	188	386203	20.000	ng	-0.01
76) Chrysene-d12	14.063	240	222868	20.000	ng	-0.01
86) Perylene-d12	15.563	264	239278	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	804897	111.329	ng	0.01
7) Phenol-d6	6.522	99	915526	105.196	ng	-0.01
23) Nitrobenzene-d5	7.457	82	631906	75.090	ng	-0.02
42) 2,4,6-Tribromophenol	10.728	330	307832	115.591	ng	-0.01
45) 2-Fluorobiphenyl	9.257	172	1309254	73.217	ng	-0.01
79) Terphenyl-d14	13.010	244	1177240	72.184	ng	0.00
Target Compounds						
32) Benzoic acid	7.893	122	29286	6.727	ng	95
50) Dimethylphthalate	9.645	163	78078	4.931	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053025\
Data File : BF142577.D
Acq On : 30 May 2025 18:46
Operator : RC/JU
Sample : Q2146-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP04-MHN-WC

Quant Time: May 30 22:28:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration

