


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF053117\
 Data File : BF095584.D
 Acq On : 1 Jun 2017 3:04
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF050217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 01 03:03:02 2017



AutoFind: Scans 2299, 2300, 2301; Background Corrected with Scan 2294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.8	18119	PASS
68	69	0.00	2	0.8	135	PASS
69	198	0.00	100	36.6	17555	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.9	23952	PASS
197	198	0.00	2	0.4	180	PASS
198	198	100	100	100.0	47976	PASS
199	198	5	9	6.3	3005	PASS
275	198	10	60	23.0	11036	PASS
365	198	1	100	2.0	983	PASS
441	442	0.01	24	16.3	6502	PASS
442	198	50	100	83.0	39810	PASS
443	442	15	24	19.1	7609	PASS