

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053123\
 Data File : BF133544.D
 Acq On : 31 May 2023 17:49
 Operator : CG\JU
 Sample : 02957-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-9-0A

Quant Time: May 31 18:14:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 30 12:58:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	138464	20.000	ng	0.00
21) Naphthalene-d8	8.110	136	539472	20.000	ng	# 0.00
39) Acenaphthene-d10	9.869	164	270311	20.000	ng	0.00
64) Phenanthrene-d10	11.351	188	535536	20.000	ng	-0.01
76) Chrysene-d12	13.992	240	389526	20.000	ng	-0.01
86) Perylene-d12	15.451	264	316357	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	1028698	123.956	ng	0.01
7) Phenol-d6	6.457	99	1172312	114.752	ng	0.00
23) Nitrobenzene-d5	7.392	82	890093	110.164	ng	-0.01
42) 2,4,6-Tribromophenol	10.657	330	423397	157.134	ng	0.00
45) 2-Fluorobiphenyl	9.192	172	1542169	88.056	ng	0.00
79) Terphenyl-d14	12.951	244	1910831	87.220	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF053123\
Data File : BF133544.D
Acq On : 31 May 2023 17:49
Operator : CG\JU
Sample : 02957-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-9-0A

Quant Time: May 31 18:14:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 30 12:58:40 2023
Response via : Initial Calibration

