


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF053124\  
Data File : BF138083.D  
Acq On : 31 May 2024 12:32  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF053124.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Sat Jun 01 01:21:05 2024






AutoFind: Scans 1652, 1653, 1654; Background Corrected with Scan 1647

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 28.7         | 41136      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.5          | 657        | PASS                |
| 69             | 198             | 0.00            | 100             | 30.0         | 42941      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.4          | 186        | PASS                |
| 127            | 198             | 10              | 80              | 44.1         | 63187      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 143275     | PASS                |
| 199            | 198             | 5               | 9               | 6.6          | 9482       | PASS                |
| 275            | 198             | 10              | 60              | 29.8         | 42755      | PASS                |
| 365            | 198             | 1               | 100             | 4.1          | 5896       | PASS                |
| 441            | 198             | 0.01            | 100             | 15.9         | 22829      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 152179     | PASS                |
| 443            | 442             | 15              | 24              | 18.9         | 28736      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF053124\  
Data File : BF138083.D  
Acq On : 31 May 2024 12:32  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: May 31 13:19:01 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF051424.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed May 15 04:12:16 2024  
Response via : Initial Calibration



TIC: BF138083.D\data.ms

**(70) Pentachlorophenol (C)**

**11.380min (+ 0.006) 31023.60 ng**




**response 129929**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.60  | 63.97  |
| 264.00 | 61.80  | 61.41  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF053124\  
Data File : BF138083.D  
Acq On : 31 May 2024 12:32  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: May 31 13:19:01 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF051424.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed May 15 04:12:16 2024  
Response via : Initial Calibration



TIC: BF138083.D\data.ms

(77) Benzidine

12.927min (+ 0.011) 0.00 ng

response 453556

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.30  | 13.65  |
| 183.00 | 11.90  | 11.22  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 5/31/2024     | BNA_F            | <u>BF138083.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 333842           | 13.739             |
| DDD           | 6357             | 13.468             |
| DDE           | 357              | 13.098             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 6714          | 340556           | 1.97               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP