

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF060120\
 Data File : BF120462.D
 Acq On : 1 Jun 2020 11:17
 Operator : JU/CG
 Sample : PB129165BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB129165BL

Quant Time: Jun 01 12:04:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF052820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 12:39:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	295691	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	1084077	20.00	ng	0.00
39) Acenaphthene-d10	9.87	164	555832	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	1058174	20.00	ng	0.00
76) Chrysene-d12	14.00	240	889824	20.00	ng	0.00
86) Perylene-d12	15.45	264	881879	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	1933850	124.64	ng	0.01
7) Phenol-d6	6.47	99	2479286	129.64	ng	0.00
23) Nitrobenzene-d5	7.40	82	1591344	96.02	ng	0.00
42) 2,4,6-Tribromophenol	10.67	330	738950	143.48	ng	0.00
45) 2-Fluorobiphenyl	9.20	172	2713197	82.82	ng	0.00
79) Terphenyl-d14	12.95	244	3221450	80.56	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF060120\
Data File : BF120462.D
Acq On : 1 Jun 2020 11:17
Operator : JU/CG
Sample : PB129165BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB129165BL

Quant Time: Jun 01 12:04:53 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF052820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 29 12:39:43 2020
Response via : Initial Calibration

