

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF060120\
 Data File : BF120467.D
 Acq On : 1 Jun 2020 13:31
 Operator : JU/CG
 Sample : L2790-11DL 5X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-3-SDL

Quant Time: Jun 01 14:05:31 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF052820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 29 12:39:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	212855	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	778661	20.00	ng	0.00
39) Acenaphthene-d10	9.87	164	367722	20.00	ng	0.00
64) Phenanthrene-d10	11.36	188	599905	20.00	ng	0.00
76) Chrysene-d12	13.99	240	456890	20.00	ng	0.00
86) Perylene-d12	15.45	264	544962	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	213832	19.14	ng	0.00
7) Phenol-d6	6.46	99	280342	20.36	ng	-0.01
23) Nitrobenzene-d5	7.39	82	163469	13.73	ng	-0.01
42) 2,4,6-Tribromophenol	10.66	330	64837	19.03	ng	0.00
45) 2-Fluorobiphenyl	9.19	172	360092	16.61	ng	0.00
79) Terphenyl-d14	12.94	244	311882	15.19	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phthalate	13.98	149	630901	40.52	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF060120\
Data File : BF120467.D
Acq On : 1 Jun 2020 13:31
Operator : JU/CG
Sample : L2790-11DL 5X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-3-SDL

Quant Time: Jun 01 14:05:31 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF052820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 29 12:39:43 2020
Response via : Initial Calibration

