

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060122\
 Data File : BF128617.D
 Acq On : 01 Jun 2022 11:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 04:33:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	54	0.00
2	1,4-Dioxane	40.000	35.181	12.0	48	-0.01
3	Pyridine	40.000	35.716	10.7	50	0.00
4	n-Nitrosodimethylamine	40.000	43.959	-9.9	59	0.00
5 S	2-Fluorophenol	80.000	77.500	3.1	55	0.04
6	Aniline	40.000	35.616	11.0	48	0.00
7 S	Phenol-d6	80.000	74.891	6.4	53	0.05
8	2-Chlorophenol	40.000	38.898	2.8	55	0.02
9	Benzaldehyde	40.000	40.570	-1.4	54	0.00
10 C	Phenol	40.000	42.711	-6.8	61	0.05
11	bis(2-Chloroethyl)ether	40.000	37.738	5.7	54	0.00
12	1,3-Dichlorobenzene	40.000	39.843	0.4	56	0.00
13 C	1,4-Dichlorobenzene	40.000	40.870	-2.2	57	0.00
14	1,2-Dichlorobenzene	40.000	40.272	-0.7	57	0.00
15	Benzyl Alcohol	40.000	41.142	-2.9	57	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	37.902	5.2	53	0.00
17	2-Methylphenol	40.000	38.693	3.3	54	0.05
18	Hexachloroethane	40.000	43.456	-8.6	61	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	41.689	-4.2	60	0.00
20	3+4-Methylphenols	40.000	43.370	-8.4	62	0.04
21 I	Naphthalene-d8	20.000	20.000	0.0	56	0.00
22	Acetophenone	40.000	43.140	-7.9	63	-0.01
23 S	Nitrobenzene-d5	80.000	91.838	-14.8	63	0.00
24	Nitrobenzene	40.000	44.496	-11.2	62	0.00
25	Isophorone	40.000	39.265	1.8	56	0.00
26 C	2-Nitrophenol	40.000	47.357	-18.4	63	0.00
27	2,4-Dimethylphenol	40.000	36.873	7.8	53	0.02
28	bis(2-Chloroethoxy)methane	40.000	38.804	3.0	56	0.00
29 C	2,4-Dichlorophenol	40.000	40.525	-1.3	58	0.03
30	1,2,4-Trichlorobenzene	40.000	41.283	-3.2	59	0.00
31	Naphthalene	40.000	40.457	-1.1	59	0.00
32	Benzoic acid	40.000	30.868	22.8	40	0.04
33	4-Chloroaniline	40.000	39.244	1.9	56	0.00
34 C	Hexachlorobutadiene	40.000	46.342	-15.9	66	0.00
35	Caprolactam	40.000	37.361	6.6	52	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.043	-7.6	61	0.04
37	2-Methylnaphthalene	40.000	41.023	-2.6	60	0.00
38	1-Methylnaphthalene	40.000	40.498	-1.2	59	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	57	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.232	-8.1	64	0.00
41 P	Hexachlorocyclopentadiene	40.000	48.011	-20.0	67	0.00
42 S	2,4,6-Tribromophenol	80.000	95.664	-19.6	69	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.786	-2.0	59	0.01
44	2,4,5-Trichlorophenol	40.000	38.814	3.0	56	0.04
45 S	2-Fluorobiphenyl	80.000	86.213	-7.8	66	0.00
46	1,1'-Biphenyl	40.000	41.197	-3.0	62	0.00
47	2-Chloronaphthalene	40.000	40.998	-2.5	60	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060122\
 Data File : BF128617.D
 Acq On : 01 Jun 2022 11:20
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 04:33:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 27 01:51:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	50.313	-25.8#	68	0.00
49	Acenaphthylene	40.000	39.745	0.6	59	0.00
50	Dimethylphthalate	40.000	41.633	-4.1	62	0.00
51	2,6-Dinitrotoluene	40.000	45.102	-12.8	61	0.00
52 C	Acenaphthene	40.000	41.890	-4.7	62	0.00
53	3-Nitroaniline	40.000	42.941	-7.4	59	0.00
54 P	2,4-Dinitrophenol	40.000	48.604	-21.5	76	0.00
55	Dibenzofuran	40.000	40.271	-0.7	59	-0.01
56 P	4-Nitrophenol	40.000	31.901	20.2	43	0.06
57	2,4-Dinitrotoluene	40.000	50.329	-25.8#	65	0.00
58	Fluorene	40.000	43.000	-7.5	65	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.679	-6.7	62	0.00
60	Diethylphthalate	40.000	43.484	-8.7	63	-0.01
61	4-Chlorophenyl-phenylether	40.000	44.528	-11.3	67	-0.01
62	4-Nitroaniline	40.000	46.772	-16.9	63	0.00
63	Azobenzene	40.000	41.703	-4.3	62	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	58	0.00
65	4,6-Dinitro-2-methylphenol	40.000	50.548	-26.4#	79	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.864	-2.2	61	0.00
67	4-Bromophenyl-phenylether	40.000	43.479	-8.7	65	0.00
68	Hexachlorobenzene	40.000	43.336	-8.3	65	0.00
69	Atrazine	40.000	43.093	-7.7	64	0.00
70 C	Pentachlorophenol	40.000	33.378	16.6	47	0.00
71	Phenanthrene	40.000	40.676	-1.7	62	0.00
72	Anthracene	40.000	41.170	-2.9	63	0.00
73	Carbazole	40.000	39.671	0.8	60	0.00
74	Di-n-butylphthalate	40.000	42.240	-5.6	63	0.00
75 C	Fluoranthene	40.000	40.293	-0.7	61	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	59	0.00
77	Benzidine	40.000	36.889	7.8	51	0.00
78	Pyrene	40.000	41.472	-3.7	63	0.00
79 S	Terphenyl-d14	80.000	92.274	-15.3	71	0.00
80	Butylbenzylphthalate	40.000	41.771	-4.4	61	0.00
81	Benzo(a)anthracene	40.000	41.493	-3.7	64	0.00
82	3,3'-Dichlorobenzidine	40.000	38.585	3.5	58	0.00
83	Chrysene	40.000	39.544	1.1	60	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	46.672	-16.7	70	-0.01
85 c	Di-n-octyl phthalate	40.000	41.379	-3.4	61	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	52	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.433	8.9	48	0.00
88	Benzo(b)fluoranthene	40.000	41.261	-3.2	52	0.00
89	Benzo(k)fluoranthene	40.000	42.996	-7.5	59	0.00
90 C	Benzo(a)pyrene	40.000	41.562	-3.9	55	0.00
91	Dibenzo(a,h)anthracene	40.000	36.513	8.7	49	0.00
92	Benzo(g,h,i)perylene	40.000	35.552	11.1	47	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060122\
Data File : BF128617.D
Acq On : 01 Jun 2022 11:20
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 02 04:33:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 27 01:51:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0