

Data Path : Z:\HPCHEM1\BNA_F\Data\BF060315\
 Data File : BF079674.D
 Acq On : 3 Jun 2015 21:07
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Jun 04 03:43:10 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 15:20:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	85649	20.00	ng	0.00
21) Naphthalene-d8	8.81	136	329701	20.00	ng	0.00
38) Acenaphthene-d10	10.59	164	168852	20.00	ng	0.00
63) Phenanthrene-d10	12.11	188	352178	20.00	ng	-0.02
75) Chrysene-d12	15.19	240	372882	20.00	ng	-0.19
86) Perylene-d12	17.19	264	338985	20.00	ng	-0.30

System Monitoring Compounds						
5) 2-Fluorophenol	6.14	112	380601	80.69	ng	0.00
7) Phenol-d6	7.11	99	533734	86.06	ng	0.00
23) Nitrobenzene-d5	8.08	82	455886	81.07	ng	0.00
41) 2,4,6-Tribromophenol	11.38	330	115086	82.58	ng	-0.01
44) 2-Fluorobiphenyl	9.88	172	971425	87.28	ng	0.00
78) Terphenyl-d14	13.85	244	1267014	86.27	ng	-0.10

Target Compounds					Qvalue	
2) 1,4-Dioxane	3.60	88	86935	39.80	ng	99
3) Pyridine	4.31	79	249876	39.80	ng	99
4) n-Nitrosodimethylamine	4.23	42	118300	42.27	ng	99
6) Aniline	7.18	93	364507	40.94	ng	99
8) 2-Chlorophenol	7.30	128	230835	42.04	ng	97
9) Benzaldehyde	7.06	77	149422	41.55	ng	99
10) Phenol	7.12	94	281821	42.37	ng	99
11) bis(2-Chloroethyl)ether	7.23	93	218490	41.20	ng	97
12) 1,3-Dichlorobenzene	7.46	146	256688	40.88	ng	99
13) 1,4-Dichlorobenzene	7.54	146	259646	40.44	ng	98
14) 1,2-Dichlorobenzene	7.54	146	259646	44.67	ng	99
15) Benzyl Alcohol	7.63	79	182159	40.28	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.78	45	348098	39.37	ng	99
17) 2-Methylphenol	7.74	107	184039	40.60	ng	98
18) Hexachloroethane	8.04	117	83932	39.93	ng	93
19) n-Nitroso-di-n-propylamine	7.91	70	179967	42.39	ng	98
20) 3+4-Methylphenols	7.88	107	245136	40.83	ng	99
22) Acetophenone	7.92	105	313544	40.64	ng	97
24) Nitrobenzene	8.09	77	229669	41.57	ng	98
25) Isophorone	8.33	82	467717	41.62	ng	99
26) 2-Nitrophenol	8.41	139	73554	40.31	ng	96
27) 2,4-Dimethylphenol	8.43	122	194186	37.76	ng	98
28) bis(2-Chloroethoxy)methane	8.52	93	259306	40.16	ng	99
29) 2,4-Dichlorophenol	8.65	162	195807	42.69	ng	97
30) 1,2,4-Trichlorobenzene	8.75	180	205027	39.66	ng	98
31) Naphthalene	8.83	128	685647	39.78	ng	100
32) Benzoic acid	8.48	122	57491	41.45	ng	97
33) 4-Chloroaniline	8.87	127	385493	41.13	ng	98
34) Hexachlorobutadiene	8.95	225	113112	39.04	ng	95
35) Caprolactam	9.20	113	56170	44.48	ng	# 62
36) 4-Chloro-3-methylphenol	9.32	107	198175	42.33	ng	97
37) 2-Methylnaphthalene	9.53	142	478052	41.58	ng	99
39) 1,2,4,5-Tetrachlorobenzene	9.70	216	224356	41.40	ng	98
40) Hexachlorocyclopentadiene	9.69	237	93351	36.13	ng	100


Data Path : Z:\HPCHEM1\BNA_F\Data\BF060315\
 Data File : BF079674.D
 Acq On : 3 Jun 2015 21:07
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Jun 04 03:43:10 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 15:20:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.80	196	126080	42.79	ng	99
43) 2,4,5-Trichlorophenol	9.84	196	150406	46.07	ng	96
45) 1,1'-Biphenyl	9.99	154	534388	41.04	ng	95
46) 2-Chloronaphthalene	10.02	162	426837	41.84	ng	# 90
47) 2-Nitroaniline	10.10	65	101561	48.35	ng	92
48) Acenaphthylene	10.46	152	702638	41.55	ng	99
49) Dimethylphthalate	10.27	163	513213	41.48	ng	99
50) 2,6-Dinitrotoluene	10.34	165	92002	46.81	ng	91
51) Acenaphthene	10.63	154	426226	41.59	ng	98
52) 3-Nitroaniline	10.51	138	120083	49.53	ng	93
53) 2,4-Dinitrophenol	10.62	184	22583	40.41	ng	# 1
54) Dibenzofuran	10.80	168	629602	42.33	ng	97
55) 4-Nitrophenol	10.65	139	82677	42.28	ng	94
56) 2,4-Dinitrotoluene	10.75	165	105079	43.57	ng	88
57) Fluorene	11.14	166	527213	43.37	ng	99
58) 2,3,4,6-Tetrachlorophenol	10.90	232	107326	47.44	ng	99
59) Diethylphthalate	10.98	149	538538	45.33	ng	97
60) 4-Chlorophenyl-phenylether	11.13	204	235603	42.94	ng	91
61) 4-Nitroaniline	11.13	138	118629	49.60	ng	95
62) Azobenzene	11.29	77	466779	42.02	ng	86
64) 4,6-Dinitro-2-methylphenol	11.16	198	38469	39.34	ng	# 73
65) n-Nitrosodiphenylamine	11.23	169	440245	39.81	ng	99
66) 4-Bromophenyl-phenylether	11.62	248	145455	41.16	ng	95
67) Hexachlorobenzene	11.71	284	164257	39.56	ng	95
68) Atrazine	11.76	200	127953	45.17	ng	96
69) Pentachlorophenol	11.90	266	72912	47.93	ng	98
70) Phenanthrene	12.14	178	812021	41.70	ng	100
71) Anthracene	12.19	178	798261	40.75	ng	99
72) Carbazole	12.34	167	743644	41.34	ng	98
73) Di-n-butylphthalate	14.41	149	329070	44.37	ng	99
74) Fluoranthene	13.44	202	865111	41.21	ng	93
76) Benzidine	13.55	184	297645	35.09	ng	99
77) Pyrene	13.70	202	880185	41.00	ng	99
79) Butylbenzylphthalate	14.41	149	329070	45.02	ng	97
80) Benzo(a)anthracene	15.17	228	857445	42.05	ng	99
81) 3,3'-Dichlorobenzidine	15.11	252	289406	47.51	ng	98
82) Chrysene	15.22	228	843601	41.12	ng	98
83) Bis(2-ethylhexyl)phthalate	15.13	149	528158	43.49	ng	# 97
84) Di-n-octyl phthalate	15.98	149	816730	45.58	ng	100
85) Indeno(1,2,3-cd)pyrene	19.22	276	909447	45.33	ng	100
87) Benzo(b)fluoranthene	16.59	252	790908	45.15	ng	98
88) Benzo(k)fluoranthene	16.64	252	846301	41.24	ng	99
89) Benzo(a)pyrene	17.11	252	768313	44.86	ng	# 97
90) Dibenzo(a,h)anthracene	19.26	278	750618	46.77	ng	99
91) Benzo(g,h,i)perylene	19.84	276	758081	44.53	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.52 min Scan# 554
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 152 Resp: 85649
Ion Ratio Lower Upper
152 100
150 159.4 126.7 190.1
115 61.2 49.2 73.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 39.80 ng
RT: 3.60 min Scan# 211
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion: 88 Resp: 86935
Ion Ratio Lower Upper
88 100
58 80.3 64.2 96.2
43 32.6 24.4 36.6



Abundance Ion 88.00 (87.70 to 88.70): BF0
Ion 58.00 (57.70 to 58.70): BF0
Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 39.80 ng

RT: 4.31 min Scan# 273

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 79 Resp: 249876

Ion Ratio Lower Upper

79 100

52 73.5 60.1 90.1

51 34.9 28.2 42.4



#4

n-Nitrosodimethylamine

Concen: 42.27 ng

RT: 4.23 min Scan# 266

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07




Tgt Ion: 42 Resp: 118300


Ion Ratio Lower Upper

42 100

74 132.9 107.5 161.3

44 7.8 6.2 9.2









#5
 2-Fluorophenol
 Concen: 80.69 ng
 RT: 6.14 min Scan# 433
 Delta R.T. 0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
112	100		
64	51.0	41.1	61.7
63	25.8	20.5	30.7



#6
 Aniline
 Concen: 40.94 ng
 RT: 7.18 min Scan# 524
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion	Ratio	Lower	Upper
93	100		
66	36.1	28.5	42.7
65	18.3	15.0	22.6





#7

Phenol-d6

Concen: 86.06 ng

RT: 7.11 min Scan# 518

Delta R.T. 0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 99 Resp: 533734

Ion Ratio Lower Upper

99 100

42 18.9 15.2 22.8




71 33.2 25.9 38.9



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#8

2-Chlorophenol

Concen: 42.04 ng

RT: 7.30 min Scan# 535

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion: 128 Resp: 230835

Ion Ratio Lower Upper

128 100

130 32.4 12.7 52.7



64 46.4 29.0 69.0




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF0





#9

Benzaldehyde

Concen: 41.55 ng

RT: 7.06 min Scan# 514

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 77 Resp: 149422

Ion Ratio Lower Upper

77 100

105 87.4 68.2 108.2




106 84.6 64.9 104.9



Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): BF0

Ion 106.00 (105.70 to 106.70): BF0



#10

Phenol

Concen: 42.37 ng

RT: 7.12 min Scan# 519

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion: 94 Resp: 281821

Ion Ratio Lower Upper

94 100

65 30.0 10.9 50.9



66 37.6 18.1 58.1




Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0









#11
 bis(2-Chloroethyl)ether
 Concen: 41.20 ng
 RT: 7.23 min Scan# 529
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion: 93 Resp: 218490
 Ion Ratio Lower Upper
 93 100
 63 72.0 54.7 94.7
 95 32.4 13.5 53.5



#12
 1,3-Dichlorobenzene
 Concen: 40.88 ng
 RT: 7.46 min Scan# 549
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion: 146 Resp: 256688
 Ion Ratio Lower Upper
 146 100
 148 63.5 51.1 76.7
 75 33.6 28.2 42.4









#13
 1,4-Dichlorobenzene
 Concen: 40.44 ng
 RT: 7.54 min Scan# 556
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:146 Resp: 259646
 Ion Ratio Lower Upper
 146 100
 148 66.0 52.1 78.1
 111 35.0 29.1 43.7



#14
 1,2-Dichlorobenzene
 Concen: 44.67 ng
 RT: 7.54 min Scan# 556
 Delta R.T. -0.16 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:146 Resp: 259646
 Ion Ratio Lower Upper
 146 100
 148 66.0 52.6 79.0
 111 35.0 27.3 40.9





#15

Benzyl Alcohol

Concen: 40.28 ng

RT: 7.63 min Scan# 564

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 79 Resp: 182159

Ion Ratio Lower Upper

79 100

108 68.7 56.2 84.4

77 64.1 50.7 76.1



#16

2,2'-oxybis(1-Chloropropane)

Concen: 39.37 ng

RT: 7.78 min Scan# 577

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07




Tgt Ion: 45 Resp: 348098


Ion Ratio Lower Upper

45 100

77 13.2 0.0 32.9

79 10.3 0.0 29.8





#17

2-Methylphenol

Concen: 40.60 ng

RT: 7.74 min Scan# 573

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:107 Resp: 184039


Ion Ratio Lower Upper

107 100

108 111.8 90.9 136.3

77 47.2 38.6 57.8

79 46.5 38.0 57.0






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#18

Hexachloroethane

Concen: 39.93 ng

RT: 8.04 min Scan# 600

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:117 Resp: 83932

Ion Ratio Lower Upper

117 100

119 94.0 76.7 115.1



201 87.8 61.7 92.5






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 42.39 ng
RT: 7.91 min Scan# 588
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07


Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 70 Resp: 179967
Ion Ratio Lower Upper
70 100
42 58.1 47.7 71.5
101 9.1 7.4 11.0
130 20.2 15.8 23.8



#20
3+4-Methylphenols
Concen: 40.83 ng
RT: 7.88 min Scan# 586
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion: 107 Resp: 245136
Ion Ratio Lower Upper
107 100
108 85.9 66.2 106.2
77 34.0 12.9 52.9
79 28.1 7.3 47.3





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.81 min Scan# 667

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:136 Resp: 329701


Ion Ratio Lower Upper

136 100

137 11.2 9.3 13.9

54 9.6 8.3 12.5

68 6.9 6.1 9.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 40.64 ng

RT: 7.92 min Scan# 589

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Tgt Ion:105 Resp: 313544


Ion Ratio Lower Upper

105 100

71 0.3 0.2 0.4

51 22.4 19.2 28.8

120 22.9 17.6 26.4






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 81.07 ng
 RT: 8.08 min Scan# 603
 Delta R.T. 0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion: 82 Resp: 455886
 Ion Ratio Lower Upper
 82 100
 128 52.3 39.4 59.0
 54 45.2 37.4 56.0



#24
 Nitrobenzene
 Concen: 41.57 ng
 RT: 8.09 min Scan# 604
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion: 77 Resp: 229669
 Ion Ratio Lower Upper
 77 100
 123 39.4 30.6 45.8
 65 13.6 10.5 15.7





#25

Isophorone

Concen: 41.62 ng

RT: 8.33 min Scan# 625

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 82 Resp: 467717

Ion Ratio Lower Upper

82 100

95 7.0 5.4 8.2




138 18.6 14.4 21.6



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 138.00 (137.70 to 138.70): BF0



#26

2-Nitrophenol

Concen: 40.31 ng

RT: 8.41 min Scan# 632

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion: 139 Resp: 73554

Ion Ratio Lower Upper

139 100

109 29.2 23.6 35.4



65 67.5 57.4 86.0




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0





#27

2,4-Dimethylphenol

Concen: 37.76 ng

RT: 8.43 min Scan# 634

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:122 Resp: 194186

Ion Ratio Lower Upper

122 100

107 107.6 85.0 127.6




121 60.3 46.3 69.5



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 40.16 ng

RT: 8.52 min Scan# 642

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion: 93 Resp: 259306

Ion Ratio Lower Upper

93 100

95 32.0 25.8 38.6



123 9.5 6.8 10.2




Abundance Ion 93.00 (92.70 to 93.70): BF0

Ion 95.00 (94.70 to 95.70): BF0

Ion 123.00 (122.70 to 123.70): E





#29

2,4-Dichlorophenol

Concen: 42.69 ng

RT: 8.65 min Scan# 653

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:162 Resp: 195807

Ion Ratio Lower Upper

162 100

164 64.2 43.0 83.0




98 37.1 19.8 59.8



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF0



#30

1,2,4-Trichlorobenzene

Concen: 39.66 ng

RT: 8.75 min Scan# 662

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:180 Resp: 205027

Ion Ratio Lower Upper

180 100

182 96.8 76.1 114.1



145 26.1 21.8 32.8




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 39.78 ng

RT: 8.83 min Scan# 669

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:128 Resp: 685647

Ion Ratio Lower Upper

128 100

129 11.7 9.3 13.9




127 13.3 10.4 15.6



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 41.45 ng

RT: 8.48 min Scan# 638

Delta R.T. 0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:122 Resp: 57491

Ion Ratio Lower Upper

122 100

105 126.4 109.7 149.7



77 102.7 86.1 126.1




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0





#33

4-Chloroaniline

Concen: 41.13 ng

RT: 8.87 min Scan# 672

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:127 Resp: 385493


Ion Ratio Lower Upper

127 100

129 33.9 26.5 39.7

65 27.3 23.0 34.6

92 18.2 14.7 22.1




Abundance Ion 127.00 (126.70 to 127.70): E


Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0



Time-->



#34

Hexachlorobutadiene

Concen: 39.04 ng

RT: 8.95 min Scan# 679

Delta R.T. -0.01 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:225 Resp: 113112

Ion Ratio Lower Upper

225 100

223 66.8 49.6 74.4


227 62.4 52.3 78.5




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E







Time-->



#35
 Caprolactam
 Concen: 44.48 ng
 RT: 9.20 min Scan# 701
 Delta R.T. -0.01 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:113 Resp: 56170
 Ion Ratio Lower Upper
 113 100
 55 216.0 142.3 182.3#
 56 159.4 99.8 139.8#



#36
 4-Chloro-3-methylphenol
 Concen: 42.33 ng
 RT: 9.32 min Scan# 712
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:107 Resp: 198175
 Ion Ratio Lower Upper
 107 100
 144 23.0 17.0 25.4
 142 71.0 55.1 82.7









#37
 2-Methylnaphthalene
 Concen: 41.58 ng
 RT: 9.53 min Scan# 730
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:142 Resp: 478052
 Ion Ratio Lower Upper
 142 100
 141 89.8 71.4 107.0
 115 32.8 27.3 40.9



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 10.59 min Scan# 823
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:164 Resp: 168852
 Ion Ratio Lower Upper
 164 100
 162 105.8 82.0 123.0
 160 46.1 35.9 53.9





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.40 ng

RT: 9.70 min Scan# 745

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC

Tgt Ion:216 Resp: 224356


Ion Ratio Lower Upper

216 100

214 80.3 62.8 94.2

179 20.8 16.6 24.8

108 17.3 14.2 21.4




Abundance Ion 216.00 (215.70 to 216.70): E



Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



Time-->



#40

Hexachlorocyclopentadiene

Concen: 36.13 ng

RT: 9.69 min Scan# 744

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:237 Resp: 93351

Ion Ratio Lower Upper

237 100

235 61.6 41.7 81.7


272 11.7 0.0 31.1




Abundance Ion 236.80 (236.50 to 237.50): E


Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



Time-->









#41
 2,4,6-Tribromophenol
 Concen: 82.58 ng
 RT: 11.38 min Scan# 892
 Delta R.T. -0.01 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:330 Resp: 115086
 Ion Ratio Lower Upper
 330 100
 332 97.7 77.3 115.9
 141 39.1 27.4 41.2



#42
 2,4,6-Trichlorophenol
 Concen: 42.79 ng
 RT: 9.80 min Scan# 754
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:196 Resp: 126080
 Ion Ratio Lower Upper
 196 100
 198 97.5 77.3 115.9
 200 31.1 24.0 36.0









#43
 2,4,5-Trichlorophenol
 Concen: 46.07 ng
 RT: 9.84 min Scan# 757
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:196 Resp: 150406
 Ion Ratio Lower Upper
 196 100
 198 96.0 75.3 112.9
 97 39.9 37.6 56.4
 132 25.8 21.5 32.3




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 87.28 ng
 RT: 9.88 min Scan# 761
 Delta R.T. 0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:172 Resp: 971425
 Ion Ratio Lower Upper
 172 100
 171 36.1 28.8 43.2
 170 24.5 19.2 28.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 41.04 ng
 RT: 9.99 min Scan# 770
 Delta R.T. -0.01 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:154 Resp: 534388
 Ion Ratio Lower Upper
 154 100
 153 40.3 19.6 59.6
 76 18.0 0.0 29.5




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF0






#46
 2-Chloronaphthalene
 Concen: 41.84 ng
 RT: 10.02 min Scan# 773
 Delta R.T. -0.01 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:162 Resp: 426837
 Ion Ratio Lower Upper
 162 100
 127 42.1 25.6 38.4#
 164 32.1 26.3 39.5



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 48.35 ng
RT: 10.10 min Scan# 780
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion: 65 Resp: 101561
Ion Ratio Lower Upper
65 100
92 64.1 49.7 74.5
138 88.9 62.6 94.0




Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): BF0






#48
Acenaphthylene
Concen: 41.55 ng
RT: 10.46 min Scan# 811
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion: 152 Resp: 702638
Ion Ratio Lower Upper
152 100
151 19.7 16.2 24.4
153 13.1 10.4 15.6



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#49

Dimethylphthalate

Concen: 41.48 ng

RT: 10.27 min Scan# 795

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:163 Resp: 513213

Ion Ratio Lower Upper

163 100

194 3.4 2.6 3.8




164 10.2 8.6 12.8



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 46.81 ng

RT: 10.34 min Scan# 801

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:165 Resp: 92002

Ion Ratio Lower Upper

165 100

63 48.1 45.9 68.9



89 53.7 46.8 70.2




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 41.59 ng

RT: 10.63 min Scan# 826

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:154 Resp: 426226

Ion Ratio Lower Upper

154 100

153 114.2 89.8 134.6




152 54.2 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 49.53 ng

RT: 10.51 min Scan# 816

Delta R.T. -0.00 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07


Tgt Ion:138 Resp: 120083

Ion Ratio Lower Upper

138 100

108 11.3 9.1 13.7



92 118.0 101.4 152.2




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 40.41 ng
 RT: 10.62 min Scan# 825
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040EC

Tgt Ion:184 Resp: 22583
 Ion Ratio Lower Upper
 184 100
 63 154.1 114.0 171.0
 154 678.4 297.0 445.4#




Abundance Ion 184.00 (183.70 to 184.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 154.00 (153.70 to 154.70): E






#54
 Dibenzofuran
 Concen: 42.33 ng
 RT: 10.80 min Scan# 841
 Delta R.T. -0.00 min
 Lab File: BF079674.D
 Acq: 3 Jun 2015 21:07

Tgt Ion:168 Resp: 629602
 Ion Ratio Lower Upper
 168 100
 139 34.3 29.1 43.7
 169 13.2 10.7 16.1



Abundance Ion 168.00 (167.70 to 168.70): E
 Ion 139.00 (138.70 to 139.70): E
 Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 42.28 ng
RT: 10.65 min Scan# 828
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:139 Resp: 82677
Ion Ratio Lower Upper
139 100
109 58.9 37.1 77.1
65 61.2 49.2 89.2



#56
2,4-Dinitrotoluene
Concen: 43.57 ng
RT: 10.75 min Scan# 837
Delta R.T. 0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion:165 Resp: 105079
Ion Ratio Lower Upper
165 100
63 25.5 25.5 38.3
89 51.3 48.5 72.7
182 2.3 1.9 2.9





#57

Fluorene

Concen: 43.37 ng

RT: 11.14 min Scan# 871

Delta R.T. -0.01 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040EC


Tgt Ion:166 Resp: 527213

Ion Ratio Lower Upper

166 100

165 98.8 79.8 119.6




167 13.0 10.6 15.8



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 47.44 ng

RT: 10.90 min Scan# 850

Delta R.T. -0.01 min

Lab File: BF079674.D

Acq: 3 Jun 2015 21:07

Tgt Ion:232 Resp: 107326


Ion Ratio Lower Upper

232 100

131 44.7 36.6 54.8

130 2.0 1.6 2.4

166 30.6 24.1 36.1






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 45.33 ng
RT: 10.98 min Scan# 857
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC


Tgt Ion:149 Resp: 538538
Ion Ratio Lower Upper
149 100
177 22.4 16.1 24.1
150 12.8 10.2 15.2



#60
4-Chlorophenyl-phenylether
Concen: 42.94 ng
RT: 11.13 min Scan# 870
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion:204 Resp: 235603
Ion Ratio Lower Upper
204 100
206 34.2 27.5 41.3
141 43.2 42.8 64.2









#61
4-Nitroaniline
Concen: 49.60 ng
RT: 11.13 min Scan# 870
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040EC

Tgt Ion:138 Resp: 118629
Ion Ratio Lower Upper
138 100
92 54.5 40.7 80.7
108 73.6 56.0 96.0




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BF0
Ion 108.00 (107.70 to 108.70): E



#62
Azobenzene
Concen: 42.02 ng
RT: 11.29 min Scan# 884
Delta R.T. -0.00 min
Lab File: BF079674.D
Acq: 3 Jun 2015 21:07

Tgt Ion: 77 Resp: 466779
Ion Ratio Lower Upper
77 100
182 35.6 1.3 41.3
105 24.9 2.3 42.3
51 25.5 9.9 49.9



Abundance Ion 77.00 (76.70 to 77.70): BF0
Ion 182.00 (181.70 to 182.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 51.00 (50.70 to 51.70): BF0

