

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060315\
 Data File : BF079663.D
 Acq On : 3 Jun 2015 15:03
 Operator : TP/IZ
 Sample : G2446-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T-15000

Manual Integrations
 APPROVED

apatel
 6/5/2015 8:05:49 AM

Quant Time: Jun 03 16:22:30 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 03 15:20:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	80846	20.00	ng	0.00
21) Naphthalene-d8	8.81	136	312328	20.00	ng	0.00
38) Acenaphthene-d10	10.59	164	147481	20.00	ng	0.00
63) Phenanthrene-d10	12.09	188	306358	20.00	ng	-0.05
75) Chrysene-d12	14.95	240	334790m	20.00	ng	-0.43
86) Perylene-d12	16.87	264	297353m	20.00	ng	-0.62
System Monitoring Compounds						
5) 2-Fluorophenol	6.14	112	218733	49.35	ng	0.00
7) Phenol-d6	7.11	99	184772	30.29	ng	0.00
23) Nitrobenzene-d5	8.07	82	501624	92.50	ng	-0.01
41) 2,4,6-Tribromophenol	11.38	330	111700	91.77	ng	-0.01
44) 2-Fluorobiphenyl	9.88	172	1160985	122.04	ng	0.00
78) Terphenyl-d14	13.74	244	1561252	120.27	ng	-0.22

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060315\
Data File : BF079663.D
Acq On : 3 Jun 2015 15:03
Operator : TP/IZ
Sample : G2446-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
T-15000

Manual Integrations
APPROVED

apatel
6/5/2015 8:05:49 AM

Quant Time: Jun 03 16:22:30 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 03 15:20:47 2015
Response via : Initial Calibration

