

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060315\
 Data File : BF079690.D
 Acq On : 4 Jun 2015 6:24
 Operator : TP/IZ
 Sample : G2427-09
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HB-9A-DEL3(0-2)

Manual Integrations
 APPROVED

apatel
 6/5/2015 8:10:48 AM

Quant Time: Jun 05 12:01:36 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 05 10:52:02 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	101277	20.00	ng	0.00
21) Naphthalene-d8	8.81	136	389456	20.00	ng	0.00
38) Acenaphthene-d10	10.59	164	207609	20.00	ng	0.00
63) Phenanthrene-d10	12.11	188	408701	20.00	ng	-0.02
75) Chrysene-d12	15.18	240	424400	20.00	ng	-0.21
86) Perylene-d12	17.18	264	409164	20.00	ng	-0.31
System Monitoring Compounds						
5) 2-Fluorophenol	6.14	112	443222	79.47	ng	0.00
7) Phenol-d6	7.11	99	598536	81.51	ng	0.00
23) Nitrobenzene-d5	8.08	82	316045	49.50	ng	0.00
41) 2,4,6-Tribromophenol	11.39	330	123041	71.81	ng	0.00
44) 2-Fluorobiphenyl	9.88	172	760148	52.96	ng	0.00
78) Terphenyl-d14	13.84	244	869525	50.02	ng	-0.11
Target Compounds						Qvalue
37) 2-Methylnaphthalene	9.53	142	30832	2.27	ng	99
49) Dimethylphthalate	10.27	163	331554	21.80	ng	98
70) Phenanthrene	12.14	178	157036	6.95	ng	98
74) Fluoranthene	13.43	202	199868	8.20	ng	98
77) Pyrene	13.70	202	179540	7.35	ng	98
79) Butylbenzylphthalate	14.40	149	43712	7.26	ng	98
80) Benzo(a)anthracene	15.15	228	104475	4.50	ng	98
82) Chrysene	15.20	228	93134	3.99	ng	95
83) Bis(2-ethylhexyl)phthalate	15.12	149	128755	11.84	ng	# 96
85) Indeno(1,2,3-cd)pyrene	19.21	276	53570	2.35	ng	91
87) Benzo(b)fluoranthene	16.58	252	122977m	5.82	ng	
89) Benzo(a)pyrene	17.10	252	82404	3.99	ng	99
91) Benzo(a,h,i)perylene	19.83	276	45583	2.22	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060315\
Data File : BF079690.D
Acq On : 4 Jun 2015 6:24
Operator : TP/IZ
Sample : G2427-09
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HB-9A-DEL3(0-2)

Manual Integrations
APPROVED

apatel
6/5/2015 8:10:48 AM

Quant Time: Jun 05 12:01:36 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 05 10:52:02 2015
Response via : Initial Calibration

