


Hexachlorocyclopentadiene

Response Ratio




Resp Ratio = $3.32 \times 10^{-1} \times \text{Amt} - 4.87 \times 10^{-2}$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060415.M
Calibration Table Last Updated: Fri Jun 05 19:20:13 2015


2,6-Dinitrotoluene

Response Ratio




Resp Ratio = $3.11 \times 10^{-1} \times \text{Amt} - 4.74 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060415.M
Calibration Table Last Updated: Fri Jun 05 19:22:39 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060415.M
Calibration Table Last Updated: Fri Jun 05 19:22:57 2015

4,6-Dinitro-2-methylphenol



$R = 1.68e-003 A^2 + 5.43e-002 A - 8.48e-003$
Coef of Det (r^2) = 0.990 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060415.M
Calibration Table Last Updated: Fri Jun 05 19:22:57 2015