


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF060416\
 Data File : BF087656.D
 Acq On : 4 Jun 2016 10:13
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060416.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Jun 04 11:07:28 2016



AutoFind: Scans 879, 880, 881; Background Corrected with Scan 876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.5	13431	PASS
68	69	0.00	2	1.5	185	PASS
69	198	0.00	100	40.9	12628	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	57.5	17760	PASS
197	198	0.00	2	0.5	162	PASS
198	198	100	100	100.0	30890	PASS
199	198	5	9	6.6	2054	PASS
275	198	10	60	24.0	7402	PASS
365	198	1	100	2.9	900	PASS
441	442	0.01	24	15.6	4184	PASS
442	198	50	100	86.7	26786	PASS
443	442	15	24	19.2	5154	PASS