


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060420\
 Data File : BF120528.D
 Acq On : 4 Jun 2020 12:41
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF052820.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 04 16:38:34 2020



AutoFind: Scans 1652, 1653, 1654; Background Corrected with Scan 1647

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.5	36700	PASS
68	69	0.00	2	1.5	529	PASS
69	198	0.00	100	32.3	35416	PASS
70	69	0.00	2	0.3	109	PASS
127	198	10	80	46.1	50517	PASS
197	198	0.00	2	0.5	553	PASS
198	198	100	100	100.0	109602	PASS
199	198	5	9	6.6	7229	PASS
275	198	10	60	28.2	30898	PASS
365	198	1	100	3.5	3849	PASS
441	198	0.01	100	15.4	16846	PASS
442	198	50	100	98.8	108232	PASS
443	442	15	24	18.8	20298	PASS