

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060421\
 Data File : BF124129.D
 Acq On : 4 Jun 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 10:54:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 04 09:56:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	36.941	7.6	100	0.00
3	Pyridine	40.000	39.507	1.2	106	0.00
4	n-Nitrosodimethylamine	40.000	39.024	2.4	105	0.00
5 S	2-Fluorophenol	80.000	76.772	4.0	104	0.00
6	Aniline	40.000	37.936	5.2	105	0.00
7 S	Phenol-d6	80.000	76.226	4.7	105	0.00
8	2-Chlorophenol	40.000	37.930	5.2	102	0.00
9	Benzaldehyde	40.000	36.789	8.0	105	0.00
10 C	Phenol	40.000	38.570	3.6	106	0.00
11	bis(2-Chloroethyl)ether	40.000	38.525	3.7	103	0.00
12	1,3-Dichlorobenzene	40.000	38.800	3.0	105	0.00
13 C	1,4-Dichlorobenzene	40.000	38.706	3.2	105	0.00
14	1,2-Dichlorobenzene	40.000	38.354	4.1	104	0.00
15	Benzyl Alcohol	40.000	37.327	6.7	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.889	5.3	106	0.00
17	2-Methylphenol	40.000	38.609	3.5	105	0.00
18	Hexachloroethane	40.000	37.535	6.2	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.117	2.2	112	0.00
20	3+4-Methylphenols	40.000	39.050	2.4	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	108	0.00
22	Acetophenone	40.000	37.715	5.7	105	0.00
23 S	Nitrobenzene-d5	80.000	73.830	7.7	105	0.00
24	Nitrobenzene	40.000	37.229	6.9	102	0.00
25	Isophorone	40.000	37.444	6.4	106	0.00
26 C	2-Nitrophenol	40.000	37.999	5.0	106	0.00
27	2,4-Dimethylphenol	40.000	36.948	7.6	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.108	7.2	109	0.00
29 C	2,4-Dichlorophenol	40.000	38.832	2.9	109	0.00
30	1,2,4-Trichlorobenzene	40.000	36.543	8.6	104	0.00
31	Naphthalene	40.000	36.693	8.3	103	0.00
32	Benzoic acid	40.000	31.435	21.4	94	0.00
33	4-Chloroaniline	40.000	38.224	4.4	108	0.00
34 C	Hexachlorobutadiene	40.000	36.738	8.2	103	0.00
35	Caprolactam	40.000	37.283	6.8	107	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.818	8.0	109	0.00
37	2-Methylnaphthalene	40.000	37.168	7.1	104	0.00
38	1-Methylnaphthalene	40.000	36.996	7.5	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.157	9.6	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.975	10.1	106	0.00
42 S	2,4,6-Tribromophenol	80.000	72.659	9.2	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.182	7.0	105	0.00
44	2,4,5-Trichlorophenol	40.000	36.017	10.0	103	0.00
45 S	2-Fluorobiphenyl	80.000	73.560	8.0	106	0.00
46	1,1'-Biphenyl	40.000	36.232	9.4	106	0.00
47	2-Chloronaphthalene	40.000	35.793	10.5	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060421\
 Data File : BF124129.D
 Acq On : 4 Jun 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 04 10:54:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 04 09:56:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.801	5.5	110	0.00
49	Acenaphthylene	40.000	36.555	8.6	106	0.00
50	Dimethylphthalate	40.000	35.314	11.7	106	0.00
51	2,6-Dinitrotoluene	40.000	35.246	11.9	106	0.00
52 C	Acenaphthene	40.000	35.910	10.2	105	0.00
53	3-Nitroaniline	40.000	34.526	13.7	102	0.00
54 P	2,4-Dinitrophenol	40.000	36.516	8.7	109	0.00
55	Dibenzofuran	40.000	36.172	9.6	106	0.00
56 P	4-Nitrophenol	40.000	34.635	13.4	96	0.00
57	2,4-Dinitrotoluene	40.000	36.373	9.1	101	0.00
58	Fluorene	40.000	36.972	7.6	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.598	6.0	112	0.00
60	Diethylphthalate	40.000	36.476	8.8	106	0.00
61	4-Chlorophenyl-phenylether	40.000	37.007	7.5	110	0.00
62	4-Nitroaniline	40.000	35.573	11.1	101	0.00
63	Azobenzene	40.000	35.401	11.5	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.622	8.4	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.188	9.5	104	0.00
67	4-Bromophenyl-phenylether	40.000	36.236	9.4	109	0.00
68	Hexachlorobenzene	40.000	35.833	10.4	104	0.00
69	Atrazine	40.000	39.877	0.3	105	0.00
70 C	Pentachlorophenol	40.000	35.092	12.3	102	0.00
71	Phenanthrene	40.000	36.789	8.0	105	0.00
72	Anthracene	40.000	36.178	9.6	107	0.00
73	Carbazole	40.000	35.738	10.7	104	0.00
74	Di-n-butylphthalate	40.000	36.185	9.5	106	0.00
75 C	Fluoranthene	40.000	36.121	9.7	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzydine	40.000	44.269	-10.7	112	0.00
78	Pyrene	40.000	37.383	6.5	106	0.00
79 S	Terphenyl-d14	80.000	75.282	5.9	106	0.00
80	Butylbenzylphthalate	40.000	39.267	1.8	112	0.00
81	Benzo(a)anthracene	40.000	38.194	4.5	111	0.00
82	3,3'-Dichlorobenzidine	40.000	40.607	-1.5	117	0.00
83	Chrysene	40.000	37.664	5.8	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.233	-3.1	117	0.00
85 c	Di-n-octyl phthalate	40.000	40.640	-1.6	118	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.738	10.7	103	0.00
88	Benzo(b)fluoranthene	40.000	38.310	4.2	112	0.00
89	Benzo(k)fluoranthene	40.000	38.257	4.4	109	0.00
90 C	Benzo(a)pyrene	40.000	37.853	5.4	109	0.00
91	Dibenzo(a,h)anthracene	40.000	35.872	10.3	105	0.00
92	Benzo(g,h,i)perylene	40.000	35.688	10.8	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060421\
Data File : BF124129.D
Acq On : 4 Jun 2021 9:19
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 04 10:54:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 04 09:56:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0