




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060425\
Data File : BF142601.D
Acq On : 04 Jun 2025 11:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 04 13:03:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration



TIC: BF142601.D\data.ms

(70) Pentachlorophenol (C)




11.228min (-0.006) 34098.66 ng

response 188828

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.60	62.83
264.00	63.00	64.15
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060425\
Data File : BF142601.D
Acq On : 04 Jun 2025 11:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 04 13:03:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration



TIC: BF142601.D\data.ms

(77) Benzidine

12.763min (+ 0.000) 0.00 ng

response 822752

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	14.00
183.00	11.20	11.05
0.00	0.00	0.00