

Data Path : Y:\HPCHEM1\BNA F\DATA\BF060515\
 Data File : BF079757.D
 Acq On : 6 Jun 2015 5:58
 Operator : TP/IZ
 Sample : G2450-18 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)

Quant Time: Jun 08 04:48:29 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 04:31:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	51409	20.00	ng	0.00
21) Naphthalene-d8	8.79	136	214669	20.00	ng	0.00
38) Acenaphthene-d10	10.56	164	121954	20.00	ng	0.00
63) Phenanthrene-d10	12.07	188	236915	20.00	ng	0.00
75) Chrysene-d12	15.01	240	264207	20.00	ng	-0.01
86) Perylene-d12	16.95	264	255651	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	6.10	112	145632	23.74	ng	0.00
7) Phenol-d6	7.07	99	185871	23.46	ng	-0.01
23) Nitrobenzene-d5	8.04	82	100152	14.32	ng	0.00
41) 2,4,6-Tribromophenol	11.36	330	47662	20.43	ng	0.00
44) 2-Fluorobiphenyl	9.86	172	237232	13.18	ng	0.00
78) Terphenyl-d14	13.74	244	273258	11.84	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
6) Aniline	7.26	93	319	0.06	ng	# 1
9) Benzaldehyde	7.04	77	1298	0.51	ng	# 69
10) Phenol	7.10	94	868	0.20	ng	# 87
11) bis(2-Chloroethyl)ether	7.26	93	319	0.10	ng	# 1
15) Benzyl Alcohol	7.64	79	2334	0.71	ng	# 39
18) Hexachloroethane	8.02	117	427	0.33	ng	# 1
20) 3+4-Methylphenols	7.86	107	224	0.06	ng	# 19
22) Acetophenone	7.88	105	1402	0.27	ng	# 74
24) Nitrobenzene	8.04	77	556	0.16	ng	# 47
25) Isophorone	8.31	82	332	0.05	ng	# 1
27) 2,4-Dimethylphenol	8.42	122	206	0.06	ng	# 5
31) Naphthalene	8.80	128	18000	1.60	ng	# 97
32) Benzoic acid	8.42	122	206	4.12	ng	# 31
33) 4-Chloroaniline	8.80	127	2100	0.37	ng	# 32
35) Caprolactam	9.19	113	644	0.71	ng	# 3
36) 4-Chloro-3-methylphenol	9.32	107	1247	0.38	ng	# 13
37) 2-Methylnaphthalene	9.50	142	62648	8.75	ng	# 98
43) 2,4,5-Trichlorophenol	9.95	196	270	0.10	ng	# 1
45) 1,1'-Biophenyl	9.96	154	11031	1.04	ng	# 91
46) 2-Chloronaphthalene	9.98	162	956	0.11	ng	# 76
47) 2-Nitroaniline	10.07	65	1315	0.72	ng	# 62
48) Acenaphthylene	10.42	152	16889	1.18	ng	# 75
49) Dimethylphthalate	10.24	163	27058	2.88	ng	# 97
50) 2,6-Dinitrotoluene	10.32	165	959	1.79	ng	# 13
51) Acenaphthene	10.59	154	9998	1.23	ng	# 61
52) 3-Nitroaniline	10.51	138	2404	1.21	ng	# 60
54) Dibenzofuran	10.76	168	24295	2.11	ng	# 84
55) 4-Nitrophenol	10.65	139	12944	8.73	ng	# 63
56) 2,4-Dinitrotoluene	10.72	165	2044	3.02	ng	# 69
57) Fluorene	11.11	166	16362	1.65	ng	# 41
59) Diethylphthalate	10.95	149	1171	0.13	ng	# 18
60) 4-Chlorophenyl-phenylether	11.07	204	441	0.10	ng	# 1
61) 4-Nitroaniline	11.11	138	1763	0.81	ng	# 1
62) Azobenzene	11.27	77	5070	0.55	ng	# 1


Data Path : Y:\HPCHEM1\BNA F\DATA\BF060515\
 Data File : BF079757.D
 Acq On : 6 Jun 2015 5:58
 Operator : TP/IZ
 Sample : G2450-18 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)

Quant Time: Jun 08 04:48:29 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 04:31:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
64) 4,6-Dinitro-2-methylphenol	11.14	198	2048	5.34	nq	# 1
65) n-Nitrosodiphenylamine	11.22	169	52252	6.60	nq	# 40
66) 4-Bromophenyl-phenylether	11.67	248	211	0.08	nq	# 1
68) Atrazine	11.71	200	1320	0.56	nq	# 1
70) Phenanthrene	12.09	178	424044	30.57	nq	99
71) Anthracene	12.09	178	424044	31.00	nq	97
72) Carbazole	12.30	167	5356	0.41	nq	# 1
73) Di-n-butylphthalate	12.62	149	3205	0.22	nq	# 1
74) Fluoranthene	13.35	202	96004	6.32	nq	94
76) Benzidine	13.45	184	1293	0.16	nq	# 1
77) Pyrene	13.61	202	471809	29.13	nq	97
79) Butylbenzylphthalate	14.29	149	881	0.14	nq	# 2
80) Benzo(a)anthracene	14.99	228	161001	9.95	nq	# 78
81) 3,3'-Dichlorobenzidine	14.91	252	1614	0.29	nq	# 74
82) Chrysene	15.04	228	152171	10.18	nq	# 93
83) Bis(2-ethylhexyl)phthalate	14.95	149	3876	0.38	nq	# 56
84) Di-n-octyl phthalate	15.74	149	1568	0.09	nq	# 60
85) Indeno(1,2,3-cd)pyrene	18.95	276	43583	2.52	nq	95
87) Benzo(b)fluoranthene	16.37	252	93235	5.87	nq	# 80
88) Benzo(k)fluoranthene	16.37	252	93235	6.12	nq	# 83
89) Benzo(a)pyrene	16.87	252	115155	7.85	nq	# 91
90) Dibenzo(a,h)anthracene	18.97	278	21504	1.52	nq	# 55
91) Benzo(g,h,i)perylene	19.55	276	150183	10.27	nq	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.50 min Scan# 552
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)

Tot Ion	Ratio	Lower	Upper
152	100		
150	151.5	128.1	192.1
115	46.3	49.7	74.5




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 23.74 ng
RT: 6.10 min Scan# 430
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion	Ratio	Lower	Upper
112	100		
64	63.1	49.6	74.4
63	31.5	24.9	37.3



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0









#6
 Aniline
 Concen: 0.06 ng
 RT: 7.26 min Scan# 531
 Delta R.T. 0.11 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58




Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)


Tot Ion: 93 Resp: 319
 Ion Ratio Lower Upper
 93 100
 66 9828.8 30.3 45.5#
 65 861.5 15.8 23.6#



#7
 Phenol-d6
 Concen: 23.46 ng
 RT: 7.07 min Scan# 515
 Delta R.T. -0.01 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Tgt Ion: 99 Resp: 185871
 Ion Ratio Lower Upper
 99 100
 42 23.5 18.0 27.0
 71 37.8 28.4 42.6









#9
Benzaldehyde
Concen: 0.51 ng
RT: 7.04 min Scan# 512
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleID :
RT440NB6.3(2)


Tot Ion: 77 Resp: 1298
Ion Ratio Lower Upper
77 100
105 118.3 66.1 106.1#
106 58.1 61.9 101.9#



#10
Phenol
Concen: 0.20 ng
RT: 7.10 min Scan# 517
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion: 94 Resp: 868
Ion Ratio Lower Upper
94 100
65 36.2 5.3 45.3
66 34.5 11.3 51.3









#11
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.26 min Scan# 531
Delta R.T. 0.06 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion: 93 Resp: 319
Ion Ratio Lower Upper
93 100
63 396.1 58.7 98.7#
95 244.1 12.9 52.9#



#15
Benzyl Alcohol
Concen: 0.71 ng
RT: 7.64 min Scan# 565
Delta R.T. 0.03 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion: 79 Resp: 2334
Ion Ratio Lower Upper
79 100
108 19.8 68.4 102.6#
77 97.7 50.0 75.0#









#18
Hexachloroethane
Concen: 0.33 ng
RT: 8.02 min Scan# 598
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:117 Resp: 427
Ion Ratio Lower Upper
117 100
119 223.8 75.5 113.3#
201 0.0 66.4 99.6#



#20
3+4-Methylphenols
Concen: 0.06 ng
RT: 7.86 min Scan# 584
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:107 Resp: 224
Ion Ratio Lower Upper
107 100
108 0.0 75.2 115.2#
77 0.0 10.3 50.3#
79 0.0 9.3 49.3#





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.79 min Scan# 665

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)

Tot Ion:136 Resp: 214669


Ion Ratio Lower Upper

136 100

137 10.4 8.8 13.2

54 6.0 8.7 13.1#

68 4.6 5.8 8.8#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#22

Acetophenone

Concen: 0.27 ng

RT: 7.88 min Scan# 586

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Tgt Ion:105 Resp: 1402


Ion Ratio Lower Upper

105 100

71 0.0 0.9 1.3#

51 32.9 21.3 31.9#

120 0.0 17.1 25.7#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF0

Ion 51.00 (50.70 to 51.70): BF0

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 14.32 ng
 RT: 8.04 min Scan# 600
 Delta R.T. 0.00 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)

Tot Ion	82	Resp	100152
Ion Ratio	100	Lower	Upper
128	47.3	37.6	56.4
54	47.8	39.4	59.2




Abundance Ion 82.00 (81.70 to 82.70): BF0
 Ion 128.00 (127.70 to 128.70): B
 Ion 54.00 (53.70 to 54.70): BF0






#24
 Nitrobenzene
 Concen: 0.16 ng
 RT: 8.04 min Scan# 600
 Delta R.T. -0.02 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Tgt Ion	77	Resp	556
Ion Ratio	100	Lower	Upper
123	0.0	28.1	42.1#
65	0.0	10.9	16.3#



Abundance Ion 77.00 (76.70 to 77.70): BF0
 Ion 123.00 (122.70 to 123.70): B
 Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 0.05 ng

RT: 8.31 min Scan# 623

Delta R.T. 0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion: 82 Resp: 332

Ion Ratio Lower Upper

82 100

95 110.3 5.3 7.9#

138 0.0 13.0 19.4#






Abundance Ion 82.00 (81.70 to 82.70): BF079741.D (-619) (-)

Ion 95.00 (94.70 to 95.70): BF079741.D (-619) (-)

Ion 138.00 (137.70 to 138.70): BF079741.D (-619) (-)

Time-->



#27

2,4-Dimethylphenol

Concen: 0.06 ng

RT: 8.42 min Scan# 633

Delta R.T. 0.02 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion: 122 Resp: 206

Ion Ratio Lower Upper

122 100

107 0.0 88.5 132.7#

121 0.0 46.5 69.7#






Abundance Ion 122.00 (121.70 to 122.70): BF079741.D (-627) (-)

Ion 107.00 (106.70 to 107.70): BF079741.D (-627) (-)

Ion 121.00 (120.70 to 121.70): BF079741.D (-627) (-)

Time-->









#31
Naphthalene
Concen: 1.60 ng
RT: 8.80 min Scan# 666
Delta R.T. -0.01 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)

Tot Ion:128 Resp: 18000
Ion Ratio Lower Upper
128 100
129 12.6 9.4 14.2
127 11.9 10.7 16.1




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 4.12 ng
RT: 8.42 min Scan# 633
Delta R.T. -0.05 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:122 Resp: 206
Ion Ratio Lower Upper
122 100
105 192.0 106.7 146.7#
77 169.3 68.9 108.9#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 0.37 ng
RT: 8.80 min Scan# 666
Delta R.T. -0.03 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:127	Resp:	2100
Ion Ratio	Lower	Upper
127	100	
129	105.6	26.8 40.2#
65	20.3	22.9 34.3#
92	0.0	15.2 22.8#



#35
Caprolactam
Concen: 0.71 ng
RT: 9.19 min Scan# 700
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:113	Resp:	644
Ion Ratio	Lower	Upper
113	100	
55	394.5	186.6 226.6#
56	244.2	137.8 177.8#





#36

4-Chloro-3-methylphenol

Concen: 0.38 ng

RT: 9.32 min Scan# 712

Delta R.T. 0.02 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:107 Resp: 1247

Ion Ratio Lower Upper

107 100

144 0.0 24.5 36.7#




142 0.0 74.6 112.0#




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 10.56 min Scan# 820
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:164 Resp: 121954
Ion Ratio Lower Upper
164 100
162 103.9 82.6 123.8
160 45.5 36.8 55.2



#41
2,4,6-Tribromophenol
Concen: 20.43 ng
RT: 11.36 min Scan# 890
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:330 Resp: 47662
Ion Ratio Lower Upper
330 100
332 98.5 77.4 116.0
141 58.3 30.2 45.2#









#43
2,4,5-Trichlorophenol
Concen: 0.10 ng
RT: 9.95 min Scan# 767
Delta R.T. 0.14 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:196 Resp: 270
Ion Ratio Lower Upper
196 100
198 0.0 76.9 115.3#
97 4135.1 38.9 58.3#
132 107.9 23.9 35.9#



#44
2-Fluorobiphenyl
Concen: 13.18 ng
RT: 9.86 min Scan# 759
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:172 Resp: 237232
Ion Ratio Lower Upper
172 100
171 34.8 29.4 44.0
170 23.1 19.7 29.5





#45

1,1'-Biphenyl

Concen: 1.04 ng

RT: 9.96 min Scan# 768

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:154 Resp: 11031

Ion Ratio Lower Upper

154 100

153 45.5 19.9 59.9




76 12.0 0.0 28.2



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 0.11 ng

RT: 9.98 min Scan# 769

Delta R.T. -0.02 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:162 Resp: 956

Ion Ratio Lower Upper

162 100

127 61.4 35.4 53.2#



164 44.9 26.3 39.5#




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 0.72 ng
RT: 10.07 min Scan# 777
Delta R.T. -0.01 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion: 65 Resp: 1315
Ion Ratio Lower Upper
65 100
92 53.3 49.0 73.6
138 31.9 67.7 101.5#



#48
Acenaphthylene
Concen: 1.18 ng
RT: 10.42 min Scan# 808
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion: 152 Resp: 16889
Ion Ratio Lower Upper
152 100
151 30.4 16.2 24.4#
153 24.9 10.4 15.6#





#49

Dimethylphthalate

Concen: 2.88 ng

RT: 10.24 min Scan# 792

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)





Tot Ion:163 Resp: 27058

Ion Ratio Lower Upper

163 100

194 4.5 2.6 3.8#

164 11.8 8.5 12.7



#50

2,6-Dinitrotoluene

Concen: 1.79 ng

RT: 10.32 min Scan# 799

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58




Tgt Ion:165 Resp: 959


Ion Ratio Lower Upper

165 100

63 176.7 45.2 67.8#

89 66.1 45.9 68.9





#51

Acenaphthene

Concen: 1.23 ng

RT: 10.59 min Scan# 823

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)





Tot Ion:154 Resp: 9998

Ion Ratio Lower Upper

154 100

153 140.6 90.1 135.1#

152 101.3 43.4 65.0#



#52

3-Nitroaniline

Concen: 1.21 ng

RT: 10.51 min Scan# 816

Delta R.T. 0.02 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58




Tgt Ion:138 Resp: 2404


Ion Ratio Lower Upper

138 100

108 169.0 8.9 13.3#

92 131.2 99.0 148.4









#54
Dibenzo(furan)
Concen: 2.11 ng
RT: 10.76 min Scan# 838
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion	Ratio	Lower	Upper
168	100		
139	40.2	27.7	41.5
169	26.2	11.0	16.4



#55
4-Nitrophenol
Concen: 8.73 ng
RT: 10.65 min Scan# 828
Delta R.T. 0.02 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion	Ratio	Lower	Upper
139	100		
109	61.4	44.2	84.2
65	34.9	72.3	112.3





#56

2,4-Dinitrotoluene

Concen: 3.02 ng

RT: 10.72 min Scan# 834

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)

Tot Ion:165 Resp: 2044



Ion Ratio Lower Upper

165 100

63 28.9 24.7 37.1

89 29.5 45.7 68.5#

182 86.5 1.8 2.8#





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E



#57

Fluorene

Concen: 1.65 ng

RT: 11.11 min Scan# 868

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58



Tgt Ion:166 Resp: 16362

Ion Ratio Lower Upper

166 100

165 142.8 79.4 119.0#


167 82.6 11.0 16.4#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#59

Diethylphthalate

Concen: 0.13 ng

RT: 10.95 min Scan# 854

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:149 Resp: 1171

Ion Ratio Lower Upper

149 100

177 63.5 18.1 27.1#




150 43.5 10.5 15.7#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 0.10 ng

RT: 11.07 min Scan# 865

Delta R.T. -0.02 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:204 Resp: 441

Ion Ratio Lower Upper

204 100

206 34.3 25.8 38.6



141 563.4 56.5 84.7#




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 0.81 ng

RT: 11.11 min Scan# 868

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)





Tot Ion:138 Resp: 1763

Ion Ratio Lower Upper

138 100

92 177.5 38.2 78.2#

108 170.5 53.8 93.8#



#62

Azobenzene

Concen: 0.55 ng

RT: 11.27 min Scan# 882

Delta R.T. 0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Tgt Ion: 77 Resp: 5070




Ion Ratio Lower Upper


77 100

182 447.6 0.0 35.2#

105 87.5 0.0 39.5#

51 45.8 12.4 52.4









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 12.07 min Scan# 952
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:188 Resp: 236915
Ion Ratio Lower Upper
188 100
94 10.9 5.0 7.4#
80 12.3 5.3 7.9#



#64
4,6-Dinitro-2-methylphenol
Concen: 5.34 ng
RT: 11.14 min Scan# 871
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:198 Resp: 2048
Ion Ratio Lower Upper
198 100
51 190.6 34.4 74.4#
105 468.2 32.7 72.7#





#65

n-Nitrosodiphenylamine

Concen: 6.60 ng

RT: 11.22 min Scan# 878

Delta R.T. 0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)





Tot Ion:169 Resp: 52252

Ion Ratio Lower Upper

169 100

168 11.1 54.8 82.2#

167 13.9 29.7 44.5#



#66

4-Bromophenyl-phenylether

Concen: 0.08 ng

RT: 11.67 min Scan# 917

Delta R.T. 0.07 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58




Tgt Ion:248 Resp: 211


Ion Ratio Lower Upper

248 100

250 0.0 76.1 114.1#

141 1735.5 64.6 96.8#









#68
Atrazine
Concen: 0.56 ng
RT: 11.71 min Scan# 921
Delta R.T. -0.01 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:200 Resp: 1320
Ion Ratio Lower Upper
200 100
173 165.4 5.5 45.5#
215 215.0 28.5 68.5#



#70
Phenanthrene
Concen: 30.57 ng
RT: 12.09 min Scan# 954
Delta R.T. -0.01 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:178 Resp: 424044
Ion Ratio Lower Upper
178 100
176 20.1 16.2 24.4
179 16.6 12.6 19.0









#71
 Anthracene
 Concen: 31.00 ng
 RT: 12.09 min Scan# 954
 Delta R.T. -0.06 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58




Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)


Tot Ion:178 Resp: 424044
 Ion Ratio Lower Upper
 178 100
 176 20.1 14.8 22.2
 179 16.6 12.2 18.2



#72
 Carbazole
 Concen: 0.41 ng
 RT: 12.30 min Scan# 972
 Delta R.T. 0.00 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Tgt Ion:167 Resp: 5356
 Ion Ratio Lower Upper
 167 100
 166 89.9 17.7 26.5#
 139 40.5 9.3 13.9#





#73

Di-n-butylphthalate

Concen: 0.22 ng

RT: 12.62 min Scan# 1000

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:149 Resp: 3205

Ion Ratio Lower Upper

149 100

150 163.5 9.8 14.8#




104 23.6 12.9 19.3#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 6.32 ng

RT: 13.35 min Scan# 1064

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:202 Resp: 96004

Ion Ratio Lower Upper

202 100

101 12.3 0.0 31.3



203 22.3 0.0 38.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 15.01 min Scan# 1209

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:240 Resp: 264207

Ion Ratio Lower Upper

240 100

120 12.1 6.2 9.2#




236 26.6 21.0 31.4



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 0.16 ng

RT: 13.45 min Scan# 1073

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:184 Resp: 1293

Ion Ratio Lower Upper

184 100

185 74.3 12.2 18.2#



183 127.5 9.3 13.9#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pvrene

Concen: 29.13 ng

RT: 13.61 min Scan# 1087

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:202 Resp: 471809

Ion Ratio Lower Upper

202 100

200 21.6 17.5 26.3




203 20.9 14.8 22.2



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 11.84 ng

RT: 13.74 min Scan# 1098

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:244 Resp: 273258

Ion Ratio Lower Upper

244 100

212 7.5 6.1 9.1



122 14.3 9.3 13.9#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 0.14 ng
RT: 14.29 min Scan# 1146
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion:149 Resp: 881
Ion Ratio Lower Upper
149 100
91 154.3 64.4 96.6#
206 91.6 14.8 22.2#



#80
Benzo(a)anthracene
Concen: 9.95 ng
RT: 14.99 min Scan# 1208
Delta R.T. -0.01 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion:228 Resp: 161001
Ion Ratio Lower Upper
228 100
226 33.5 22.2 33.4#
229 38.4 16.6 25.0#









#81
 3,3'-Dichlorobenzidine
 Concen: 0.29 ng
 RT: 14.91 min Scan# 1201
 Delta R.T. -0.02 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58




Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)


Tot Ion:252 Resp: 1614
 Ion Ratio Lower Upper
 252 100
 254 58.2 52.0 78.0
 126 63.7 6.9 10.3#



#82
 Chrysene
 Concen: 10.18 ng
 RT: 15.04 min Scan# 1212
 Delta R.T. -0.01 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Tgt Ion:228 Resp: 152171
 Ion Ratio Lower Upper
 228 100
 226 31.9 25.2 37.8
 229 28.2 16.2 24.4#









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 0.38 ng
 RT: 14.95 min Scan# 1204
 Delta R.T. -0.01 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58




Instrument :
 BNA_F
 ClientSampleId :
 RT440NB6.3(2)


Tot Ion:149 Resp: 3876
 Ion Ratio Lower Upper
 149 100
 167 49.0 20.9 31.3#
 279 15.1 2.2 3.2#



#84
 Di-n-octyl phthalate
 Concen: 0.09 ng
 RT: 15.74 min Scan# 1273
 Delta R.T. -0.01 min
 Lab File: BF079757.D
 Acq: 6 Jun 2015 5:58

Tgt Ion:149 Resp: 1568
 Ion Ratio Lower Upper
 149 100
 167 43.1 1.3 1.9#
 43 0.0 9.0 13.4#









#85
Indeno(1,2,3-cd)pyrene
Concen: 2.52 ng
RT: 18.95 min Scan# 1554
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58




Instrument :
BNA_F
ClientSampleId :
RT440NB6.3(2)


Tot Ion	Ratio	Lower	Upper
276	100		
138	27.0	20.8	31.2
277	28.8	20.0	30.0



#86
Perylene-d12
Concen: 20.00 ng
RT: 16.95 min Scan# 1379
Delta R.T. 0.00 min
Lab File: BF079757.D
Acq: 6 Jun 2015 5:58

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.7	18.6	27.8
265	24.2	18.2	27.4





#87

Benzo(b)fluoranthene

Concen: 5.87 ng

RT: 16.37 min Scan# 1328

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:252 Resp: 93235

Ion Ratio Lower Upper

252 100

253 30.1 17.4 26.2#




125 17.7 6.5 9.7#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 6.12 ng

RT: 16.37 min Scan# 1328

Delta R.T. -0.05 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:252 Resp: 93235

Ion Ratio Lower Upper

252 100

253 30.1 17.8 26.8#



125 17.7 8.4 12.6#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 7.85 ng

RT: 16.87 min Scan# 1372

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:252 Resp: 115155

Ion Ratio Lower Upper

252 100

253 26.1 17.6 26.4




125 12.3 7.4 11.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 1.52 ng

RT: 18.97 min Scan# 1556

Delta R.T. 0.00 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58


Tgt Ion:278 Resp: 21504

Ion Ratio Lower Upper

278 100

139 27.7 12.7 19.1#



279 52.5 19.4 29.2#




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(a,h,i)perylene

Concen: 10.27 ng

RT: 19.55 min Scan# 1607

Delta R.T. -0.01 min

Lab File: BF079757.D

Acq: 6 Jun 2015 5:58

Instrument :

BNA_F

ClientSampleId :

RT440NB6.3(2)


Tot Ion:276 Resp: 150183

Ion Ratio Lower Upper

276 100

277 25.9 18.8 28.2

138 25.1 18.2 27.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

