

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060515\
 Data File : BF079763.D
 Acq On : 6 Jun 2015 8:51
 Operator : TP/IZ
 Sample : G2464-12 2X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-16AS

Manual Integrations
 APPROVED

apatel
 6/9/2015 8:11:11 AM

Quant Time: Jun 08 16:55:23 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 15:29:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	53636	20.00	ng	0.00
21) Naphthalene-d8	8.79	136	224095	20.00	ng	0.00
38) Acenaphthene-d10	10.56	164	123496	20.00	ng	0.00
63) Phenanthrene-d10	12.08	188	254750	20.00	ng	0.01
75) Chrysene-d12	15.07	240	262818	20.00	ng	0.14
86) Perylene-d12	17.05	264	257052	20.00	ng	0.23
System Monitoring Compounds						
5) 2-Fluorophenol	6.10	112	174221	54.44	ng	0.00
7) Phenol-d6	7.08	99	248208	60.05	ng	0.00
23) Nitrobenzene-d5	8.04	82	125539	34.39	ng	0.00
41) 2,4,6-Tribromophenol	11.36	330	64234	54.38	ng	0.00
44) 2-Fluorobiphenyl	9.86	172	298289	32.72	ng	0.00
78) Terphenyl-d14	13.78	244	369683	32.22	ng	0.07
Target Compounds						
49) Dimethylphthalate	10.24	163	64456	6.78	ng	99
70) Phenanthrene	12.10	178	269894	18.10	ng	98
71) Anthracene	12.16	178	77514	5.27	ng	98
74) Fluoranthene	13.38	202	551931	33.80	ng	96
77) Pyrene	13.65	202	546786	33.94	ng	99
80) Benzo(a)anthracene	15.06	228	303972	18.89	ng	98
82) Chrysene	15.11	228	254296m	17.10	ng	
85) Indeno(1,2,3-cd)pyrene	19.05	276	164324	9.55	ng	97
87) Benzo(b)fluoranthene	16.47	252	373696m	23.40	ng	
88) Benzo(k)fluoranthene	16.50	252	122129m	7.97	ng	
89) Benzo(a)pyrene	16.96	252	277801	18.85	ng	# 93
90) Dibenzo(a,h)anthracene	19.07	278	41348	2.91	ng	# 89
91) Benzo(a,h,i)perylene	19.67	276	170588	11.60	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060515\
Data File : BF079763.D
Acq On : 6 Jun 2015 8:51
Operator : TP/IZ
Sample : G2464-12 2X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-16AS

Manual Integrations
APPROVED

apatel
6/9/2015 8:11:11 AM

Quant Time: Jun 08 16:55:23 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 15:29:07 2015
Response via : Initial Calibration

