

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060515\
 Data File : BF079765.D
 Acq On : 6 Jun 2015 9:49
 Operator : TP/IZ
 Sample : G2499-01 2X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FK-SF-02-002-A

Manual Integrations
 APPROVED

apatel
 6/9/2015 8:11:21 AM

Quant Time: Jun 08 17:09:55 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 08 15:29:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	53094	20.00	ng	0.00
21) Naphthalene-d8	8.79	136	223107	20.00	ng	0.00
38) Acenaphthene-d10	10.56	164	112609	20.00	ng	0.00
63) Phenanthrene-d10	12.09	188	247324	20.00	ng	0.02
75) Chrysene-d12	15.23	240	267701m	20.00	ng	0.30
86) Perylene-d12	17.28	264	245997m	20.00	ng	0.46
System Monitoring Compounds						
5) 2-Fluorophenol	6.10	112	55041	17.38	ng	0.00
7) Phenol-d6	7.08	99	210735	51.51	ng	0.00
23) Nitrobenzene-d5	8.04	82	112064	30.84	ng	0.00
41) 2,4,6-Tribromophenol	11.36	330	2358	2.19	ng	0.00
44) 2-Fluorobiphenyl	9.86	172	280530	33.75	ng	0.00
78) Terphenyl-d14	13.86	244	362634	31.02	ng	0.15
Target Compounds						Qvalue
15) Benzyl Alcohol	7.61	79	96104	28.42	ng	94
49) Dimethylphthalate	10.24	163	54640	6.30	ng	99
70) Phenanthrene	12.11	178	190979	13.19	ng	98
71) Anthracene	12.17	178	56504	3.96	ng	99
74) Fluoranthene	13.44	202	240119	15.15	ng	96
77) Pyrene	13.71	202	208722	12.72	ng	98
79) Butylbenzylphthalate	14.45	149	24945m	3.81	ng	
80) Benzo(a)anthracene	15.21	228	117886m	7.19	ng	
82) Chrysene	15.27	228	110217m	7.28	ng	
83) Bis(2-ethylhexyl)phthalate	15.19	149	69062m	6.68	ng	
85) Indeno(1,2,3-cd)pyrene	19.30	276	54854m	3.13	ng	
87) Benzo(b)fluoranthene	16.69	252	129790m	8.49	ng	
88) Benzo(k)fluoranthene	16.72	252	39172m	2.67	ng	
89) Benzo(a)pyrene	17.19	252	88390m	6.27	ng	
91) Benzo(a,h,i)perylene	19.91	276	52164m	3.71	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060515\
Data File : BF079765.D
Acq On : 6 Jun 2015 9:49
Operator : TP/IZ
Sample : G2499-01 2X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FK-SF-02-002-A

Manual Integrations
APPROVED

apatel
6/9/2015 8:11:21 AM

Quant Time: Jun 08 17:09:55 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 08 15:29:07 2015
Response via : Initial Calibration

