





Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

2,6-Dinitrotoluene

Response Ratio




$R = 2.44e-002 A^2 + 2.01e-001 A - 1.03e-002$
Coef of Det (r^2) = 0.998 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

2,4-Dinitrophenol

Response Ratio




$R = 1.19e-002 A^2 + 3.90e-002 A - 3.89e-003$
Coef of Det (r^2) = 1.000 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

2,4-Dinitrotoluene

Response Ratio




$R = 3.09e-002 A^2 + 2.41e-001 A - 2.04e-002$
Coef of Det (r^2) = 0.999 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

4-Nitroaniline


Response Ratio



$R = 4.09e-002 A^2 + 2.29e-001 A + 5.08e-003$
Coef of Det (r^2) = 1.000 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

4,6-Dinitro-2-methylphenol



$R = 1.11e-002 A^2 + 2.51e-002 A + 1.16e-003$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic


Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015



Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Fri Jun 12 16:20:49 2015

2-Nitroaniline

Response Ratio



Resp Ratio = $3.74 \times 10^{-1} \times \text{Amt} - 1.22 \times 10^{-1}$
Coef of Det (r^2) = 0.992 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
Calibration Table Last Updated: Mon Jun 15 11:19:44 2015