

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060516\
 Data File : BF087734.D
 Acq On : 6 Jun 2016 4:21
 Operator : UM/SJ
 Sample : H3397-02
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EFFWW

Quant Time: Jun 06 05:10:27 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 04 11:07:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	106410	20.00	ng	-0.01
21) Naphthalene-d8	8.15	136	389202	20.00	ng	0.00
38) Acenaphthene-d10	9.91	164	196440	20.00	ng	0.00
63) Phenanthrene-d10	11.38	188	283499	20.00	ng	0.00
75) Chrysene-d12	14.03	240	220131	20.00	ng	0.01
86) Perylene-d12	15.46	264	217135	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	77331	11.75	ng	0.01
7) Phenol-d6	6.51	99	244093	29.56	ng	0.02
23) Nitrobenzene-d5	7.43	82	570585	84.82	ng	0.00
41) 2,4,6-Tribromophenol	10.70	330	159814	81.05	ng	0.01
44) 2-Fluorobiphenyl	9.23	172	1127884	89.76	ng	0.00
78) Terphenyl-d14	12.99	244	294412	32.64	ng	0.02
Target Compounds						
15) Benzyl Alcohol	7.04	79	48736	7.99	ng	Qvalue 97
22) Acetophenone	7.28	105	25013	2.84	ng	# 95
45) 1,1'-Biphenyl	9.32	154	33634	2.11	ng	97
49) Dimethylphthalate	9.62	163	29892	2.09	ng	99
54) Dibenzofuran	9.94	168	146003	8.42	ng	# 64
84) Di-n-octyl phthalate	14.71	149	967369	65.54	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060516\
Data File : BF087734.D
Acq On : 6 Jun 2016 4:21
Operator : UM/SJ
Sample : H3397-02
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EFFWW

Quant Time: Jun 06 05:10:27 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 04 11:07:28 2016
Response via : Initial Calibration

