

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060516\
 Data File : BF087714.D
 Acq On : 5 Jun 2016 18:21
 Operator : UM/SJ
 Sample : H3362-07
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP16-LF3MW4

Quant Time: Jun 06 04:10:54 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 04 11:07:28 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	147872	20.00	ng	-0.01
21) Naphthalene-d8	8.15	136	635718	20.00	ng	0.00
38) Acenaphthene-d10	9.90	164	278115	20.00	ng	-0.01
63) Phenanthrene-d10	11.38	188	562361	20.00	ng	0.00
75) Chrysene-d12	14.01	240	404339	20.00	ng	-0.01
86) Perylene-d12	15.44	264	315765	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	578635	63.25	ng	0.01
7) Phenol-d6	6.48	99	412111	35.92	ng	-0.01
23) Nitrobenzene-d5	7.43	82	857445	78.04	ng	0.00
41) 2,4,6-Tribromophenol	10.68	330	314760	112.75	ng	0.00
44) 2-Fluorobiphenyl	9.22	172	1472549	79.64	ng	-0.01
78) Terphenyl-d14	12.97	244	1113820	67.22	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060516\
Data File : BF087714.D
Acq On : 5 Jun 2016 18:21
Operator : UM/SJ
Sample : H3362-07
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP16-LF3MW4

Quant Time: Jun 06 04:10:54 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 04 11:07:28 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.86 min Scan# 461

Delta R.T. -0.01 min

Lab File: BF087714.D

Acq: 5 Jun 2016 18:21

Instrument :

BNA_F

ClientSampleId :

SP16-LF3MW4


Tot Ion:152 Resp: 147872

Ion Ratio Lower Upper

152 100

150 157.9 123.8 185.6




115 66.0 39.6 59.4#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 63.25 ng

RT: 5.46 min Scan# 339

Delta R.T. 0.01 min

Lab File: BF087714.D

Acq: 5 Jun 2016 18:21


Tgt Ion:112 Resp: 578635

Ion Ratio Lower Upper

112 100

64 44.4 42.2 63.2



63 22.8 21.8 32.8




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 35.92 ng

RT: 6.48 min Scan# 428

Delta R.T. -0.01 min

Lab File: BF087714.D

Acq: 5 Jun 2016 18:21

Instrument :

BNA_F

ClientSampleId :

SP16-LF3MW4


Tot Ion: 99 Resp: 412111

Ion Ratio Lower Upper

99 100

42 18.9 12.2 18.2#




71 33.4 23.4 35.0



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.15 min Scan# 574

Delta R.T. -0.00 min

Lab File: BF087714.D

Acq: 5 Jun 2016 18:21

Tgt Ion: 136 Resp: 635718


Ion Ratio Lower Upper

136 100

137 10.8 9.1 13.7

54 6.2 6.6 10.0#

68 5.2 5.1 7.7






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 78.04 ng
 RT: 7.43 min Scan# 511
 Delta R.T. -0.00 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21

Instrument :
 BNA_F
 ClientSampleId :
 SP16-LF3MW4

Tot Ion: 82 Resp: 857445
 Ion Ratio Lower Upper
 82 100
 128 48.6 35.9 53.9
 54 49.1 40.9 61.3




Abundance Ion 82.00 (81.70 to 82.70): BF0
 Ion 128.00 (127.70 to 128.70): B
 Ion 54.00 (53.70 to 54.70): BF0






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.90 min Scan# 727
 Delta R.T. -0.01 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21

Tgt Ion:164 Resp: 278115
 Ion Ratio Lower Upper
 164 100
 162 104.1 85.4 128.2
 160 47.0 39.5 59.3



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 112.75 ng
 RT: 10.68 min Scan# 796
 Delta R.T. -0.00 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21




Instrument :
 BNA_F
 ClientSampleId :
 SP16-LF3MW4


Tot Ion:330 Resp: 314760
 Ion Ratio Lower Upper
 330 100
 332 95.6 0.0 0.0#
 141 32.5 0.0 0.0#



#44
 2-Fluorobiphenyl
 Concen: 79.64 ng
 RT: 9.22 min Scan# 668
 Delta R.T. -0.01 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21

Tgt Ion:172 Resp: 1472549
 Ion Ratio Lower Upper
 172 100
 171 35.9 28.6 43.0
 170 24.1 19.2 28.8









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.38 min Scan# 857
Delta R.T. -0.00 min
Lab File: BF087714.D
Acq: 5 Jun 2016 18:21

Instrument :
BNA_F
ClientSampleId :
SP16-LF3MW4

Tot Ion:188 Resp: 562361
Ion Ratio Lower Upper
188 100
94 9.6 9.0 13.6
80 9.6 10.0 15.0#




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#75
Chrysene-d12
Concen: 20.00 ng
RT: 14.01 min Scan# 1087
Delta R.T. -0.01 min
Lab File: BF087714.D
Acq: 5 Jun 2016 18:21

Tgt Ion:240 Resp: 404339
Ion Ratio Lower Upper
240 100
120 14.5 6.8 10.2#
236 26.4 20.2 30.2



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#78
 Terphenyl-d14
 Concen: 67.22 ng
 RT: 12.97 min Scan# 996
 Delta R.T. 0.00 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21

Instrument :
 BNA_F
 ClientSampled :
 SP16-LF3MW4

Tot Ion:244 Resp: 1113820
 Ion Ratio Lower Upper
 244 100
 212 7.9 6.0 9.0
 122 13.6 11.2 16.8




Abundance Ion 244.00 (243.70 to 244.70): E
 Ion 212.00 (211.70 to 212.70): E
 Ion 122.00 (121.70 to 122.70): E



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 15.44 min Scan# 1212
 Delta R.T. -0.00 min
 Lab File: BF087714.D
 Acq: 5 Jun 2016 18:21

Tgt Ion:264 Resp: 315765
 Ion Ratio Lower Upper
 264 100
 260 23.8 19.2 28.8
 265 21.7 16.9 25.3



Abundance Ion 264.00 (263.70 to 264.70): E
 Ion 260.00 (259.70 to 260.70): E
 Ion 265.00 (264.70 to 265.70): E

