

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060524\
 Data File : BF138125.D
 Acq On : 05 Jun 2024 17:06
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF060524

Quant Time: Jun 06 05:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	38.897	2.8	114	0.02
3	Pyridine	40.000	40.123	-0.3	116	0.02
4	n-Nitrosodimethylamine	40.000	39.925	0.2	115	0.03
5 S	2-Fluorophenol	80.000	78.443	1.9	113	0.00
6	Aniline	40.000	40.233	-0.6	116	0.00
7 S	Phenol-d6	80.000	80.682	-0.9	117	0.00
8	2-Chlorophenol	40.000	40.697	-1.7	117	0.00
9	Benzaldehyde	40.000	48.342	-20.9	126	0.00
10 C	Phenol	40.000	41.227	-3.1	119	0.00
11	bis(2-Chloroethyl)ether	40.000	39.648	0.9	114	0.00
12	1,3-Dichlorobenzene	40.000	39.471	1.3	114	0.00
13 C	1,4-Dichlorobenzene	40.000	39.476	1.3	114	0.00
14	1,2-Dichlorobenzene	40.000	39.377	1.6	113	0.00
15	Benzyl Alcohol	40.000	41.396	-3.5	117	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.135	2.2	113	0.00
17	2-Methylphenol	40.000	40.521	-1.3	117	0.00
18	Hexachloroethane	40.000	41.105	-2.8	117	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.119	2.2	115	0.00
20	3+4-Methylphenols	40.000	40.840	-2.1	115	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	38.293	4.3	113	0.00
23 S	Nitrobenzene-d5	80.000	94.171	-17.7	129	0.00
24	Nitrobenzene	40.000	44.706	-11.8	125	0.00
25	Isophorone	40.000	39.398	1.5	117	0.00
26 C	2-Nitrophenol	40.000	48.149	-20.4#	163	0.00
27	2,4-Dimethylphenol	40.000	41.480	-3.7	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.419	1.5	116	0.00
29 C	2,4-Dichlorophenol	40.000	42.165	-5.4	120	0.00
30	1,2,4-Trichlorobenzene	40.000	39.780	0.5	117	0.00
31	Naphthalene	40.000	38.889	2.8	115	0.00
32	Benzoic acid	40.000	50.093	-25.2#	187	0.01
33	4-Chloroaniline	40.000	39.724	0.7	118	0.00
34 C	Hexachlorobutadiene	40.000	39.797	0.5	116	0.00
35	Caprolactam	40.000	40.498	-1.2	118	0.01
36 C	4-Chloro-3-methylphenol	40.000	41.349	-3.4	119	0.00
37	2-Methylnaphthalene	40.000	39.005	2.5	115	0.00
38	1-Methylnaphthalene	40.000	38.890	2.8	114	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.962	0.1	117	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.036	-2.6	113	0.00
42 S	2,4,6-Tribromophenol	80.000	86.893	-8.6	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.061	-10.2	124	0.00
44	2,4,5-Trichlorophenol	40.000	42.692	-6.7	120	0.00
45 S	2-Fluorobiphenyl	80.000	74.736	6.6	112	0.00
46	1,1'-Biphenyl	40.000	38.427	3.9	114	0.00
47	2-Chloronaphthalene	40.000	39.082	2.3	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060524\
 Data File : BF138125.D
 Acq On : 05 Jun 2024 17:06
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF060524

Quant Time: Jun 06 05:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.317	-8.3	136	0.00
49	Acenaphthylene	40.000	38.634	3.4	114	0.00
50	Dimethylphthalate	40.000	38.958	2.6	116	0.00
51	2,6-Dinitrotoluene	40.000	43.167	-7.9	131	0.00
52 C	Acenaphthene	40.000	39.197	2.0	117	0.00
53	3-Nitroaniline	40.000	42.929	-7.3	130	0.00
54 P	2,4-Dinitrophenol	40.000	48.906	-22.3	190	0.00
55	Dibenzofuran	40.000	38.511	3.7	114	0.00
56 P	4-Nitrophenol	40.000	46.216	-15.5	129	0.00
57	2,4-Dinitrotoluene	40.000	44.446	-11.1	134	0.00
58	Fluorene	40.000	38.184	4.5	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.274	-10.7	121	0.00
60	Diethylphthalate	40.000	38.638	3.4	113	0.00
61	4-Chlorophenyl-phenylether	40.000	38.006	5.0	113	0.00
62	4-Nitroaniline	40.000	46.527	-16.3	122	0.00
63	Azobenzene	40.000	38.697	3.3	114	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.895	-17.2	159	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.184	-0.5	113	0.00
67	4-Bromophenyl-phenylether	40.000	41.067	-2.7	115	0.00
68	Hexachlorobenzene	40.000	40.099	-0.2	114	0.00
69	Atrazine	40.000	39.208	2.0	111	0.00
70 C	Pentachlorophenol	40.000	47.929	-19.8	123	0.00
71	Phenanthrene	40.000	38.402	4.0	109	0.00
72	Anthracene	40.000	38.720	3.2	111	0.00
73	Carbazole	40.000	38.578	3.6	110	0.00
74	Di-n-butylphthalate	40.000	38.552	3.6	108	0.00
75 C	Fluoranthene	40.000	37.445	6.4	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	40.289	-0.7	76	0.00
78	Pyrene	40.000	42.015	-5.0	105	0.00
79 S	Terphenyl-d14	80.000	80.035	-0.0	105	0.00
80	Butylbenzylphthalate	40.000	47.145	-17.9	109	0.00
81	Benzo(a)anthracene	40.000	39.065	2.3	101	0.00
82	3,3'-Dichlorobenzidine	40.000	40.745	-1.9	103	0.00
83	Chrysene	40.000	39.746	0.6	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.325	-10.8	107	0.00
85 c	Di-n-octyl phthalate	40.000	45.161	-12.9	111	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.460	-3.7	108	0.00
88	Benzo(b)fluoranthene	40.000	40.789	-2.0	110	0.00
89	Benzo(k)fluoranthene	40.000	38.655	3.4	101	0.00
90 C	Benzo(a)pyrene	40.000	41.206	-3.0	105	0.00
91	Dibenzo(a,h)anthracene	40.000	42.392	-6.0	110	0.00
92	Benzo(g,h,i)perylene	40.000	40.288	-0.7	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060524\
Data File : BF138125.D
Acq On : 05 Jun 2024 17:06
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF060524

Quant Time: Jun 06 05:23:23 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 06 05:19:29 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1