

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060524\
 Data File : BF138125.D
 Acq On : 05 Jun 2024 17:06
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF060524

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024

Quant Time: Jun 06 05:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	490377	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	1912746	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	1086256	20.000	ng	0.00
64) Phenanthrene-d10	11.427	188	1856310	20.000	ng	0.00
76) Chrysene-d12	14.068	240	956610	20.000	ng	0.00
86) Perylene-d12	15.545	264	733906	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	2257124	78.443	ng	0.00
7) Phenol-d6	6.534	99	2937300	80.682	ng	0.00
23) Nitrobenzene-d5	7.475	82	2647447	94.171	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	908362	86.893	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	4642880	74.736	ng	0.00
79) Terphenyl-d14	13.015	244	5118700	80.035	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.751	88	512968	38.897	ng	98
3) Pyridine	3.504	79	1296545	40.123	ng	97
4) n-Nitrosodimethylamine	3.475	42	633367	39.925	ng	96
6) Aniline	6.575	93	1485110	40.233	ng	100
8) 2-Chlorophenol	6.692	128	1266877	40.697	ng	99
9) Benzaldehyde	6.463	77	825433	48.342	ng	96
10) Phenol	6.551	94	1542021	41.227	ng	98
11) bis(2-Chloroethyl)ether	6.645	93	1188833	39.648	ng	98
12) 1,3-Dichlorobenzene	6.851	146	1407819	39.471	ng	99
13) 1,4-Dichlorobenzene	6.928	146	1418022	39.476	ng	100
14) 1,2-Dichlorobenzene	7.081	146	1327751	39.377	ng	99
15) Benzyl Alcohol	7.051	79	1066341	41.396	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.181	45	1722669	39.135	ng	99
17) 2-Methylphenol	7.151	107	1020961	40.521	ng	98
18) Hexachloroethane	7.422	117	503959	41.105	ng	97
19) n-Nitroso-di-n-propyla...	7.328	70	908085	39.119	ng	97
20) 3+4-Methylphenols	7.310	107	1280438	40.840	ng	97
22) Acetophenone	7.322	105	1686221	38.293	ng	# 98
24) Nitrobenzene	7.492	77	1369130	44.706	ng	100
25) Isophorone	7.733	82	2440118	39.398	ng	98
26) 2-Nitrophenol	7.804	139	618831	48.149	ng	98
27) 2,4-Dimethylphenol	7.839	122	892366	41.480	ng	98
28) bis(2-Chloroethoxy)met...	7.939	93	1487130	39.419	ng	98
29) 2,4-Dichlorophenol	8.045	162	1093277	42.165	ng	97
30) 1,2,4-Trichlorobenzene	8.128	180	1244535	39.780	ng	100
31) Naphthalene	8.216	128	3597830	38.889	ng	99
32) Benzoic acid	7.969	122	835546m	50.093	ng	
33) 4-Chloroaniline	8.257	127	1381226	39.724	ng	100
34) Hexachlorobutadiene	8.328	225	731644	39.797	ng	99
35) Caprolactam	8.645	113	350629	40.498	ng	94
36) 4-Chloro-3-methylphenol	8.733	107	1098676	41.349	ng	99
37) 2-Methylnaphthalene	8.904	142	2432621	39.005	ng	100
38) 1-Methylnaphthalene	9.004	142	2374818	38.890	ng	100
40) 1,2,4,5-Tetrachloroben...	9.063	216	1225557	39.962	ng	100
41) Hexachlorocyclopentadiene	9.051	237	415449	41.036	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	820423	44.061	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060524\
 Data File : BF138125.D
 Acq On : 05 Jun 2024 17:06
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF060524


Quant Time: Jun 06 05:23:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	875945	42.692	ng	99
46) 1,1'-Biphenyl	9.369	154	2932004	38.427	ng	99
47) 2-Chloronaphthalene	9.392	162	2376495	39.082	ng	99
48) 2-Nitroaniline	9.486	65	646658	43.317	ng	100
49) Acenaphthylene	9.810	152	3324467	38.634	ng	99
50) Dimethylphthalate	9.669	163	2659047	38.958	ng	99
51) 2,6-Dinitrotoluene	9.727	165	633990	43.167	ng	98
52) Acenaphthene	9.980	154	2279003	39.197	ng	99
53) 3-Nitroaniline	9.898	138	631618	42.929	ng	98
54) 2,4-Dinitrophenol	9.998	184	205331	48.906	ng	99
55) Dibenzofuran	10.151	168	3167004	38.511	ng	99
56) 4-Nitrophenol	10.045	139	479785	46.216	ng	100
57) 2,4-Dinitrotoluene	10.127	165	779209	44.446	ng	96
58) Fluorene	10.492	166	2494209	38.184	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	716507	44.274	ng	99
60) Diethylphthalate	10.369	149	2574456	38.638	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	1261213	38.006	ng	99
62) 4-Nitroaniline	10.516	138	592759	46.527	ng	96
63) Azobenzene	10.645	77	2483468	38.697	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	318459	46.895	ng	78
66) n-Nitrosodiphenylamine	10.604	169	2227546	40.184	ng	100
67) 4-Bromophenyl-phenylether	10.974	248	868000	41.067	ng	99
68) Hexachlorobenzene	11.039	284	985324	40.099	ng	99
69) Atrazine	11.133	200	699276	39.208	ng	99
70) Pentachlorophenol	11.227	266	630451	47.929	ng	99
71) Phenanthrene	11.457	178	3431774	38.402	ng	99
72) Anthracene	11.510	178	3451109	38.720	ng	99
73) Carbazole	11.663	167	3144516	38.578	ng	100
74) Di-n-butylphthalate	11.992	149	3693647	38.552	ng	100
75) Fluoranthene	12.645	202	3610506	37.445	ng	99
77) Benzidine	12.757	184	396608	40.289	ng	98
78) Pyrene	12.874	202	3634887	42.015	ng	99
80) Butylbenzylphthalate	13.492	149	1346834	47.145	ng	96
81) Benzo(a)anthracene	14.057	228	2357713	39.065	ng	100
82) 3,3'-Dichlorobenzidine	14.021	252	623278	40.745	ng	97
83) Chrysene	14.092	228	2302167	39.746	ng	99
84) Bis(2-ethylhexyl)phtha...	14.051	149	1844863	44.325	ng	98
85) Di-n-octyl phthalate	14.668	149	2587328	45.161	ng	99
87) Indeno(1,2,3-cd)pyrene	17.033	276	1636174	41.460	ng	99
88) Benzo(b)fluoranthene	15.109	252	1881891	40.789	ng	98
89) Benzo(k)fluoranthene	15.145	252	1700737	38.655	ng	99
90) Benzo(a)pyrene	15.480	252	1439803	41.206	ng	99
91) Dibenzo(a,h)anthracene	17.056	278	1340135	42.392	ng	100
92) Benzo(g,h,i)perylene	17.486	276	1320045	40.288	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.910 min Scan# 81
 Delta R.T. 0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:152 Resp: 49037
 Ion Ratio Lower Upper
 152 100
 150 155.7 121.8 182.6
 115 57.2 46.2 69.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024



#2
 1,4-Dioxane
 Concen: 38.897 ng
 RT: 2.751 min Scan# 113
 Delta R.T. 0.023 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06


Tgt Ion: 88 Resp: 512968
 Ion Ratio Lower Upper
 88 100
 58 74.1 60.6 90.8
 43 28.6 23.6 35.4





#3
Pyridine
Concen: 40.123 ng
RT: 3.504 min Scan# 241
Delta R.T. 0.023 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524




Tgt Ion: 79 Resp: 129654
Ion Ratio Lower Upper
79 100
52 69.4 57.4 86.2
51 35.1 29.1 43.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#4
n-Nitrosodimethylamine
Concen: 39.925 ng
RT: 3.475 min Scan# 236
Delta R.T. 0.029 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion: 42 Resp: 633367
Ion Ratio Lower Upper
42 100
74 130.3 100.0 150.0
44 6.6 5.7 8.5





#5
2-Fluorophenol
Concen: 78.443 ng
RT: 5.528 min Scan# 585
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :





ICVBF060524

Tgt Ion:112 Resp: 2257124
Ion Ratio Lower Upper
112 100
64 58.2 44.0 66.0
63 29.1 22.1 33.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#6
Aniline
Concen: 40.233 ng
RT: 6.575 min Scan# 763
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion: 93 Resp: 1485110
Ion Ratio Lower Upper
93 100
66 38.6 30.9 46.3
65 19.7 16.0 24.0






#7
Phenol-d6
Concen: 80.682 ng
RT: 6.534 min Scan# 756
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion: 99 Resp: 2937300
Ion Ratio Lower Upper
99 100
42 18.0 14.4 21.6
71 31.9 25.5 38.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#8
2-Chlorophenol
Concen: 40.697 ng
RT: 6.692 min Scan# 783
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:128 Resp: 1266877
Ion Ratio Lower Upper
128 100
130 32.5 12.9 52.9
64 47.2 26.6 66.6









#9
Benzaldehyde
Concen: 48.342 ng
RT: 6.463 min Scan# 744
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :
BNA_F
ClientSampleId :
ICVBF060524

Tgt Ion: 77 Resp: 82543
Ion Ratio Lower Upper
77 100
105 100.2 75.8 115.8
106 96.9 73.9 113.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#10
Phenol
Concen: 41.227 ng
RT: 6.551 min Scan# 759
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion: 94 Resp: 1542021
Ion Ratio Lower Upper
94 100
65 26.7 7.9 47.9
66 34.1 15.5 55.5






#11
bis(2-Chloroethyl)ether
Concen: 39.648 ng
RT: 6.645 min Scan# 775
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion: 93 Resp: 118883

Ion	Ratio	Lower	Upper
93	100		
63	77.4	55.3	95.3
95	33.2	12.5	52.5


Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024






#12
1,3-Dichlorobenzene
Concen: 39.471 ng
RT: 6.851 min Scan# 810
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:146 Resp: 1407819

Ion	Ratio	Lower	Upper
146	100		
148	64.4	52.2	78.2
75	28.0	23.0	34.6





#13
1,4-Dichlorobenzene
Concen: 39.476 ng
RT: 6.928 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F





ClientSampleId :

ICVBF060524

Tgt Ion:146 Resp: 141802
Ion Ratio Lower Upper
146 100
148 65.9 52.6 78.8
111 40.4 31.9 47.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#14
1,2-Dichlorobenzene
Concen: 39.377 ng
RT: 7.081 min Scan# 849
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:146 Resp: 1327751
Ion Ratio Lower Upper
146 100
148 65.3 51.6 77.4
111 43.7 34.3 51.5






#15
Benzyl Alcohol
Concen: 41.396 ng
RT: 7.051 min Scan# 843
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion: 79 Resp: 106634
Ion Ratio Lower Upper
79 100
108 83.3 64.5 96.7
77 62.2 51.1 76.7


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.135 ng
RT: 7.181 min Scan# 866
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion: 45 Resp: 1722669
Ion Ratio Lower Upper
45 100
77 13.7 0.0 33.2
79 10.3 0.0 29.9






#17
2-Methylphenol
Concen: 40.521 ng
RT: 7.151 min Scan# 861
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:107 Resp: 102096
Ion Ratio Lower Upper
107 100
108 115.2 94.2 141.2
77 44.8 36.7 55.1
79 44.3 35.8 53.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#18
Hexachloroethane
Concen: 41.105 ng
RT: 7.422 min Scan# 907
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:117 Resp: 503959
Ion Ratio Lower Upper
117 100
119 94.5 77.7 116.5
201 104.6 81.7 122.5






#19
n-Nitroso-di-n-propylamine
Concen: 39.119 ng
RT: 7.328 min Scan# 890
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion: 70 Resp: 908085
Ion Ratio Lower Upper
70 100
42 56.3 46.3 69.5
101 10.3 7.7 11.5
130 24.1 17.8 26.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#20
3+4-Methylphenols
Concen: 40.840 ng
RT: 7.310 min Scan# 888
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:107 Resp: 1280438
Ion Ratio Lower Upper
107 100
108 92.4 73.0 113.0
77 68.5 53.2 93.2
79 27.6 8.8 48.8






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1038
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:136 Resp: 1912740
Ion Ratio Lower Upper
136 100
137 11.2 9.1 13.7
54 7.5 6.2 9.2
68 6.0 5.2 7.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#22
Acetophenone
Concen: 38.293 ng
RT: 7.322 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:105 Resp: 1686221
Ion Ratio Lower Upper
105 100
71 6.1 6.3 9.5#
51 26.1 19.8 29.6
120 24.9 19.8 29.6






#23
Nitrobenzene-d5
Concen: 94.171 ng
RT: 7.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion: 82 Resp: 264744
Ion Ratio Lower Upper
82 100
128 48.7 37.7 56.5
54 50.7 40.8 61.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#24
Nitrobenzene
Concen: 44.706 ng
RT: 7.492 min Scan# 919
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion: 77 Resp: 1369130
Ion Ratio Lower Upper
77 100
123 47.9 38.2 57.2
65 13.7 10.7 16.1





#25
Isophorone
Concen: 39.398 ng
RT: 7.733 min Scan# 959
Delta R.T. 0.005 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :





ICVBF060524

Tgt Ion: 82 Resp: 2440118
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 20.4 15.5 23.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#26
2-Nitrophenol
Concen: 48.149 ng
RT: 7.804 min Scan# 972
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:139 Resp: 618831
Ion Ratio Lower Upper
139 100
109 26.8 21.8 32.6
65 48.2 40.1 60.1






#27
2,4-Dimethylphenol
Concen: 41.480 ng
RT: 7.839 min Scan# 91
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:122 Resp: 892360
Ion Ratio Lower Upper
122 100
107 107.7 84.2 126.2
121 58.2 45.4 68.0


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#28
bis(2-Chloroethoxy)methane
Concen: 39.419 ng
RT: 7.939 min Scan# 995
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion: 93 Resp: 1487130
Ion Ratio Lower Upper
93 100
95 34.3 26.6 39.8
123 13.3 10.2 15.2





#29
2,4-Dichlorophenol
Concen: 42.165 ng
RT: 8.045 min Scan# 1013
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :





ICVBF060524

Tgt Ion:162 Resp: 109327
Ion Ratio Lower Upper
162 100
164 65.1 44.1 84.1
98 33.7 17.0 57.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#30
1,2,4-Trichlorobenzene
Concen: 39.780 ng
RT: 8.128 min Scan# 1027
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:180 Resp: 1244535
Ion Ratio Lower Upper
180 100
182 95.8 76.4 114.6
145 30.5 24.8 37.2






#31
Naphthalene
Concen: 38.889 ng
RT: 8.216 min Scan# 1042
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:128 Resp: 3597830
Ion Ratio Lower Upper
128 100
129 12.2 9.4 14.0
127 14.7 11.2 16.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#32
Benzoic acid
Concen: 50.093 ng m
RT: 7.969 min Scan# 1000
Delta R.T. 0.012 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:122 Resp: 835546
Ion Ratio Lower Upper
122 100
105 121.0 102.8 142.8
77 90.1 71.2 111.2






#33
4-Chloroaniline
Concen: 39.724 ng
RT: 8.257 min Scan# 1049
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:127 Resp: 1381220
Ion Ratio Lower Upper
127 100
129 33.4 26.5 39.7
65 30.3 24.0 36.0
92 18.8 15.0 22.6

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#34
Hexachlorobutadiene
Concen: 39.797 ng
RT: 8.328 min Scan# 1061
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06


Tgt Ion:225 Resp: 731644
Ion Ratio Lower Upper
225 100
223 63.4 50.9 76.3
227 63.3 51.2 76.8





#35
Caprolactam
Concen: 40.498 ng
RT: 8.645 min Scan# 1113
Delta R.T. 0.012 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524



Tgt Ion:113 Resp: 350629
Ion Ratio Lower Upper
113 100
55 156.7 144.5 184.5
56 116.3 102.0 142.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#36
4-Chloro-3-methylphenol
Concen: 41.349 ng
RT: 8.733 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:107 Resp: 1098676
Ion Ratio Lower Upper
107 100
144 27.4 21.7 32.5
142 83.2 67.0 100.4





#37
2-Methylnaphthalene
Concen: 39.005 ng
RT: 8.904 min Scan# 1159
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :





ICVBF060524

Tgt Ion:142 Resp: 243262
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#38
1-Methylnaphthalene
Concen: 38.890 ng
RT: 9.004 min Scan# 1176
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:142 Resp: 2374818
Ion Ratio Lower Upper
142 100
141 91.8 73.3 109.9





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1186

Delta R.T. -0.000 min

Lab File: BF138125.D


Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524



Tgt Ion:164 Resp: 1086250

Ion Ratio Lower Upper

164 100

162 98.4 80.9 121.3




160 45.5 37.0 55.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024

Supervised By :mohammad ahmed 06/07/2024



#40

1,2,4,5-Tetrachlorobenzene

Concen: 39.962 ng

RT: 9.063 min Scan# 1186

Delta R.T. -0.000 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06

Tgt Ion:216 Resp: 1225557




Ion Ratio Lower Upper


216 100

214 79.4 63.3 94.9

179 21.2 16.9 25.3

108 17.6 14.2 21.2





#41
Hexachlorocyclopentadiene
Concen: 41.036 ng
RT: 9.051 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :





ICVBF060524

Tgt Ion:237 Resp: 415449
Ion Ratio Lower Upper
237 100
235 63.0 43.4 83.4
272 14.5 0.0 33.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#42
2,4,6-Tribromophenol
Concen: 86.893 ng
RT: 10.733 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:330 Resp: 908362
Ion Ratio Lower Upper
330 100
332 97.0 75.8 113.8
141 28.3 23.2 34.8





#43

2,4,6-Trichlorophenol

Concen: 44.061 ng

RT: 9.175 min Scan# 1205

Delta R.T. -0.000 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524





Tgt Ion:196 Resp: 82042

Ion	Ratio	Lower	Upper
196	100		
198	99.2	78.5	117.7
200	32.2	25.4	38.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#44

2,4,5-Trichlorophenol

Concen: 42.692 ng

RT: 9.210 min Scan# 1211




Delta R.T. -0.000 min


Lab File: BF138125.D

Acq: 05 Jun 2024 17:06

Tgt Ion:196 Resp: 875945

Ion	Ratio	Lower	Upper
196	100		
198	98.3	77.4	116.0
97	48.9	39.5	59.3
132	28.2	22.2	33.2






#45
2-Fluorobiphenyl
Concen: 74.736 ng
RT: 9.269 min Scan# 1221
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:172 Resp: 4642880
Ion Ratio Lower Upper
172 100
171 39.9 31.6 47.4
170 27.3 21.4 32.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#46
1,1'-Biphenyl
Concen: 38.427 ng
RT: 9.369 min Scan# 1238
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:154 Resp: 2932004
Ion Ratio Lower Upper
154 100
153 43.1 22.7 62.7
76 13.5 0.0 33.3






#47
2-Chloronaphthalene
Concen: 39.082 ng
RT: 9.392 min Scan# 1242
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:162 Resp: 237649
Ion Ratio Lower Upper
162 100
127 41.6 33.0 49.4
164 34.7 27.5 41.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#48
2-Nitroaniline
Concen: 43.317 ng
RT: 9.486 min Scan# 1258
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion: 65 Resp: 646658
Ion Ratio Lower Upper
65 100
92 71.5 57.3 85.9
138 120.0 95.4 143.2





#49

Acenaphthylene

Concen: 38.634 ng

RT: 9.810 min Scan# 11

Delta R.T. -0.000 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524

Tgt Ion:152 Resp: 332446

Ion Ratio Lower Upper

152 100

151 21.9 17.1 25.7





153 14.7 11.6 17.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024

Supervised By :mohammad ahmed 06/07/2024



#50

Dimethylphthalate

Concen: 38.958 ng

RT: 9.669 min Scan# 1289

Delta R.T. -0.000 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06




Tgt Ion:163 Resp: 2659047


Ion Ratio Lower Upper

163 100

194 4.4 3.5 5.3

164 11.1 8.6 12.8






#51
2,6-Dinitrotoluene
Concen: 43.167 ng
RT: 9.727 min Scan# 1299
Delta R.T. -0.001 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:165 Resp: 633990
Ion Ratio Lower Upper
165 100
63 54.7 42.9 64.3
89 54.8 45.0 67.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#52
Acenaphthene
Concen: 39.197 ng
RT: 9.980 min Scan# 1342
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:154 Resp: 2279003
Ion Ratio Lower Upper
154 100
153 110.2 87.6 131.4
152 54.2 43.0 64.6






#53
3-Nitroaniline
Concen: 42.929 ng
RT: 9.898 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:138 Resp: 631618
Ion Ratio Lower Upper
138 100
108 10.5 10.3 15.5
92 109.8 89.5 134.3


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#54
2,4-Dinitrophenol
Concen: 48.906 ng
RT: 9.998 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Tgt Ion:184 Resp: 205331
Ion Ratio Lower Upper
184 100
63 55.9 45.4 68.0
154 56.5 46.0 69.0





#55
Dibenzenofuran
Concen: 38.511 ng
RT: 10.151 min Scan# 1371
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524




Tgt Ion:168 Resp: 3167004
Ion Ratio Lower Upper
168 100
139 39.8 31.1 46.7
169 14.8 11.6 17.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#56
4-Nitrophenol
Concen: 46.216 ng
RT: 10.045 min Scan# 1353
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:139 Resp: 479785
Ion Ratio Lower Upper
139 100
109 65.8 45.5 85.5
65 89.5 69.1 109.1






#57
2,4-Dinitrotoluene
Concen: 44.446 ng
RT: 10.127 min Scan# 1367
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:165 Resp: 779209
Ion Ratio Lower Upper
165 100
63 43.5 36.8 55.2
89 71.3 60.2 90.2
182 2.9 2.2 3.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#58
Fluorene
Concen: 38.184 ng
RT: 10.492 min Scan# 1429
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:166 Resp: 2494209
Ion Ratio Lower Upper
166 100
165 96.9 76.3 114.5
167 14.2 11.2 16.8






#59
2,3,4,6-Tetrachlorophenol
Concen: 44.274 ng
RT: 10.263 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:232 Resp: 716501
Ion Ratio Lower Upper
232 100
131 41.6 33.7 50.5
130 2.1 1.7 2.5
166 29.1 23.0 34.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#60
Diethylphthalate
Concen: 38.638 ng
RT: 10.369 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Tgt Ion:149 Resp: 2574456
Ion Ratio Lower Upper
149 100
177 25.3 19.9 29.9
150 13.4 10.7 16.1





#61
4-Chlorophenyl-phenylether
Concen: 38.006 ng
RT: 10.486 min Scan# 1428
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524



Tgt Ion:204 Resp: 126121
Ion Ratio Lower Upper
204 100
206 33.5 26.5 39.7
141 56.9 45.2 67.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#62
4-Nitroaniline
Concen: 46.527 ng
RT: 10.516 min Scan# 1433
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06


Tgt Ion:138 Resp: 592759
Ion Ratio Lower Upper
138 100
92 49.1 32.3 72.3
108 69.6 53.0 93.0





#63
Azobenzene
Concen: 38.697 ng
RT: 10.645 min Scan# 1455
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524




Tgt Ion: 77 Resp: 2483468
Ion Ratio Lower Upper
77 100
182 28.9 7.2 47.2
105 22.8 3.0 43.0
51 30.0 10.3 50.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Tgt Ion:188 Resp: 1856310
Ion Ratio Lower Upper
188 100
94 10.4 7.4 11.2
80 11.4 7.7 11.5





#65
4,6-Dinitro-2-methylphenol
Concen: 46.895 ng
RT: 10.539 min Scan# 1436
Delta R.T. 0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524



Tgt Ion:198 Resp: 318459
Ion Ratio Lower Upper
198 100
51 30.1 28.2 68.2
105 35.5 26.8 66.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#66
n-Nitrosodiphenylamine
Concen: 40.184 ng
RT: 10.604 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:169 Resp: 2227546
Ion Ratio Lower Upper
169 100
168 67.0 53.4 80.2
167 37.4 29.8 44.6






#67
4-Bromophenyl-phenylether
Concen: 41.067 ng
RT: 10.974 min Scan# 1511
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:248 Resp: 868000
Ion Ratio Lower Upper
248 100
250 97.3 78.7 118.1
141 69.0 56.5 84.7


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#68
Hexachlorobenzene
Concen: 40.099 ng
RT: 11.039 min Scan# 1522
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:284 Resp: 985324
Ion Ratio Lower Upper
284 100
142 27.1 21.4 32.0
249 29.7 23.3 34.9






#69
 Atrazine
 Concen: 39.208 ng
 RT: 11.133 min Scan# 1538
 Delta R.T. -0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:200 Resp: 699270
 Ion Ratio Lower Upper
 200 100
 173 25.3 5.7 45.7
 215 49.1 29.4 69.4


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024




#70
 Pentachlorophenol
 Concen: 47.929 ng
 RT: 11.227 min Scan# 1554
 Delta R.T. -0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06



Tgt Ion:266 Resp: 630451
 Ion Ratio Lower Upper
 266 100
 268 63.3 51.4 77.0
 264 63.8 50.1 75.1






#71
Phenanthrene
Concen: 38.402 ng
RT: 11.457 min Scan# 1593
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:178 Resp: 3431774
Ion Ratio Lower Upper
178 100
176 21.8 17.2 25.8
179 17.2 13.6 20.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#72
Anthracene
Concen: 38.720 ng
RT: 11.510 min Scan# 1602
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:178 Resp: 3451109
Ion Ratio Lower Upper
178 100
176 20.9 16.5 24.7
179 17.2 13.6 20.4






#73
 Carbazole
 Concen: 38.578 ng
 RT: 11.663 min Scan# 1628
 Delta R.T. -0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:167 Resp: 3144510
 Ion Ratio Lower Upper
 167 100
 166 23.1 18.6 27.8
 139 13.4 11.0 16.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024




#74
 Di-n-butylphthalate
 Concen: 38.552 ng
 RT: 11.992 min Scan# 1684
 Delta R.T. -0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06



Tgt Ion:149 Resp: 3693647
 Ion Ratio Lower Upper
 149 100
 150 10.8 8.5 12.7
 104 5.5 4.4 6.6






#75
Fluoranthene
Concen: 37.445 ng
RT: 12.645 min Scan# 1795
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:202 Resp: 3610500
Ion Ratio Lower Upper
202 100
101 11.5 0.0 32.2
203 19.0 0.0 39.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.068 min Scan# 2037
Delta R.T. 0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06


Tgt Ion:240 Resp: 956610
Ion Ratio Lower Upper
240 100
120 10.8 9.8 14.8
236 26.9 21.2 31.8





#77
Benzidine
Concen: 40.289 ng
RT: 12.757 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06




Instrument :
BNA_F
ClientSampleId :
ICVBF060524




Tgt Ion:184 Resp: 396608
Ion Ratio Lower Upper
184 100
185 16.6 12.2 18.2
183 11.0 9.0 13.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024




#78
Pyrene
Concen: 42.015 ng
RT: 12.874 min Scan# 1834
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06



Tgt Ion:202 Resp: 3634887
Ion Ratio Lower Upper
202 100
200 23.4 19.1 28.7
203 19.0 15.6 23.4






#79
 Terphenyl-d14
 Concen: 80.035 ng
 RT: 13.015 min Scan# 1858
 Delta R.T. -0.006 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:244 Resp: 5118700
 Ion Ratio Lower Upper
 244 100
 212 8.2 6.4 9.6
 122 12.7 9.4 14.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
 Supervised By :mohammad ahmed 06/07/2024



#80
 Butylbenzylphthalate
 Concen: 47.145 ng
 RT: 13.492 min Scan# 1939
 Delta R.T. -0.000 min
 Lab File: BF138125.D
 Acq: 05 Jun 2024 17:06

Tgt Ion:149 Resp: 1346834
 Ion Ratio Lower Upper
 149 100
 91 70.6 59.6 89.4
 206 23.8 17.8 26.6





#81
Benzo(a)anthracene
Concen: 39.065 ng
RT: 14.057 min Scan# 2035
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F





ClientSampleId :

ICVBF060524

Tgt Ion:228 Resp: 235771
Ion Ratio Lower Upper
228 100
226 28.6 22.7 34.1
229 20.9 16.7 25.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#82
3,3'-Dichlorobenzidine
Concen: 40.745 ng
RT: 14.021 min Scan# 2029
Delta R.T. -0.000 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:252 Resp: 623278
Ion Ratio Lower Upper
252 100
254 64.2 52.9 79.3
126 9.2 8.8 13.2





#83

Chrysene

Concen: 39.746 ng

RT: 14.092 min Scan# 2042

Delta R.T. -0.006 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524

Tgt Ion:228 Resp: 230216

Ion Ratio Lower Upper

228 100





226 31.6 25.6 38.4

229 20.7 16.6 25.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 44.325 ng

RT: 14.051 min Scan# 2034

Delta R.T. -0.006 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06




Tgt Ion:149 Resp: 1844863


Ion Ratio Lower Upper

149 100

167 29.2 24.2 36.2

279 5.0 4.6 6.8





#85

Di-n-octyl phthalate

Concen: 45.161 ng

RT: 14.668 min Scan# 2139

Delta R.T. -0.000 min

Lab File: BF138125.D


Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524



Tgt Ion:149 Resp: 2587328

Ion Ratio Lower Upper

149 100




167 1.7 1.4 2.0

43 10.1 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.545 min Scan# 2288

Delta R.T. -0.000 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06




Tgt Ion:264 Resp: 733906


Ion Ratio Lower Upper

264 100

260 24.6 19.4 29.2

265 21.8 17.1 25.7






#87
Indeno(1,2,3-cd)pyrene
Concen: 41.460 ng
RT: 17.033 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:276 Resp: 1636174
Ion Ratio Lower Upper
276 100
138 22.7 18.9 28.3
277 24.9 20.3 30.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#88
Benzo(b)fluoranthene
Concen: 40.789 ng
RT: 15.109 min Scan# 2214
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:252 Resp: 1881891
Ion Ratio Lower Upper
252 100
253 22.1 18.2 27.2
125 10.3 7.4 11.2





#89

Benzo(k)fluoranthene

Concen: 38.655 ng

RT: 15.145 min Scan# 2220

Delta R.T. -0.000 min

Lab File: BF138125.D


Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524



Tgt Ion:252 Resp: 170073

Ion Ratio Lower Upper

252 100

253 22.4 17.8 26.6




125 8.8 7.8 11.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024

Supervised By :mohammad ahmed 06/07/2024



#90

Benzo(a)pyrene

Concen: 41.206 ng

RT: 15.480 min Scan# 2277

Delta R.T. -0.006 min

Lab File: BF138125.D

Acq: 05 Jun 2024 17:06




Tgt Ion:252 Resp: 1439803


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 11.2 8.0 12.0






#91
Dibenzo(a,h)anthracene
Concen: 42.392 ng
RT: 17.056 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Instrument :

BNA_F

ClientSampleId :

ICVBF060524






Tgt Ion:278 Resp: 134013
Ion Ratio Lower Upper
278 100
139 14.7 11.7 17.5
279 24.1 19.2 28.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/06/2024
Supervised By :mohammad ahmed 06/07/2024



#92
Benzo(g,h,i)perylene
Concen: 40.288 ng
RT: 17.486 min Scan# 2618
Delta R.T. -0.006 min
Lab File: BF138125.D
Acq: 05 Jun 2024 17:06

Tgt Ion:276 Resp: 1320045
Ion Ratio Lower Upper
276 100
277 23.5 19.0 28.4
138 19.9 16.5 24.7

