



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060525\
Data File : BF142624.D
Acq On : 05 Jun 2025 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue May 20 16:26:47 2025






AutoFind: Scans 1611, 1612, 1613; Background Corrected with Scan 1605

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	44267	PASS
68	69	0.00	2	1.8	691	PASS
69	198	0.00	100	33.2	38579	PASS
70	69	0.00	2	0.6	236	PASS
127	198	10	80	45.4	52819	PASS
197	198	0.00	2	0.5	550	PASS
198	198	100	100	100.0	116347	PASS
199	198	5	9	6.8	7924	PASS
275	198	10	60	27.9	32424	PASS
365	198	1	100	3.5	4075	PASS
441	198	0.01	100	17.5	20310	PASS
442	442	50	100	100.0	134237	PASS
443	442	15	24	19.6	26320	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060525\
Data File : BF142624.D
Acq On : 05 Jun 2025 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jun 05 12:57:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration



TIC: BF142624.D\data.ms

(70) Pentachlorophenol (C)

11.227min (-0.006) 0.00 ng




response 138743

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.60	64.43
264.00	63.00	63.68
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060525\
Data File : BF142624.D
Acq On : 05 Jun 2025 09:07
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jun 05 12:57:43 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration



TIC: BF142624.D\data.ms

(77) Benzidine

12.763min (+ 0.000) 0.00 ng

response 543002

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.50	13.80
183.00	11.20	11.04
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/3/2025	BNA_F	<u>BF142624.D</u>
Compound Name	Response	Retention Time
DDT	279286	13.586
DDD	6433	13.286
DDE	730	12.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7163	286449	2.50

Instrument :
BNA_F
ClientSampleId :
DFTPP