


Data Path : Z:\HPCHEM1\BNA_F\Data\BF060615\
 Data File : BF079768.D
 Acq On : 6 Jun 2015 14:20
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 08 18:33:51 2015



AutoFind: Scans 971, 972, 973; Background Corrected with Scan 968

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.2	13568	PASS
68	69	0.00	2	1.0	125	PASS
69	198	0.00	100	34.4	12872	PASS
70	69	0.00	2	1.6	211	PASS
127	198	10	80	42.9	16055	PASS
197	198	0.00	2	0.3	128	PASS
198	198	100	100	100.0	37440	PASS
199	198	5	9	6.6	2457	PASS
275	198	10	60	24.2	9068	PASS
365	198	1	100	2.7	999	PASS
441	442	0.01	24	15.3	4770	PASS
442	198	50	100	83.1	31094	PASS
443	442	15	24	19.0	5921	PASS