


Data Path : Z:\HPCHEM1\BNA_F\Data\BF060617\
 Data File : BF095734.D
 Acq On : 7 Jun 2017 5:02
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 05 16:15:31 2017



AutoFind: Scans 2315, 2316, 2317; Background Corrected with Scan 2310

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.9	8697	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	36.5	8609	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.4	11874	PASS
197	198	0.00	2	1.1	271	PASS
198	198	100	100	100.0	23572	PASS
199	198	5	9	7.6	1795	PASS
275	198	10	60	23.4	5523	PASS
365	198	1	100	2.3	531	PASS
441	442	0.01	24	16.5	3174	PASS
442	198	50	100	81.8	19291	PASS
443	442	15	24	20.7	3989	PASS