

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060624\
 Data File : BF138128.D
 Acq On : 06 Jun 2024 11:28
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 11:53:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	38.778	3.1	102	0.00
3	Pyridine	40.000	38.956	2.6	101	0.00
4	n-Nitrosodimethylamine	40.000	38.583	3.5	99	0.01
5 S	2-Fluorophenol	80.000	78.706	1.6	102	0.00
6	Aniline	40.000	38.585	3.5	100	0.00
7 S	Phenol-d6	80.000	79.129	1.1	103	0.00
8	2-Chlorophenol	40.000	40.162	-0.4	103	0.00
9	Benzaldehyde	40.000	47.137	-17.8	110	0.00
10 C	Phenol	40.000	39.373	1.6	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.547	3.6	100	0.00
12	1,3-Dichlorobenzene	40.000	39.669	0.8	102	0.00
13 C	1,4-Dichlorobenzene	40.000	39.653	0.9	103	0.00
14	1,2-Dichlorobenzene	40.000	39.764	0.6	102	0.00
15	Benzyl Alcohol	40.000	40.329	-0.8	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.850	5.4	98	0.00
17	2-Methylphenol	40.000	39.427	1.4	103	0.00
18	Hexachloroethane	40.000	40.714	-1.8	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.397	4.0	101	0.00
20	3+4-Methylphenols	40.000	40.568	-1.4	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	39.059	2.4	102	0.00
23 S	Nitrobenzene-d5	80.000	91.180	-14.0	110	0.00
24	Nitrobenzene	40.000	43.406	-8.5	107	0.00
25	Isophorone	40.000	38.417	4.0	101	0.00
26 C	2-Nitrophenol	40.000	44.233	-10.6	129	0.00
27	2,4-Dimethylphenol	40.000	40.852	-2.1	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.203	2.0	102	0.00
29 C	2,4-Dichlorophenol	40.000	42.084	-5.2	106	0.00
30	1,2,4-Trichlorobenzene	40.000	39.688	0.8	103	0.00
31	Naphthalene	40.000	39.588	1.0	103	0.00
32	Benzoic acid	40.000	45.707	-14.3	144	0.00
33	4-Chloroaniline	40.000	39.208	2.0	103	0.00
34 C	Hexachlorobutadiene	40.000	39.700	0.7	102	0.00
35	Caprolactam	40.000	40.772	-1.9	105	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.793	-2.0	104	0.00
37	2-Methylnaphthalene	40.000	39.481	1.3	103	0.00
38	1-Methylnaphthalene	40.000	39.485	1.3	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.889	0.3	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.997	-5.0	103	0.00
42 S	2,4,6-Tribromophenol	80.000	89.796	-12.2	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.515	-8.8	109	0.00
44	2,4,5-Trichlorophenol	40.000	42.025	-5.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	76.889	3.9	102	0.00
46	1,1'-Biphenyl	40.000	38.919	2.7	103	0.00
47	2-Chloronaphthalene	40.000	39.093	2.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060624\
 Data File : BF138128.D
 Acq On : 06 Jun 2024 11:28
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 11:53:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 06 05:19:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.122	-5.3	117	0.00
49	Acenaphthylene	40.000	39.316	1.7	104	0.00
50	Dimethylphthalate	40.000	39.144	2.1	103	0.00
51	2,6-Dinitrotoluene	40.000	42.202	-5.5	114	0.00
52 C	Acenaphthene	40.000	39.629	0.9	105	0.00
53	3-Nitroaniline	40.000	41.822	-4.6	113	0.00
54 P	2,4-Dinitrophenol	40.000	44.411	-11.0	145	0.00
55	Dibenzofuran	40.000	39.042	2.4	103	0.00
56 P	4-Nitrophenol	40.000	45.406	-13.5	113	0.00
57	2,4-Dinitrotoluene	40.000	43.754	-9.4	117	0.00
58	Fluorene	40.000	39.035	2.4	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.952	-9.9	107	0.00
60	Diethylphthalate	40.000	39.651	0.9	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.861	2.8	102	0.00
62	4-Nitroaniline	40.000	47.878	-19.7	111	0.00
63	Azobenzene	40.000	39.113	2.2	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.502	-8.8	131	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.847	0.4	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.711	-1.8	105	0.00
68	Hexachlorobenzene	40.000	40.091	-0.2	105	0.00
69	Atrazine	40.000	39.478	1.3	103	0.00
70 C	Pentachlorophenol	40.000	47.035	-17.6	111	0.00
71	Phenanthrene	40.000	38.925	2.7	101	0.00
72	Anthracene	40.000	38.838	2.9	102	0.00
73	Carbazole	40.000	39.041	2.4	102	0.00
74	Di-n-butylphthalate	40.000	39.061	2.3	101	0.00
75 C	Fluoranthene	40.000	38.866	2.8	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	50.642	-26.6#	94	0.00
78	Pyrene	40.000	41.171	-2.9	102	0.00
79 S	Terphenyl-d14	80.000	78.943	1.3	102	0.00
80	Butylbenzylphthalate	40.000	44.780	-12.0	102	0.00
81	Benzo(a)anthracene	40.000	39.138	2.2	99	0.00
82	3,3'-Dichlorobenzidine	40.000	41.350	-3.4	103	0.00
83	Chrysene	40.000	40.158	-0.4	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.481	-6.2	101	0.00
85 c	Di-n-octyl phthalate	40.000	42.034	-5.1	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.543	-1.4	105	0.00
88	Benzo(b)fluoranthene	40.000	39.092	2.3	105	0.00
89	Benzo(k)fluoranthene	40.000	38.291	4.3	100	0.00
90 C	Benzo(a)pyrene	40.000	40.242	-0.6	103	0.00
91	Dibenzo(a,h)anthracene	40.000	40.335	-0.8	105	0.00
92	Benzo(g,h,i)perylene	40.000	40.347	-0.9	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060624\
Data File : BF138128.D
Acq On : 06 Jun 2024 11:28
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 06 11:53:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF060524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 06 05:19:29 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0