

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060717\
 Data File : BF095751.D
 Acq On : 7 Jun 2017 17:55
 Operator : SJ/MA
 Sample : I3479-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-9(0-2)20170603

Manual Integrations
 APPROVED

mohammad
 6/8/2017 12:18:49 PM

Quant Time: Jun 08 04:56:12 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 05 16:15:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	79432	20.00	ng	-0.02
21) Naphthalene-d8	9.42	136	301691	20.00	ng	-0.02
38) Acenaphthene-d10	12.24	164	127478	20.00	ng	-0.02
63) Phenanthrene-d10	14.63	188	189191	20.00	ng	-0.02
75) Chrysene-d12	18.34	240	145782	20.00	ng	-0.01
86) Perylene-d12	20.01	264	112261	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.27	112	467160	93.72	ng	-0.02
7) Phenol-d6	6.95	99	568779	95.31	ng	-0.01
23) Nitrobenzene-d5	8.30	82	331672	68.28	ng	-0.02
41) 2,4,6-Tribromophenol	13.53	330	93730	83.10	ng	-0.02
44) 2-Fluorobiphenyl	11.20	172	628206	68.58	ng	-0.02
78) Terphenyl-d14	17.00	244	366893	62.46	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	11.89	163	107346	11.09	ng	97
74) Fluoranthene	16.45	202	40599	3.76	ng	98
77) Pyrene	16.75	202	35474	3.26	ng	# 95
80) Benzo(a)anthracene	18.33	228	18653	2.20	ng	# 94
82) Chrysene	18.38	228	17778	2.10	ng	# 93
87) Benzo(b)fluoranthene	19.61	252	18782m	2.67	ng	
89) Benzo(a)pyrene	19.97	252	13099	2.03	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF060717\
Data File : BF095751.D
Acq On : 7 Jun 2017 17:55
Operator : SJ/MA
Sample : I3479-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-9(0-2)20170603

Manual Integrations
APPROVED

mohammad
6/8/2017 12:18:49 PM

Quant Time: Jun 08 04:56:12 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration

