

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060721\
 Data File : BF124155.D
 Acq On : 7 Jun 2021 14:01
 Operator : JU/CG
 Sample : M2583-22
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 54-57

Quant Time: Jun 07 14:22:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 13:32:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	37298	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	151816	20.000	ng	# 0.00
39) Acenaphthene-d10	9.910	164	85115	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	128536	20.000	ng	0.00
76) Chrysene-d12	14.039	240	85759	20.000	ng	0.00
86) Perylene-d12	15.527	264	86484	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	202819	81.857	ng	0.01
7) Phenol-d6	6.522	99	171043	51.358	ng	0.02
23) Nitrobenzene-d5	7.434	82	324353	84.736	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	135503	112.369	ng	0.01
45) 2-Fluorobiphenyl	9.234	172	592313	87.569	ng	0.01
79) Terphenyl-d14	12.980	244	362774	58.078	ng	0.00
Target Compounds						
						Qvalue
4) n-Nitrosodimethylamine	3.316	42	6181	3.746	ng	# 1
10) Phenol	6.534	94	171401	47.622	ng	88
15) Benzyl Alcohol	7.034	79	7871	2.622	ng	# 1
17) 2-Methylphenol	7.292	107	380583	169.813	ng	# 68
19) n-Nitroso-di-n-propyla...	7.269	70	4797	2.040	ng	# 61
20) 3+4-Methylphenols	7.292	107	380583	128.159	ng	# 62
22) Acetophenone	7.275	105	12228	2.825	ng	# 1
25) Isophorone	7.704	82	14450	2.095	ng	# 84
27) 2,4-Dimethylphenol	7.892	122	24053	9.944	ng	93
28) bis(2-Chloroethoxy)met...	7.887	93	24471	7.156	ng	# 89
32) Benzoic acid	7.892	122	24053	16.206	ng	# 19
35) Caprolactam	8.622	113	3664	4.765	ng	# 1
36) 4-Chloro-3-methylphenol	8.745	107	193022	63.034	ng	# 85
37) 2-Methylnaphthalene	8.745	142	128214	20.143	ng	# 18
38) 1-Methylnaphthalene	8.745	142	128214	21.555	ng	# 17
50) Dimethylphthalate	9.639	163	58451	8.300	ng	# 98
54) 2,4-Dinitrophenol	9.945	184	110	3.639	ng	# 1
60) Diethylphthalate	10.339	149	21413	3.013	ng	# 96
67) 4-Bromophenyl-phenylether	10.710	248	8589	4.868	ng	# 1
70) Pentachlorophenol	11.210	266	7564	11.429	ng	91
77) Benzidine	12.727	184	1275	4.645	ng	# 1
80) Butylbenzylphthalate	13.457	149	9800	2.934	ng	# 85
84) Bis(2-ethylhexyl)phtha...	14.016	149	23389	5.886	ng	94
87) Indeno(1,2,3-cd)pyrene	17.009	276	640	4.416	ng	# 1
91) Dibenzo(a,h)anthracene	17.068	278	221	4.208	ng	# 1
92) Benzo(g,h,i)perylene	17.445	276	114	3.824	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060721\
Data File : BF124155.D
Acq On : 7 Jun 2021 14:01
Operator : JU/CG
Sample : M2583-22
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
54-57

Quant Time: Jun 07 14:22:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 13:32:21 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.863 min Scan# 812
 Delta R.T. 0.000 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57


Tgt Ion:	152	Resp:	37298
Ion	Ratio	Lower	Upper
152	100		
150	155.6	127.0	190.6
115	67.4	54.3	81.5



n-Nitrosodimethylamine
 Concen: 3.746 ng
 RT: 3.316 min Scan# 209
 Delta R.T. -0.023 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	42	Resp:	6181
Ion	Ratio	Lower	Upper
42	100		
74	232.2	83.5	125.3#
44	31.8	5.7	8.5#









2-Fluorophenol
 Concen: 81.857 ng
 RT: 5.498 min Scan# 580
 Delta R.T. 0.012 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57

Tgt Ion:	112	Resp:	202819
Ion	Ratio	Lower	Upper
112	100		
64	70.7	61.4	92.0
63	41.4	33.0	49.4




Phenol-d6
 Concen: 51.358 ng
 RT: 6.522 min Scan# 754
 Delta R.T. 0.018 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	99	Resp:	171043
Ion	Ratio	Lower	Upper
99	100		
42	22.9	19.8	29.8
71	47.5	31.1	46.7#




Abundance Scan 754 (6.522 min): BF124122.D\data.ms (-747) #10





Phenol
Concen: 47.622 ng
RT: 6.534 min Scan# 756
Delta R.T. 0.018 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01


Instrument :
BNA_F
ClientSampleId :
54-57



Tgt Ion:	94	Resp:	171401
Ion	Ratio	Lower	Upper
94	100		
65	36.6	18.1	58.1
66	61.8	29.0	69.0






Abundance Scan 837 (7.010 min): BF124122.D\data.ms (-831) #15




Benzyl Alcohol
Concen: 2.622 ng
RT: 7.034 min Scan# 841
Delta R.T. 0.024 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:	79	Resp:	7871
Ion	Ratio	Lower	Upper
79	100		
108	262.0	55.8	83.8#
77	71.9	49.9	74.9




Abundance Scan 857 (7.128 min): BF124122.D\data.ms (-850) #17





2-Methylphenol
Concen: 169.813 ng
RT: 7.292 min Scan# 885
Delta R.T. 0.171 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01


Instrument :
BNA_F
ClientSampled :
54-57




Tgt Ion	Ratio	Lower	Upper
107	100		
108	76.4	89.8	134.8#
77	34.8	40.9	61.3#
79	25.9	40.2	60.4#





Abundance Scan 883 (7.281 min): BF124122.D\data.ms (-877) #19




n-Nitroso-di-n-propylamine
Concen: 2.040 ng
RT: 7.269 min Scan# 881
Delta R.T. -0.012 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01




Tgt Ion	Ratio	Lower	Upper
70	100		
42	98.5	60.1	90.1#
101	58.6	7.4	11.2#
130	0.0	14.2	21.2#








3+4-Methylphenols
 Concen: 128.159 ng
 RT: 7.292 min Scan# 885
 Delta R.T. 0.018 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01


Instrument :
 BNA_F
 ClientSampleId :
 54-57





Tgt Ion:	107	Resp:	380583
Ion	Ratio	Lower	Upper
107	100		
108	76.4	71.5	111.5
77	34.8	82.0	122.0#
79	25.9	14.1	54.1




Naphthalene-d8
 Concen: 20.000 ng
 RT: 8.145 min Scan# 1030
 Delta R.T. 0.000 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01



Tgt Ion:	136	Resp:	151816
Ion	Ratio	Lower	Upper
136	100		
137	10.2	8.7	13.1
54	17.5	9.8	14.6#
68	10.1	4.1	6.1#






Abundance Scan 883 (7.281 min): BF124122.D\data.ms (-876) #22




Acetophenone
Concen: 2.825 ng
RT: 7.275 min Scan# 882
Delta R.T. 0.000 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57

Tgt Ion:	105	Resp:	12228
Ion	Ratio	Lower	Upper
105	100		
71	104.5	4.2	6.4#
51	119.6	33.6	50.4#
120	16.2	16.3	24.5#







Abundance Scan 909 (7.434 min): BF124122.D\data.ms (-903) #23



Nitrobenzene-d5
Concen: 84.736 ng
RT: 7.434 min Scan# 909
Delta R.T. 0.006 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:	82	Resp:	324353
Ion	Ratio	Lower	Upper
82	100		
128	40.5	29.4	44.2
54	58.9	54.1	81.1









Isophorone
 Concen: 2.095 ng
 RT: 7.704 min Scan# 955
 Delta R.T. 0.018 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57


Tgt Ion:	82	Resp:	14450
Ion	Ratio	Lower	Upper
82	100		
95	8.3	5.8	8.6
138	5.7	11.8	17.8#



2,4-Dimethylphenol
 Concen: 9.944 ng
 RT: 7.892 min Scan# 987
 Delta R.T. 0.088 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	122	Resp:	24053
Ion	Ratio	Lower	Upper
122	100		
107	124.9	106.3	159.5
121	61.3	45.4	68.0









#28
 bis(2-Chloroethoxy)methane
 Concen: 7.156 ng
 RT: 7.887 min Scan# 986
 Delta R.T. -0.006 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57

Tgt Ion:	93	Resp:	24471
Ion	Ratio	Lower	Upper
93	100		
95	39.2	26.5	39.7
123	15.4	8.4	12.6#




#32
 Benzoic acid
 Concen: 16.206 ng
 RT: 7.892 min Scan# 987
 Delta R.T. -0.041 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	122	Resp:	24053
Ion	Ratio	Lower	Upper
122	100		
105	5.7	107.8	147.8#
77	51.9	81.6	121.6#






Abundance Scan 1107 (8.598 min): BF124122.D\data.ms (-11035)




Caprolactam
Concen: 4.765 ng
RT: 8.622 min Scan# 1111
Delta R.T. 0.030 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57

Tgt Ion:113 Resp: 3664
Ion Ratio Lower Upper
113 100
55 11714.1 254.9 294.9#
56 5449.5 179.4 219.4#





Abundance Scan 1125 (8.704 min): BF124122.D\data.ms (-11206)




4-Chloro-3-methylphenol
Concen: 63.034 ng
RT: 8.745 min Scan# 1132
Delta R.T. 0.041 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:107 Resp: 193022
Ion Ratio Lower Upper
107 100
144 28.2 16.3 24.5#
142 73.3 50.1 75.1




Abundance Scan 1132 (8.745 min): BF124155.D\data.ms (-1074) (-)






Abundance Scan 1152 (8.863 min): BF124122.D\data.ms (-11437)

2-Methylnaphthalene
Concen: 20.143 ng
RT: 8.745 min Scan# 1132
Delta R.T. -0.112 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampled :
54-57




Tgt Ion:142 Resp: 128214
Ion Ratio Lower Upper
142 100
141 12.2 71.8 107.6#






Abundance Scan 1169 (8.963 min): BF124122.D\data.ms (-11638)


1-Methylnaphthalene
Concen: 21.555 ng
RT: 8.745 min Scan# 1132
Delta R.T. -0.217 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01



Tgt Ion:142 Resp: 128214
Ion Ratio Lower Upper
142 100
141 12.2 72.8 109.2#






Abundance Scan 1329 (9.904 min): BF124122.D\data.ms (-1329)




Acenaphthene-d10
Concen: 20.000 ng
RT: 9.910 min Scan# 1330
Delta R.T. 0.006 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57

Tgt Ion:	164	Resp:	85115
Ion	Ratio	Lower	Upper
164	100		
162	102.3	80.7	121.1
160	42.6	37.3	55.9





Abundance Scan 1464 (10.698 min): BF124122.D\data.ms (-1464) (-)




2,4,6-Tribromophenol
Concen: 112.369 ng
RT: 10.710 min Scan# 1466
Delta R.T. 0.012 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01


Tgt Ion:	330	Resp:	135503
Ion	Ratio	Lower	Upper
330	100		
332	96.8	77.1	115.7
141	38.5	27.9	41.9



Abundance Scan 1466 (10.710 min): BF124155.D\data.ms (-1413) (-)






Abundance Scan 1214 (9.228 min): BF124122.D\data.ms (-120545)




2-Fluorobiphenyl
Concen: 87.569 ng
RT: 9.234 min Scan# 1215
Delta R.T. 0.012 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57

Tgt Ion:	172	Resp:	592313
Ion	Ratio	Lower	Upper
172	100		
171	36.2	28.1	42.1
170	22.2	19.2	28.8






Abundance Scan 1282 (9.627 min): BF124122.D\data.ms (-127450)




Dimethylphthalate
Concen: 8.300 ng
RT: 9.639 min Scan# 1284
Delta R.T. 0.012 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:	163	Resp:	58451
Ion	Ratio	Lower	Upper
163	100		
194	2.6	3.1	4.7#
164	10.7	8.2	12.2






Abundance Scan 1340 (9.969 min): BF124122.D\data.ms (-13354)




2,4-Dinitrophenol
Concen: 3.639 ng
RT: 9.945 min Scan# 1336
Delta R.T. -0.017 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampled :
54-57

Tgt Ion:	184	Resp:	110
Ion	Ratio	Lower	Upper
184	100		
63	410.6	73.5	110.3#
154	201.3	52.5	78.7#






Abundance Scan 1401 (10.327 min): BF124122.D\data.ms (-13569)




Diethylphthalate
Concen: 3.013 ng
RT: 10.339 min Scan# 1403
Delta R.T. 0.018 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:	149	Resp:	21413
Ion	Ratio	Lower	Upper
149	100		
177	18.6	15.4	23.2
150	9.6	9.9	14.9#






Abundance Scan 1582 (11.392 min): BF124122.D\data.ms (-1576) (-)




Phenanthrene-d10
Concen: 20.000 ng
RT: 11.398 min Scan# 1583
Delta R.T. 0.006 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57

Tgt Ion:	188	Resp:	128536
Ion	Ratio	Lower	Upper
188	100		
94	8.8	6.9	10.3
80	10.6	8.2	12.2






Abundance Scan 1504 (10.933 min): BF124122.D\data.ms (-1467) (-)



4-Bromophenyl-phenylether
Concen: 4.868 ng
RT: 10.710 min Scan# 1466
Delta R.T. -0.223 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01


Tgt Ion:	248	Resp:	8589
Ion	Ratio	Lower	Upper
248	100		
250	206.1	77.4	116.2#
141	572.9	63.7	95.5#






Abundance Scan 1549 (11.198 min): BF124122.D\data.ms (-1498) (-)

Pentachlorophenol
Concen: 11.429 ng
RT: 11.210 min Scan# 1551
Delta R.T. 0.012 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Instrument :
BNA_F
ClientSampleId :
54-57






Tgt Ion:	266	Resp:	7564
Ion Ratio	Lower	Upper	
266	100		
268	72.0	51.6	77.4
264	59.7	52.6	78.8




Abundance Scan 2031 (14.033 min): BF124122.D\data.ms (-2076) (-)

Chrysene-d12
Concen: 20.000 ng
RT: 14.039 min Scan# 2032
Delta R.T. 0.006 min
Lab File: BF124155.D
Acq: 7 Jun 2021 14:01

Tgt Ion:	240	Resp:	85759
Ion Ratio	Lower	Upper	
240	100		
120	7.9	7.2	10.8
236	21.3	19.0	28.4









Scan 1808 (12.721 min): BF124122.D\data.ms (-1803) (-)
 Benzidine
 Concen: 4.645 ng
 RT: 12.727 min Scan# 1809
 Delta R.T. 0.006 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57


Tgt Ion:	184	Resp:	1275
Ion	Ratio	Lower	Upper
184	100		
185	1864.5	12.1	18.1#
183	141.1	7.8	11.8#



Scan 1852 (12.980 min): BF124122.D\data.ms (-1879) (-)
 Terphenyl-d14
 Concen: 58.078 ng
 RT: 12.980 min Scan# 1852
 Delta R.T. 0.006 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	244	Resp:	362774
Ion	Ratio	Lower	Upper
244	100		
212	7.7	5.7	8.5
122	9.1	9.7	14.5#









Scan 1932 (13.451 min): BF124122.D\data.ms (-1932) (-)
 Butylbenzylphthalate
 Concen: 2.934 ng
 RT: 13.457 min Scan# 1933
 Delta R.T. 0.006 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57


Tgt Ion:	149	Resp:	9800
Ion	Ratio	Lower	Upper
149	100		
91	94.0	63.6	95.4
206	13.3	14.5	21.7#



Scan 2027 (14.010 min): BF124122.D\data.ms (-2027) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 5.886 ng
 RT: 14.016 min Scan# 2028
 Delta R.T. 0.006 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01

Tgt Ion:	149	Resp:	23389
Ion	Ratio	Lower	Upper
149	100		
167	23.9	21.8	32.6
279	4.4	3.3	4.9









Perylene-d12
 Concen: 20.000 ng
 RT: 15.527 min Scan# 2285
 Delta R.T. 0.018 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01




Instrument :
 BNA_F
 ClientSampleId :
 54-57


Tgt Ion:	264	Resp:	86484
Ion	Ratio	Lower	Upper
264	100		
260	25.3	18.7	28.1
265	22.6	17.8	26.8



Indeno(1,2,3-cd)pyrene
 Concen: 4.416 ng
 RT: 17.009 min Scan# 2537
 Delta R.T. 0.000 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01


Tgt Ion:	276	Resp:	640
Ion	Ratio	Lower	Upper
276	100		
138	132.2	16.2	24.2#
277	28.3	20.1	30.1








Scan 2539 (17.021 min): BF124122.D\data.ms (-2529) (-)
 Dibenzo(a,h)anthracene
 Concen: 4.208 ng
 RT: 17.068 min Scan# 2547
 Delta R.T. 0.047 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01


Instrument :
 BNA_F
 ClientSampleId :
 54-57



Tgt Ion	Ratio	Lower	Upper
278	100		
139	169.2	10.2	15.2#
279	0.0	19.4	29.0#



Scan 2613 (17.456 min): BF124122.D\data.ms (-2602) (-)
 Benzo(g,h,i)perylene
 Concen: 3.824 ng
 RT: 17.445 min Scan# 2611
 Delta R.T. -0.017 min
 Lab File: BF124155.D
 Acq: 7 Jun 2021 14:01



Tgt Ion	Ratio	Lower	Upper
276	100		
277	0.0	18.2	27.2#
138	177.7	13.8	20.6#

