


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF060816\
 Data File : BF087831.D
 Acq On : 9 Jun 2016 3:44
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 07 18:51:55 2016



AutoFind: Scans 877, 878, 879; Background Corrected with Scan 874

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.2	10345	PASS
68	69	0.00	2	1.1	108	PASS
69	198	0.00	100	32.8	10219	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.0	15291	PASS
197	198	0.00	2	0.5	167	PASS
198	198	100	100	100.0	31201	PASS
199	198	5	9	6.5	2026	PASS
275	198	10	60	25.5	7956	PASS
365	198	1	100	3.0	943	PASS
441	442	0.01	24	14.7	4126	PASS
442	198	50	100	89.9	28036	PASS
443	442	15	24	19.4	5430	PASS