


Data Path : Z:\HPCHEM1\BNA F\Data\BF060817\
 Data File : BF095785.D
 Acq On : 8 Jun 2017 16:04
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 05 16:15:31 2017



AutoFind: Scans 2313, 2314, 2315; Background Corrected with Scan 2307

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.3	10808	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.0	11630	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.5	14521	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	30586	PASS
199	198	5	9	6.8	2075	PASS
275	198	10	60	23.6	7221	PASS
365	198	1	100	2.1	641	PASS
441	442	0.01	24	16.3	4402	PASS
442	198	50	100	88.0	26930	PASS
443	442	15	24	18.4	4967	PASS