

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060821\
 Data File : BF124173.D
 Acq On : 8 Jun 2021 10:17
 Operator : JU/CG
 Sample : PB136847BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB136847BL

Quant Time: Jun 08 11:28:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 07 13:32:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	42486	20.000	ng	-0.01
21) Naphthalene-d8	8.140	136	168101	20.000	ng	-0.01
39) Acenaphthene-d10	9.898	164	101419	20.000	ng	0.00
64) Phenanthrene-d10	11.386	188	193986	20.000	ng	# 0.00
76) Chrysene-d12	14.033	240	141390	20.000	ng	0.00
86) Perylene-d12	15.515	264	102378	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	357652	126.721	ng	0.00
7) Phenol-d6	6.504	99	441975	116.503	ng	0.00
23) Nitrobenzene-d5	7.428	82	343993	81.161	ng	0.00
42) 2,4,6-Tribromophenol	10.692	330	182743	127.182	ng	0.00
45) 2-Fluorobiphenyl	9.222	172	679366	84.293	ng	0.00
79) Terphenyl-d14	12.980	244	865544	84.048	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060821\
Data File : BF124173.D
Acq On : 8 Jun 2021 10:17
Operator : JU/CG
Sample : PB136847BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB136847BL

Quant Time: Jun 08 11:28:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 07 13:32:21 2021
Response via : Initial Calibration

