


Data Path : Z:\HPCHEM1\BNA F\DATA\BF060916\
 Data File : BF087844.D
 Acq On : 9 Jun 2016 11:13
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jun 07 18:51:55 2016



AutoFind: Scans 875, 876, 877; Background Corrected with Scan 872

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.0	15344	PASS
68	69	0.00	2	1.8	285	PASS
69	198	0.00	100	34.3	15485	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.6	22824	PASS
197	198	0.00	2	0.5	214	PASS
198	198	100	100	100.0	45110	PASS
199	198	5	9	6.6	2963	PASS
275	198	10	60	26.1	11782	PASS
365	198	1	100	2.8	1245	PASS
441	442	0.01	24	15.4	6417	PASS
442	198	50	100	92.1	41556	PASS
443	442	15	24	19.4	8060	PASS