

Data File : BF095821.D
Acq On : 9 Jun 2017 14:55
Operator : SJ/MA
Sample : I3457-17
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-16(10-14)

Quant Time: Jun 10 02:15:49 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	67061	20.00	ng	-0.03
21) Naphthalene-d8	9.40	136	274639	20.00	ng	-0.04
38) Acenaphthene-d10	12.22	164	120493	20.00	ng	-0.04
63) Phenanthrene-d10	14.62	188	192832	20.00	ng	-0.03
75) Chrysene-d12	18.34	240	138024	20.00	ng	-0.02
86) Perylene-d12	20.01	264	121504	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.26	112	511568	121.55	ng	-0.03
7) Phenol-d6	6.93	99	624854	124.02	ng	-0.03
23) Nitrobenzene-d5	8.29	82	351325	79.45	ng	-0.04
41) 2,4,6-Tribromophenol	13.52	330	110554	103.70	ng	-0.04
44) 2-Fluorobiphenyl	11.18	172	641900	74.13	ng	-0.04
78) Terphenyl-d14	16.99	244	402510	72.37	ng	-0.02
Target Compounds						Qvalue
49) Dimethylphthalate	11.87	163	108660	11.88	ng	99
70) Phenanthrene	14.65	178	52089	4.68	ng	96
74) Fluoranthene	16.43	202	64486	5.86	ng	99
77) Pyrene	16.74	202	62112	6.03	ng	96
80) Benzo(a)anthracene	18.33	228	24884	3.10	ng	# 94
82) Chrysene	18.37	228	23863	2.98	ng	# 92
87) Benzo(b)fluoranthene	19.61	252	28316m	3.72	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF095821.D

Acq On : 9 Jun 2017 14:55

Operator : SJ/MA

Sample : I3457-17

Misc :

ALS Vial : 10 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

B-16(10-14)

Quant Time: Jun 10 02:15:49 2017

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Mon Jun 05 16:15:31 2017

Response via : Initial Calibration

