

Data File : BF095836.D
Acq On : 9 Jun 2017 23:07
Operator : SJ/MA
Sample : I3566-12
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-07-COMP(0-4.0)

Quant Time: Jun 10 02:37:48 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	72799	20.00	ng	-0.03
21) Naphthalene-d8	9.40	136	292867	20.00	ng	-0.04
38) Acenaphthene-d10	12.22	164	126308	20.00	ng	-0.04
63) Phenanthrene-d10	14.62	188	196844	20.00	ng	-0.03
75) Chrysene-d12	18.33	240	154708	20.00	ng	-0.02
86) Perylene-d12	20.00	264	130689	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	33354	7.30	ng	0.04
7) Phenol-d6	6.93	99	257091	47.01	ng	-0.03
23) Nitrobenzene-d5	8.29	82	328212	69.60	ng	-0.04
41) 2,4,6-Tribromophenol	13.56	330	445	0.40	ng	0.00
44) 2-Fluorobiphenyl	11.18	172	642327	70.77	ng	-0.04
78) Terphenyl-d14	16.99	244	432650	69.40	ng	-0.02
Target Compounds						
49) Dimethylphthalate	11.87	163	84495	8.81	ng	Qvalue 97
83) Bis(2-ethylhexyl)phthalate	18.41	149	15835	2.23	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF095836.D
Acq On : 9 Jun 2017 23:07
Operator : SJ/MA
Sample : I3566-12
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-07-COMP(0-4.0)

Quant Time: Jun 10 02:37:48 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF060517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 05 16:15:31 2017
Response via : Initial Calibration

