

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060921\
 Data File : BF124205.D
 Acq On : 9 Jun 2021 12:52
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 14:34:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 14:34:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	86	0.00
2	1,4-Dioxane	40.000	37.569	6.1	83	0.00
3	Pyridine	40.000	36.498	8.8	80	0.00
4	n-Nitrosodimethylamine	40.000	36.729	8.2	80	0.00
5 S	2-Fluorophenol	80.000	77.565	3.0	85	0.00
6	Aniline	40.000	36.698	8.3	83	0.00
7 S	Phenol-d6	80.000	78.655	1.7	88	0.00
8	2-Chlorophenol	40.000	38.653	3.4	84	0.00
9	Benzaldehyde	40.000	39.041	2.4	90	0.00
10 C	Phenol	40.000	36.850	7.9	83	0.00
11	bis(2-Chloroethyl)ether	40.000	39.090	2.3	85	0.00
12	1,3-Dichlorobenzene	40.000	39.259	1.9	86	0.00
13 C	1,4-Dichlorobenzene	40.000	39.243	1.9	86	0.00
14	1,2-Dichlorobenzene	40.000	37.901	5.2	83	0.00
15	Benzyl Alcohol	40.000	35.079	12.3	79	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.894	12.8	79	0.00
17	2-Methylphenol	40.000	39.298	1.8	87	0.00
18	Hexachloroethane	40.000	38.438	3.9	85	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.768	5.6	88	0.00
20	3+4-Methylphenols	40.000	38.746	3.1	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	86	0.00
22	Acetophenone	40.000	41.283	-3.2	91	0.00
23 S	Nitrobenzene-d5	80.000	78.632	1.7	89	0.00
24	Nitrobenzene	40.000	38.029	4.9	83	0.00
25	Isophorone	40.000	37.362	6.6	84	0.00
26 C	2-Nitrophenol	40.000	38.506	3.7	85	0.00
27	2,4-Dimethylphenol	40.000	38.361	4.1	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.767	5.6	88	0.00
29 C	2,4-Dichlorophenol	40.000	39.229	1.9	88	0.00
30	1,2,4-Trichlorobenzene	40.000	39.031	2.4	89	0.00
31	Naphthalene	40.000	37.069	7.3	83	0.00
32	Benzoic acid	40.000	36.368	9.1	86	0.00
33	4-Chloroaniline	40.000	38.377	4.1	86	0.00
34 C	Hexachlorobutadiene	40.000	37.656	5.9	84	0.00
35	Caprolactam	40.000	39.058	2.4	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.336	6.7	88	0.00
37	2-Methylnaphthalene	40.000	38.502	3.7	85	0.00
38	1-Methylnaphthalene	40.000	38.602	3.5	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	83	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.980	-2.4	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.104	-7.8	103	0.00
42 S	2,4,6-Tribromophenol	80.000	78.561	1.8	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.262	1.8	86	0.00
44	2,4,5-Trichlorophenol	40.000	39.501	1.2	88	0.00
45 S	2-Fluorobiphenyl	80.000	80.102	-0.1	89	0.00
46	1,1'-Biphenyl	40.000	39.136	2.2	89	0.00
47	2-Chloronaphthalene	40.000	37.933	5.2	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060921\
 Data File : BF124205.D
 Acq On : 9 Jun 2021 12:52
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 09 14:34:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 09 14:34:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.069	2.3	88	0.00
49	Acenaphthylene	40.000	38.345	4.1	86	0.00
50	Dimethylphthalate	40.000	37.862	5.3	88	0.00
51	2,6-Dinitrotoluene	40.000	39.827	0.4	93	0.00
52 C	Acenaphthene	40.000	37.825	5.4	85	0.00
53	3-Nitroaniline	40.000	37.305	6.7	85	0.00
54 P	2,4-Dinitrophenol	40.000	38.585	3.5	90	0.00
55	Dibenzofuran	40.000	37.388	6.5	85	0.00
56 P	4-Nitrophenol	40.000	51.299	-28.2#	110	0.00
57	2,4-Dinitrotoluene	40.000	40.218	-0.5	87	0.00
58	Fluorene	40.000	37.476	6.3	83	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.778	3.1	89	0.00
60	Diethylphthalate	40.000	38.797	3.0	87	0.00
61	4-Chlorophenyl-phenylether	40.000	38.407	4.0	88	0.00
62	4-Nitroaniline	40.000	34.625	13.4	76	0.00
63	Azobenzene	40.000	38.064	4.8	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	81	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.984	7.5	81	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.525	3.7	84	0.00
67	4-Bromophenyl-phenylether	40.000	38.080	4.8	87	0.00
68	Hexachlorobenzene	40.000	37.251	6.9	82	0.00
69	Atrazine	40.000	40.948	-2.4	82	0.00
70 C	Pentachlorophenol	40.000	35.994	10.0	80	0.00
71	Phenanthrene	40.000	38.550	3.6	84	0.00
72	Anthracene	40.000	37.968	5.1	86	0.00
73	Carbazole	40.000	37.884	5.3	84	0.00
74	Di-n-butylphthalate	40.000	38.076	4.8	85	0.00
75 C	Fluoranthene	40.000	37.460	6.3	82	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	78	0.00
77	Benzydine	40.000	37.516	6.2	67	0.00
78	Pyrene	40.000	38.652	3.4	79	0.00
79 S	Terphenyl-d14	80.000	78.258	2.2	80	0.00
80	Butylbenzylphthalate	40.000	39.420	1.4	81	0.00
81	Benzo(a)anthracene	40.000	37.404	6.5	78	0.00
82	3,3'-Dichlorobenzidine	40.000	34.670	13.3	72	0.00
83	Chrysene	40.000	37.490	6.3	78	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.127	-0.3	82	0.00
85 c	Di-n-octyl phthalate	40.000	41.112	-2.8	86	0.00
86 I	Perylene-d12	20.000	20.000	0.0	83	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.712	8.2	83	0.00
88	Benzo(b)fluoranthene	40.000	37.624	5.9	87	0.00
89	Benzo(k)fluoranthene	40.000	38.042	4.9	85	0.00
90 C	Benzo(a)pyrene	40.000	37.568	6.1	85	0.00
91	Dibenzo(a,h)anthracene	40.000	36.502	8.7	84	0.00
92	Benzo(g,h,i)perylene	40.000	36.128	9.7	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060921\
Data File : BF124205.D
Acq On : 9 Jun 2021 12:52
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 09 14:34:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF060321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 09 14:34:04 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0